



IMEP-24: Analysis of eight heavy metals in toys according to EN 71-3:1994

Interlaboratory Comparison Report

Ines Baer, Johannes van de Kreeke, Inge Verbist, Danny Vendelbo, Thomas Linsinger, Piotr Robouch, Philip Taylor, Beatriz de la Calle



© Radius Images

EUR 24094 EN - 2009

The mission of the JRC-IRMM is to promote a common and reliable European measurement system in support of EU policies.

European Commission
Joint Research Centre
Institute for Reference Materials and Measurements

Contact information

Ms. Beatriz de la Calle
European Commission
Joint Research Centre
Institute for Reference Materials and Measurements
Retieseweg 111
2440 Geel, Belgium

E-mail: maria.de-la-calle@ec.europa.eu

Tel.: +32 (0) 14 571252
Fax: +32 (0) 14 571865

<http://irmm.jrc.ec.europa.eu/>
<http://www.jrc.ec.europa.eu/>

Legal Notice

Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of this publication.

***Europe Direct is a service to help you find answers
to your questions about the European Union***

**Freephone number (*):
00 800 6 7 8 9 10 11**

(*) Certain mobile telephone operators do not allow access to 00 800 numbers or these calls may be billed.

A great deal of additional information on the European Union is available on the Internet. It can be accessed through the Europa server <http://europa.eu/>

JRC 55720

EUR 24094 EN
ISBN 978-92-79-14273-4
ISSN 1018-5593
DOI 10.2787/191

Luxembourg: Office for Official Publications of the European Communities

© European Communities, 2009
Cover Illustration: © Radius Images
Reproduction is authorised provided the source is acknowledged

Printed in Belgium

IMEP-24: Analysis of eight heavy metals in toys according to EN 71-3:1994

Interlaboratory Comparison Report

November 2009

Ines Baer (*a*)
Johannes van de Kreeke (*e*)
Inge Verbist (*d*)
Danny Vendelbo (*d*)
Thomas Linsinger (*c*)
Piotr Robouch (*c*)
Philip Taylor (*e*)
Beatriz de la Calle (*b,c*)

(a) ILC coordinator, (b) IMEP programme coordinator,
(c) technical / scientific support, (d) administrative support,
(e) ILC conception



Contents

| | | |
|-----------|--|-----------|
| 1 | Summary | 5 |
| 2 | IMEP support to EU policy | 5 |
| 3 | Introduction | 6 |
| 4 | Scope and aim | 7 |
| 5 | Time frame | 7 |
| 6 | Invitation, registration and distribution | 7 |
| | 6.1 Confidentiality | 8 |
| | 6.2 Distribution | 8 |
| | 6.3 Procedure to apply | 9 |
| 7 | Test material | 9 |
| | 7.1 BCR 620 | 10 |
| | 7.2 Homogeneity and stability study | 10 |
| | 7.2.1 Homogeneity | 10 |
| | 7.2.2 Stability | 11 |
| 8 | Reference values and their uncertainties | 11 |
| | 8.1 Target values | 11 |
| | 8.2 Establishing reference values and uncertainties (X_{ref} , U_{ref})..... | 12 |
| | 8.2.1 Comparison with BCR-620 certification | 12 |
| | 8.2.2 Reference value for mercury..... | 13 |
| | 8.2.3 The standard deviation for proficiency assessment $\hat{\sigma}$ | 14 |
| 9 | Reported results | 15 |
| | 9.1 General observations..... | 15 |
| | 9.2 Measurement results..... | 15 |
| | 9.3 Scoring of results..... | 17 |
| | 9.3.1 Scores and evaluation criteria | 17 |
| | 9.3.2 Scoring the reported measurement results | 18 |
| | 9.4 Results interpretation by participants | 19 |
| | 9.4.1 Reporting of results..... | 20 |
| | 9.4.2 Assessment of compliance with EN 71-3:1994 | 20 |
| | 9.5 Further information extracted from the questionnaire..... | 22 |
| 10 | Conclusion | 24 |
| 11 | Acknowledgements | 25 |
| | Abbreviations | 27 |
| | References | 28 |
| | Annexes | 30 |

1 Summary

The Institute for Reference Materials and Measurements (IRMM) of the Joint Research Centre (JRC), a Directorate-General of the European Commission, operates the International Measurement Evaluation Programme (IMEP). It organises interlaboratory comparisons (ILC's) in support to EU policies. This report presents the results of an ILC which focussed on the determination of soluble antimony, arsenic, barium, cadmium, chromium, lead, mercury, and selenium according to European Standard EN 71-3:1994.

The principle of the procedure in EN 71-3:1994 [1] consists in the extraction of soluble elements from toy material under the conditions simulating the material remaining in contact with stomach acid for a period of time after swallowing.

Forty participants from eighteen countries registered to the exercise, of which 33 reported results for As, 35 for Ba and Se, 37 for Cr, Pb, and Sb, 38 for Hg, and 39 for Cd. For seven measurands the test material had already been certified in the past. The validity of the certificate was reconfirmed and the certified values were taken as the reference values for this ILC. As no certified value was available for Hg, the mean value of the results provided by four expert laboratories was used together with the corresponding uncertainty. Participants were invited to report the uncertainty on their measurements. This was done by 35 of the 39 laboratories having submitted results in this exercise.

Laboratory results were rated with z and zeta scores in accordance with ISO 13528 [2]. The standard deviations for proficiency assessment were based on the analytical correction laid down in EN 71-3:1994.

2 IMEP support to EU policy

The International Measurement Evaluation Programme (IMEP) is organised by the Joint Research Centre - Institute for Reference Materials and Measurements. IMEP provides support to the European measurement infrastructure in the following ways:

IMEP **disseminates metrology** from the highest level down to the field laboratories. These laboratories can benchmark their measurement result against the IMEP certified reference value. This value is established according to metrological best practice.

IMEP helps laboratories to assess their estimate of **measurement uncertainty**. The participants are invited to report the uncertainty on their measurement result. IMEP integrates the estimate into the scoring, and provides assistance for the interpretation.

IMEP **supports EU policies** by organising intercomparisons in the frame of specific EU Directives, or on request of a specific Directorate-General. In the case of the IMEP-24, it was realised in the context of the Toy Safety Directive [3] applying the European Standard EN 71-3:1994.

IMEP-24 provided specific **support to the European Co-operation for Accreditation (EA)** in the frame of a Memorandum of Understanding (MoU) on a number of metrological issues, including the organisation of intercomparisons. National accreditation bodies were invited to nominate a limited number of laboratories for free participation in IMEP-24. The Swedish Board for Accreditation and Conformity Assessment (SWEDAC) liaised between EA and IMEP for this intercomparison.

3 Introduction

The requirements set up in EN 71-3:1994 are for the migration of heavy metals from the following toy materials: coatings, polymeric and similar materials, paper and paper board, textiles, glass/ceramic/metallic materials, materials intended to leave a trace, pliable modelling materials, paints and other materials [1]. The material of interest for this inter-laboratory comparison is "coating", which is defined in the standard as:

"All layers of material formed or deposited on the base material or toy and includes material which includes paints, varnishes, lacquers, inks, polymers or other substances of a similar nature, whether they contain metallic particles or not, no matter how it has been applied to the toy and which can be removed by scrapping with a sharp blade."

Concerned heavy metals are antimony, arsenic, barium, cadmium, chromium, lead, mercury, and selenium. Their migration from toys should comply with the limits listed in Table 1 when tested according to the procedure given in the standard. An analytical correction is allowed for each element and is listed in the same table. Concretely, it means that the analytical result can be reduced by the given percentage when the analytical result equals or exceeds the set limit.

Table 1 – The eight heavy metals and their maximum limits set in EN 71-3:1994

| Element | Sb | As | Ba | Cd | Cr | Pb | Hg | Se |
|---|----|----|------|----|----|----|----|-----|
| Limit migrated element (X_{EN}) [mg/kg] | 60 | 25 | 1000 | 75 | 60 | 90 | 60 | 500 |
| Analytical correction (AC) [%] | 60 | 60 | 30 | 30 | 30 | 30 | 50 | 60 |

4 Scope and aim

The aim of this proficiency test is to enable laboratories performing tests on toy products to monitor their performance and to compare with other laboratories from Europe and abroad. Another aim is to identify problems related to technique and methodology. This was particularly interesting in this exercise, since the sample preparation procedure to be applied is known to cause great spread of results. The observation of this spread in former interlaboratory trials actually led to the introduction of the analytical correction into the EN 71-3:1994 [1].

5 Time frame

The project started on 18 March 2009. The EA coordinator Annika Norling then informed the national accreditation bodies. Four expert laboratories contributing to the establishment of the reference value were invited to register on the 26 March 2009 and were sent the samples on 27 April 2009. The exercise was publicly announced on the IMEP webpage¹ in the beginning of May 2009. In parallel, laboratories specialised in toy safety related analyses were contacted.

Interested laboratories could register until Friday 22 May 2009. Samples were sent out to the laboratories on 28 May 2009. For all laboratories the deadline for reporting results was on 3 July 2009.

6 Invitation, registration and distribution

Invitations for participation were sent to the EA coordinator (Annex 1) for distribution to nominated laboratories. An invitation letter was also sent to four expert laboratories to ask them to register for the establishment of the reference value (Annex 2). Notified bodies

¹ http://irmm.jrc.ec.europa.eu/html/interlaboratory_comparisons/

from the NANDO list were sent an email (Annex 3) inviting them to take part in the exercise, after having retrieved their contact information from the NANDO webpage². NANDO lists notified bodies fulfilling the relevant requirements and which can be designated to carry out conformity assessment according to a directive, which in this case is the Toy Safety Directive. Finally, a call for participation was also released on the IRMM website (Annex 4).

Instructions on measurands, sample storage, reconstitution and measurement were sent to the participants in an accompanying letter together with the samples. The letter also contained the individual "code for access" to the result reporting website and the deadline for reporting (Annex 5). The reporting website included a questionnaire to collect additional information related to the experimental work (Annex 6).

The participants who submitted their result received the reference values two weeks after the reporting deadline. Forty laboratories registered out of which thirty-nine submitted results. Fig 1 represents the participating countries.

6.1 Confidentiality

EA was invited to nominate laboratories for participation. The following confidentiality statement was made to EA: "Confidentiality of the participants and their results towards third parties is guaranteed. However, IMEP will disclose details of the participants that have been nominated by EA to the EA working group for ILCs in Testing. The EA accreditation bodies may wish to inform the nominees of this disclosure."


6.2 Distribution

ILC samples were dispatched by IRMM on 27 April 2009 to the certifying laboratories and on 28 May 2009 to participants. Each laboratory received one package containing a coated mild steel plate, the 'Confirmation of receipt' form (Annex 7) and an accompanying letter with instructions on sample handling, procedure and timelines (Annex 5).

The dispatch was followed by the courier's parcel tracking system on internet and in most of the cases the sample was delivered within a couple of days. In one case, the dispatch took 2 weeks. It was however assumed that the parcel was not submitted to high enough temperatures or long enough time to have an impact on the samples' stability.

² <http://ec.europa.eu/enterprise/newapproach/nando/>

Fig 1– Participating countries, number of laboratories



6.3 Procedure to apply

As this exercise was run to verify the performance of the laboratories when applying the EN 71-3:1994 [1], they were recommended to apply the corresponding procedure. Concerning the quantitative analysis of migrated elements, the standard recommends the use of methods having a detection limit of a maximum of 1/10 of the values to be determined.

7 Test material

The test material used for this exercise is the certified reference material BCR 620 and consists of alkyd resin paint on a mild steel plate. This material was certified in 1998 for levels of toxic element migration using the method specified in the EN 71-3:1994 [1]. All elements except mercury were certified. The BCR 620 was taken out of sales because of doubts of stability observed during monitoring analysis. Four expert laboratories analysed the test material again for confirmation. Their reported values would be taken as reference values (X_{ref}) in case the assumed degradation would be confirmed and the original certified values could not be used.

7.1 BCR 620

The certification report is not available for the public since the material is not commercialised anymore. However, details about the certification can be found in publications [4, 5] and are summarised hereafter. The paint was ordered at a specialised paint manufacturing company Trimite Ltd (UK). It was adulterated with 8 toxic elements at concentrations sufficient to yield soluble element concentrations at or around the maximum permissible levels. The paint was produced using dark grey "base" paint and adding a series of "tinters" each containing one of the eight toxic elements. A large single batch was produced, thoroughly mixed and dispensed into a number of 5 L tins before passing to the paint spraying contractor, Auto Imagination Ltd (UK). Before the spraying each tin was thoroughly re-mixed and as part of the spraying process the paint was passed through a turbulent mixing chamber in front of the spray jet. The amount sprayed on each panel was such that it was possible to scrape off at least 1 g of paint. The used mild steel plates were of size 150 x 100 mm and were degreased and abraded by sand blasting before application of the paint.

The certification measurements were carried out by 10 expert laboratories following the sample preparation procedure given in EN 71-3:1994 but using different instrumental techniques to analyse the sample extracts. The mean of the laboratory means was adopted as the certified value for each element in BCR 620 (Table 3).

7.2 Homogeneity and stability study

Since the material has been withdrawn from the market it was decided to carry out a homogeneity and a short-term stability study before starting the ILC. The study was performed by the Istituto Italiano Sicurezza dei Giocattoli S.r.l., Cabiato – Co (IT). Homogeneity and stability were done on the soluble elements as listed in EN 71-3:1994 (hence including Hg). Samples stored at 18 °C were used for the homogeneity study, and samples stored at 4 °C for six years were used as reference samples for the stability study.

7.2.1 Homogeneity


The laboratory performing the homogeneity study received 10 randomly chosen plates from the sample set stored at 18 °C and analyses were performed in duplicate following the procedure given in EN 71-3:1994 [1]. Results were evaluated according to ISO 13528 [2] and the IUPAC Harmonised Protocol guidelines [6] which describe tests to determine whether an ILC material is sufficiently homogeneous for its purpose.

The results of the homogeneity study can be found in Annex 8. The material passed the test with all elements according to the IUPAC calculations. This was considered sufficient to carry out this ILC, although ISO 13528 criteria were not met for all elements.

7.2.2 Stability

Stability was checked with the samples stored at 4 °C for six years. The results were compared to the results obtained with the samples analysed in the homogeneity study, stored at 18 °C, as well as to the certified reference values established ten years ago (cf. Ch 8). The various results were overlapping within their uncertainties (Fig 2) despite being analysed at such long time interval. Taking furthermore into account the sample preparation inherent variation, no stability issues were expected for the duration of the trial.

Fig 2 – Results of the stability & homogeneity study, shown together with the BCR values. (Concentrations and expanded uncertainties in logarithmic scale)



8 Reference values and their uncertainties

8.1 Target values

By target values is meant the concentration of heavy metals aimed at when producing the material. In this exercise they were set by the elements' concentrations of the material available. This material has been specifically produced for the toy safety norm for

which the limits are set in EN 71-3:1994 [1] and target values were aimed at being close to these limits. Thus, the material was perfectly fit-for-purpose.

8.2 Establishing reference values and uncertainties (X_{ref} , U_{ref})

Four expert laboratories were contacted to perform accurate analysis so that their values could be used to either confirm the reference values from the certificate, or for the establishment of new reference values. Additionally, a reference value had to be determined for mercury, where no certified value was available. The four expert laboratories are:

- Istituto di Ricerche e Collaudi, M. Masini S.r.l., Milano (IT)
- Istituto Italiano Sicurezza dei Giocattoli S.r.l., Cabiato – Co (IT)
- LGC Ltd, Teddington (UK)
- Swedish National Testing and Research Institute, Borås (SE)

8.2.1 Comparison with BCR-620 certification

Concerning the reference values of antimony, arsenic, barium, cadmium, chromium, lead and selenium, the results of the four expert laboratories were compared with each other and then with the certified values. If the four laboratories reported results which agreed within their expanded uncertainties, the mass fraction of the different measurands was calculated as the mean of the four independent reported results, X_{exp} . The associated uncertainty is calculated by propagating contributions from characterisation (u_{char}) and homogeneity (u_{bb}) as follows [7]:

$$u_{exp} = \sqrt{(u_{char}^2 + u_{bb}^2)} \quad (\text{all standard uncertainties}) \quad \text{Eq.1}$$

where the uncertainty of characterisation u_{char} is calculated from the uncertainties reported by the four laboratories [8]:

$$u_{char} = \sqrt{(u_1^2 + u_2^2 + u_3^2 + u_4^2)} / 4 \quad (\text{all standard uncertainties}) \quad \text{Eq.2}$$

To evaluate the significance of the difference between the values from the expert laboratories (X_{exp} , U_{exp}) on one side, and the certified values (X_{BCR} , U_{BCR}) on the other side, the E_n number is established according to:

$$E_n = \frac{|X_{\text{exp}} - X_{\text{BCR}}|}{\sqrt{U_{\text{exp}}^2 + U_{\text{BCR}}^2}} \quad \text{Eq.3}$$

If $E_n < 1$, then there is no significant difference between the measurement result and the certified value and X_{exp} and X_{BCR} are in agreement within their associated expanded uncertainties (U_{exp} and U_{BCR}) calculated with a coverage factor $k=2$ [2, 9].

The results obtained by the four expert laboratories and their expanded uncertainties are represented graphically in Annex 9 together with the BCR values. The numerical values and the calculated E_n numbers for each element can be found in Table 3. The results show that experimental values and certified values are in agreement and thus, the original certified values could be used as reference values for all measurands in this exercise (except Hg).

8.2.2 Reference value for mercury


The mass fraction of mercury had not been certified for BCR 620. Thus, the results of the four expert laboratories were used for the establishment of the assigned value. As shown in Fig 3 the four reported values are within the range covered by the expanded uncertainties. The assigned value X_{ref} is thus calculated as the mean of these values and its associated uncertainty U_{ref} by using the calculations in Eq. 1 and Eq. 2. The numerical results and corresponding calculations are summarised in Table 2 below.

Table 2 – Determination of the reference value for mercury. Values from measurements and calculations. All values in (mg kg^{-1}).

| Laboratory | LGC | Masini | Ist. Sic. Gioc. | SP | X_{ref} | U_{char} | U_{bb} | u_{ref} | U_{ref}^* |
|------------------|-----|--------|-----------------|-----|------------------|-------------------|-----------------|------------------|--------------------|
| X_{lab} | 390 | 255 | 397 | 438 | 370 | | | | |
| U_{lab} | 56 | 19 | 9 | 110 | | 31 | 56 | 64 | 127 |

* Expanded uncertainty with a coverage factor $k=2$ which corresponds to a level of confidence of about 95%[10]

Fig 3 - The results of the four laboratories for mercury together with the determined reference value and their uncertainties



8.2.3 The standard deviation for proficiency assessment $\hat{\sigma}$

The standard deviation for proficiency testing $\hat{\sigma}$ is an estimate of the expected / required variability of the trial. It has to be determined for each ILC individually. In this exercise, it was established based on the analytical corrections (AC) given in EN 71-3:1994. These were interpreted as expanded uncertainties. Thus, $\hat{\sigma}$ was set as half the AC, assuming a confidence interval of 95% (Table 3).

Values highlighted in green in Table 3 below were used for the evaluation of the analytical performance of the laboratories. Consequently, the terms X_{BCR} and U_{BCR} will be replaced by X_{ref} and U_{ref} from this point on.

Table 3 – Establishment of reference values, their uncertainties and $\hat{\sigma}$. X_{BCR} , U_{BCR} , X_{exp} , U_{exp} in (mg kg⁻¹).

| | Sb | As | Ba | Cd | Cr | Pb | Hg | Se |
|-------------------------|---------|------------|----------|----------|---------|----------|--------------|----------|
| $X_{BCR} \pm U_{BCR}^*$ | 83 ± 19 | 23.0 ± 6.3 | 430 ± 50 | 117 ± 21 | 64 ± 24 | 140 ± 40 | | 240 ± 60 |
| $X_{exp} \pm U_{exp}^*$ | 66 ± 14 | 17.9 ± 4.4 | 429 ± 88 | 138 ± 33 | 58 ± 18 | 139 ± 20 | 370 ± 127 | 181 ± 46 |
| E_n number | 0.7 | 0.7 | 0.0 | -0.5 | 0.2 | 0.0 | | 0.8 |
| AC | 60% | 60% | 30% | 30% | 30% | 30% | 50% | 60% |
| $\hat{\sigma}$ | 30% | 30% | 15% | 15% | 15% | 15% | 25% | 30% |

* Expanded uncertainty with a coverage factor k=2 which corresponds to a level of confidence of about 95%[10]

9 Reported results

9.1 General observations

From the 40 laboratories that registered, 39 submitted results and completed the associated questionnaire; 1 cancelled its participation. One laboratory reported for each element "less than" values and thus was not given any scores. A few other laboratories reported this as well, but only for 1, 2 or 5 elements. However, most of participants reported measurement results for all eight elements.

No obvious wrong result reporting was observed, although there are sometimes very low or high values. However, these were either punctually appearing for one or two elements, or they were confirmed by the participants (when re-contacting them for the analytical correction, see §9.2).

Two laboratories complained about the insufficient amount of paint and one about the insufficient amount of metal in the plate. For a reminder, the test material was conceived in view of the standard EN 71-3:1994 and the there stated amounts. One laboratory observed difficulties in sampling due to the thin lacquer, and also electrostatic loading.

9.2 Measurement results

It was observed that the reported results were spread following a bi-modal distribution. This was mostly due to the application or not of the analytical correction (AC) given in the standard EN 71-3:1994. Only a few laboratories mentioned whether or not they used the AC (8 out of 39). The situation was so unclear, that it was necessary to contact all participants again in order to request specifications.


Considering this particular situation, all calculations were done on the raw non-corrected data. For each element, the results are shown in a table including the scorings and in a graph (Annex 10 to 17). The corresponding Kernel density plots can be found in Annex 18. An overview of all scorings can be found in Annex 19.

The results are generally normally distributed around the assigned value, or at least not much deviating from it (see results' graphs and Kernel plots, Annex 10 to 18). Some sub-populations can be observed in the Kernel plots mainly due to punctual very high or very low results. It is also suspected that some sub-populations are still due to the effects of the application of the analytical correction. When requesting details about the results, five

laboratories did not answer (laboratories 029, 150, 239, 285, and 422). No answer was taken as non-application of analytical correction.

For selenium, mercury and, to a lesser degree, for arsenic there was a general tendency to underestimate the mass fraction. When using robust statistics and calculating the median from the results for the respective elements, those for mercury and selenium are lower than the corresponding X_{ref} value. Indeed, these elements are known to be volatile and difficult to analyse, but this should not result in such a bias. The instrumental detection appeared to have some influence on the results. About 85% of the laboratories used Inductively Coupled Plasma–Mass Spectrometry (ICP-MS) or –Optical Emission Spectrometry (OES) and the remaining 15% atomic absorption spectroscopy (AAS), except in the case of mercury where about one third of the participant choose AAS. When the results for Hg were verified for any differences due to the methods, results obtained with AAS appeared to be lower than those obtained with ICP techniques. However, even the latter are characterised by a lower median than X_{ref} . For illustration see Fig 4. The lower results for ICP-OES could be partly explained by difficult detection, but no explanation could be found for the other two.

Fig 4 – Mercury values shown in function of the analytical method applied.



The results for selenium and arsenic were also checked for an eventual influence of the techniques used and it appears that a similar influence on the results could be observed with selenium, however not as pronounced. No influence was observed for arsenic. Considering that low results have been observed with all three techniques, the main error contributor seems to be the sample preparation.

For the remaining measurands no influence of the analytical method used was observed, with most of the participants having used ICP-OES, some ICP-MS. Other eventual factors of influence were verified according to the questionnaire's answers, but no other tendencies could be detected.

The software used to calculate robust statistics and kernel densities was provided by the Statistical Subcommittee of the Analytical Methods Committee (AMC) of the Royal Society of Chemistry [11, 12].

9.3 Scoring of results

9.3.1 Scores and evaluation criteria

Individual laboratory performance is expressed in terms of z- and zeta scores in accordance with ISO 13528 [2] and the IUPAC International Harmonised Protocol [6]:

$$z = \frac{X_{\text{lab}} - X_{\text{ref}}}{\hat{\sigma}} \quad \text{and} \quad \text{zeta} = \frac{X_{\text{lab}} - X_{\text{ref}}}{\sqrt{u_{\text{ref}}^2 + u_{\text{lab}}^2}}$$

Where

X_{lab} is the measurement result reported by a participant

X_{ref} is the certified reference value (assigned value)

u_{ref} is the standard uncertainty of the reference value

u_{lab} is the standard uncertainty reported by a participant

$\hat{\sigma}$ is the standard deviation for proficiency assessment

Both scores can be interpreted as: satisfactory result for $|\text{score}| \leq 2$, questionable result for $2 < |\text{score}| \leq 3$ and unsatisfactory result for $|\text{score}| > 3$.

z score

The z score indicates whether a laboratory is able to perform the measurement in accordance with what can be considered as good practice within the EU. The standard deviation for proficiency assessment $\hat{\sigma}$ is accordingly based on experience described in the standard EN 71-3:1994. The $\hat{\sigma}$ applied in this ILC can be found in Table 3.

The IUPAC International Harmonised Protocol [6] suggests that participants can apply their own $\hat{\sigma}$ and recalculate the scores if the purpose of their measurements is different.

zeta score

The zeta score provides an indication of whether the estimate of uncertainty is consistent with the laboratory's deviation from the reference value [6]. It is calculated only for those results that were accompanied by an uncertainty statement. The interpretation is similar to the interpretation of the z score. An unsatisfactory zeta score may be caused by an underestimated uncertainty or by a large deviation from the reference value.

The standard uncertainty of the laboratory (u_{lab}) was calculated as follows. If an expanded uncertainty was reported, it was divided by the coverage factor k . If no coverage factor was provided, the reported uncertainty was considered as the half-width of a rectangular distribution. The reported uncertainty was then divided by $\sqrt{3}$, in accordance with recommendations issued by Eurachem and CITAC [13].

9.3.2 Scoring the reported measurement results

Scores were calculated with the raw data for all participants (not corrected for AC) except for those who reported no value or a "less than" value. These results were not used in any statistical calculation. A zeta score was calculated for results that were accompanied by an uncertainty statement. Annexes 10-17 list the scores per element and laboratory in detail, and Annex 19 summarises the scores per participant.

Table 4 summarises the distribution of scores per element. A large share of participants reported satisfactory measurement results, a small share unsatisfactory results. This is visible in good z scores, except for cadmium, lead and mercury, where satisfactory scores are below 50 % and over 35% of the laboratories got unsatisfactory results. Thus, generally the participants performed well. It should however be kept in mind that the results were evaluated with a large $\hat{\sigma}$.

The situation is slightly different for the zeta scores. Here, only two elements (barium and chromium) had over 50% of the participants getting satisfactory scores. That means that although the results reflected by the z scores are generally good, there is an obvious problem with the estimation of the uncertainty for some elements, resulting in a high number of unsatisfactory zeta scores. For all elements, except chromium, the share of satisfactory scores dropped from z score to zeta score or stayed at a similarly low level. This is particularly visible in the case of selenium, where also the number of participants having reached satisfactory scores for both, z and zeta, is the lowest for all elements.

Table 4 – Overview of scores with S(atisfactory), Q(uestionable) and U(nsatisfactory)

| z scores | | | | | | | | zeta scores | | | | | | | | z & zeta S(#*) |
|-----------|----|--------|------|-------|------|-------|------|-------------|----|--------|------|-------|------|-------|------|-------------------|
| | n* | S (#*) | S(%) | Q(#*) | Q(%) | U(#*) | U(%) | | n* | S (#*) | S(%) | Q(#*) | Q(%) | U(#*) | U(%) | |
| Sb | 37 | 29 | 78% | 5 | 14% | 3 | 8% | Sb | 33 | 15 | 45% | 6 | 18% | 12 | 36% | 15 |
| As | 33 | 26 | 79% | 6 | 18% | 1 | 3% | As | 30 | 13 | 43% | 9 | 30% | 8 | 27% | 13 |
| Ba | 35 | 29 | 83% | 2 | 6% | 4 | 11% | Ba | 32 | 17 | 53% | 5 | 16% | 10 | 31% | 16 |
| Cd | 39 | 17 | 44% | 5 | 13% | 17 | 44% | Cd | 35 | 13 | 37% | 4 | 11% | 18 | 51% | 13 |
| Cr | 37 | 26 | 70% | 5 | 14% | 6 | 16% | Cr | 33 | 28 | 85% | 0 | 0% | 5 | 15% | 24 |
| Pb | 37 | 16 | 43% | 6 | 16% | 15 | 41% | Pb | 34 | 16 | 47% | 5 | 15% | 13 | 38% | 14 |
| Hg | 38 | 15 | 39% | 9 | 24% | 14 | 37% | Hg | 35 | 12 | 34% | 4 | 11% | 19 | 54% | 12 |
| Se | 35 | 25 | 71% | 8 | 23% | 2 | 6% | Se | 32 | 9 | 28% | 10 | 31% | 13 | 41% | 9 |

* n is the number of results for which a score was given. The total number of participants is 39.
 (#) – number of laboratories

Most of the participants provided an uncertainty estimate, and most of these estimates were accompanied by a coverage factor. Only four participants did not report any uncertainty at all. This is encouraging, but contrasts with the modest share of results with a satisfactory zeta score. Considering that more than half of the participants stated in the questionnaire that they do not usually report the uncertainty to their customers, one might think that this is the reason for the lack of experience in uncertainty estimation. However, when plotting the scores in function of the reporting / non-reporting to customers, there is no tendency that those reporting uncertainties to their customers do better. As conclusion, participants are advised to verify their zeta scores, and review the principles of uncertainty estimation described in the GUM [10] and in related guidance for the field of analytical chemistry, e.g. the EURACHEM / CITAC Guide [13]. As guideline, values adopted for $\hat{\sigma}$ seem to give a realistic estimate for the average measurement uncertainty.

9.4 Results interpretation by participants

The standard EN 71-3:1994 appears to be insufficient when it comes to the question of how to deal with the analytical results and how to submit them to end-users. In §4.2 it says :

"The analytical results of materials established in clauses 7, 8 and 9 shall be adjusted by subtracting the analytical correction in Table 2 to obtain an adjusted analytical result. Materials are deemed to comply with this standard if the adjusted analytical result is less than or equal to the limits in Table1 (See Annex D)."

while Annex D.4 states :

"To achieve consistent interpretation of results, a correction factor for each element has been introduced into the standard applicable to all instrumental techniques. These are taken from the precision data in the 1988 standard and are used when an analytical result equals or exceeds the maximum limit."

Although the AC "shall be applied" according to the standard, it is not as clear whether such correction should be applied by the laboratory performing the analysis or i.e. by control authorities, or customers. This led to problems in the reporting and thus the evaluation of this exercise. For this reason, it was decided to separate the evaluation in two parts – one "classical" ILC evaluation and one evaluation about the reporting and interpretation of results. The former considers mainly the analytical performance of the laboratories and has been treated in the precedent chapters. The latter will be treated hereafter.

9.4.1 Reporting of results

When setting-up this exercise it was decided not to give any specification concerning the application of the AC, the reason being that reporting is part of the proficiency test and as such should be taken into account in the evaluation. Thus, participants were asked to submit results as they would do to customers. The way of reporting a result appears to be particularly important and this was clearly an issue in this exercise. Four different behaviours were observed:

- Application of the analytical correction 'by default' (to all elements) (13 labs): defensible for practical reasons but not always necessary; e.g. in the case of arsenic, barium and selenium, where most or even all of the laboratories had results below the legislative limit, there was no need to apply it.
- No application at all (21 labs): justified, if low results (5 labs). Otherwise not (16 labs), since the results for some elements are close or above the limit. Here the question is: "How real cases are dealt with?".
- Application only to the means (while reporting the uncorrected measurements): unclear, requires specification. (3 labs)
- Application to a part of the analysed elements: justified, but requires specification. (4 labs)

Considering that the participants were to follow the standard EN 71-3:1994, and thus were aware of the presence of the clause concerning the AC, it would be desirable if reporting of results could be improved as to include specifications about its application or non-application. This making the reporting more transparent and comprehensive.

9.4.2 Assessment of compliance with EN 71-3:1994

The different ways of reporting led us to the question of how participants dealt with the overall interpretation of their results, and thus, they were asked in the questionnaire

whether they 'would accept or reject the entrance of the material on the market' (cf. Annex 6). Their assessment should be based on a simple test for each of the eight heavy metals: does the mean of my reported values X_{lab} fall below or exceed the maximum tolerable values X_{max} :

$$X_{lab} < X_{max} : \text{compliant}$$

$$X_{lab} > X_{max} : \text{non compliant}$$

The parameter X_{max} was calculated for the purpose of this ILC and follows logically the concept of the AC laid down in EN 71-3:1994. X_{max} values are always larger than the official legal limits X_{EN} listed in EN 71-3:1994 as they compensate for the analytical correction AC as follows:

$$X_{max} = X_{EN} * 100 / (100 - AC) \quad AC \text{ in } [\%]$$

The resulting X_{max} values are listed in Table 5 together with the values for X_{EN} , X_{ref} and AC. The table also shows that the material is compliant for Sb, As, Ba, Cr and Se because $X_{ref} < X_{max}$ but not compliant for Cd, Pb and Hg because $X_{ref} > X_{max}$. This renders the test material overall not compliant (cf. EN 71-3:1994).

Table 5 – Relevant values for results interpretation. X_{EN} , X_{max} , X_{ref} in ($mg\ kg^{-1}$).

| | Sb | As | Ba | Cd | Cr | Pb | Hg | Se |
|-----------------------------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|
| X_{EN} | 60 | 25 | 1000 | 75 | 60 | 90 | 60 | 500 |
| X_{max} | 150 | 62.5 | 1429 | 107 | 86 | 129 | 120 | 1250 |
| X_{ref} | 83 | 23.0 | 430 | 117 | 64 | 140 | 370 | 240 |
| AC | 60% | 60% | 30% | 30% | 30% | 30% | 50% | 60% |


Knowing the established reference values X_{ref} of the test material this leads to four potential situations represented in the following contingency table :

| | | 'to be accepted' $X_{ref} < X_{max}$ | 'to be rejected' $X_{ref} > X_{max}$ |
|---------------|---------------------|---|---|
| Compliant | $X_{lab} < X_{max}$ | TN* | FN* |
| Non-compliant | $X_{lab} > X_{max}$ | FP* | TP* |

*TN – True Negative, FN – False Negative, FP – False Positive, TP – True Positive

The concept of this extended analysis is graphically displayed in Fig 5 for clarity. The analysis is very worthwhile, because it shows where wrong decisions were taken on the basis of a correct interpretation of wrong measurement results. An FN result may lead to the acceptance of a material which contains in reality unacceptable concentration of heavy metals, whereas an FP result may lead to financial losses because of undue material rejection.

Fig 5 – Graphical presentation of true positives(TP) / negatives(TN) and false positives(FP) / negatives(FN)



Annex 19 shows the scorings (z and zeta) of the participants for each element and their overall assessment of material compliance. It also shows the TN/FN/TP/FP labels that have been added by the ILC organiser (Int). As overall result, most laboratories have correctly assessed the material as being non compliant. However, nine laboratories were found to give an incorrect interpretation of their results: (i) either because they wrongly 'accepted the test material on the market', or (ii) because their decision was inconsistent with the results, meaning they 'reject the material' although their results do not give reason for rejection. These participants are highlighted in bold red in Annex 19.

Another critical observation was that two laboratories correctly provided a sound overall assessment, although their measurement results were very high or very low (lab 058 and 298). This implies that their overall decision might be correct, but their results are not reliable. This was also readily visible in their large z- and zeta scores.

9.5 Further information extracted from the questionnaire

In addition to submission of the results, the participants were asked to answer a number of questions relating to the measurements. All participants completed the questionnaire, although a few participants skipped a big part of it. Since this exercise was carried out using the EN 71-3:1994, many questions were related to the sample preparation described

in it. All except one laboratory indeed used the EN 71-3:1994 for the required analysis, and six participants used the standard but modified it. A sum-up of the answers to the method related questions can be found in Annex 20.

Participants were asked about the confidence level reflected by the coverage factor (k) reported. Nineteen laboratories reported a level of 95% and fourteen laboratories did not reply to this question. Two laboratories did not understand the question and six put the coverage factor k. One participant gave a level of 65% and two reported -1%. For the uncertainty estimate, several participants gave various combinations of the given choices. Twenty-three use the uncertainty of the method as determined during in-house validation. Twenty-two laboratories estimated the uncertainty using data from measurements of replicates (i.e. precision). Seven laboratories applied the ISO-GUM. Five used the known uncertainty of the standard method. It has to be emphasised that the latter should not be used on its own - the reproducibility of a standard method should always be verified by the laboratory applying it. Four laboratories made use of intercomparison data. Three made an estimation based on judgement [13] (expert guesstimate). Fifteen laboratories provide an uncertainty estimate to their customer and twenty-six do not.

Forty laboratories carry out this type of analysis regularly. However, the number of samples analysed by the forty laboratories varies a lot as can be seen in Table 6 where the number of samples per year is reported.

Table 6 – Reported number of samples analysed per year

| Number of samples per year | <50 | 50 - 250 | 250 - 1000 | >1000 |
|----------------------------|-----|----------|------------|-------|
| Number of laboratories | 14 | 6 | 7 | 13 |

All participants have a quality system in place. Thirty-eight laboratories have a quality system based on ISO 17025, three have a quality system based on both ISO 17025 and ISO 9000 series, one based on ISO 9000 series, and the remaining laboratories use either a combination of the above with another system or another system altogether. Thirty-seven laboratories are accredited and six laboratories are not. Twenty-one laboratories take part in an interlaboratory comparison on a regular basis.

Fourteen laboratories use a reference material for this type of analysis. Thirteen laboratories use the reference material for the validation and nine for calibration. Five laboratories did not specify which reference material they use. Table 7 summarises the reference materials used for the validation of the methods as reported by the participants. Considering the answers with regard to the type of reference material used, care should be taken when using a standard solution delivered in connection with an instrument as reference material. A standard solution does not allow the assessment of the overall accuracy, only a matrix (certified) reference material does.

Table 7 – Reference materials used by the participants as reported in the questionnaire

| Part Nr | Use of reference material? | Used for validation? | Used for calibration? | Which reference material? |
|---------|----------------------------|----------------------|-----------------------|---|
| 058 | yes | yes | yes | |
| 082 | yes | yes | no | IQC-026 Combined Quality Control Standard for checking the leach solution |
| 150 | yes | yes | yes | Perkin Elmer ICP |
| 176 | yes | yes | no | |
| 233 | yes | yes | yes | Merck 1000 ppm Standards |
| 285 | yes | yes | yes | CA011A |
| 298 | yes | no | yes | |
| 332 | yes | yes | no | |
| 405 | yes | yes | no | In house material |
| 489 | yes | yes | yes | CRM solution |
| 522 | yes | yes | yes | CRM solution |
| 590 | yes | yes | yes | Atomic Absorption Standard |
| 661 | yes | yes | no | |
| 951 | yes | yes | yes | Reference solutions from Baker for AAS |

10 Conclusion

In the recent years many cases were brought to the attention of the public in relation to heavy metals in toys. However, the European standard dealing with this issue from an analytic point of view, the EN 71-3:1994, is known to give a high variation in results. This standard is currently in the process of being modified. It was however judged important to see how laboratories currently perform in the analysis of heavy metals in toys.

A former CRM consisting in a mild steel plate covered with alkyd paint was used and could be distributed to 39 participants from all over Europe and outside the EU. As expected the scatter of the results was large, but showed a close to normal distribution around the reference values for five out of eight heavy metals – antimony, barium, cadmium, chromium, lead. The participants' results tended to be lower than X_{ref} in the case of arsenic, mercury and selenium, elements known to be difficult to analyse. The lower results were mainly attributed to the sample preparation, these elements being very volatile and easy to loose.

One main issue of this exercise was the reporting of results with or without the analytical correction. The latter was introduced in EN 71-3:1994 as a mean to '*achieve consistent interpretation of results*' and is recommended to be applied when values are equal to or above the maximum limits set in the standard. The correction was applied by a number of participants, but only few of them reported that they have done so and it was very unclear

which results were corrected and which not. It is recommended that more care is taken to the reporting of results, specifying clearly whether a correction was applied or not.

Participants were also evaluated on how they interpreted their own results with regard to the acceptance of the test material on the market. The majority rejected the material correctly. However, almost one third made a wrong assessment:

- accepted the material,
- rejected but it is not justified by results,
- rejected it correctly but with very unreliable results.

This alarming situation should be improved.

Finally, it should be pointed out that the reported uncertainties were often very small. This is especially noticeable when taking into account the known variation of results for this type of analysis. The questionnaire shows that half of the laboratories estimate their uncertainty using data from measurements replicates (precision only). A reliable uncertainty estimate must take into account all significant sources of uncertainty.

11 Acknowledgements

The authors thank the Istituto di Ricerche e Collaudi, M. Masini S.r.l., the Istituto Italiano Sicurezza dei Giocattoli S.r.l., the LGC Ltd, and the Swedish National Testing and Research Institute for performing high precision analyses on the test material for the establishment of the assigned values. Anne-Mette Jensen is thanked for revising the manuscript.

The laboratories participating in this exercise, listed below are kindly acknowledged.

| Organisation | Country |
|--|----------------|
| CTIB-TCHN | BELGIUM |
| Testing centre ALMI TEST | BULGARIA |
| SGS CSTC Standards Technical Services Co. Ltd Shanghai Branch Testing Centre | CHINA |
| TÜV Rheinland (Shanghai) Co., Ltd | CHINA |
| Vyzkumny ustav organických syntez a.s. | CZECH REPUBLIC |
| LABTECH s.r.o. | CZECH REPUBLIC |
| Eurofins Miljø A/S | DENMARK |
| FORCE Technology | DENMARK |
| Finnish Customs Laboratory | FINLAND |
| INTERTEK | FRANCE |
| SGS CTS | FRANCE |
| Bureau Veritas Consumer Products Services France | FRANCE |
| Hansecontrol Testing Institute | GERMANY |
| TÜV Rheinland Produkt und Umwelt GmbH | GERMANY |

| Organisation | Country |
|--|----------------|
| LGA QualiTest GmbH | GERMANY |
| Eurofins AUA GmbH | GERMANY |
| INDIKATOR GmbH | GERMANY |
| Orga Lab GmbH | GERMANY |
| Ostthüringische Materialprüfgesellschaft für Textil und Kunststoffe mbH | GERMANY |
| SGS Hong Kong Ltd. | HONG KONG |
| The Standards Institution of Israel – Chemistry Section | ISRAEL |
| The Standards Institution of Israel – Metal Section | ISRAEL |
| ECO Certificazioni Spa | ITALY |
| Istituto di Ricerche e Collaudi M. Masini S.R.L. | ITALY |
| Nuovo Istituto Italiano Sicurezza dei Giocattoli SRL | ITALY |
| Luxcontrol S.A. | LUXEMBOURG |
| Textile Research Institute | POLAND |
| CATIM | PORTUGAL |
| CITEVE-Centro Tecnológico das Industrias têxteis e Vestuário de Portugal | PORTUGAL |
| Centro de Investigación y Control de la Calidad | SPAIN |
| STFI-Packforsk (Innventia) | SWEDEN |
| ALS Scandinavia AB | SWEDEN |
| SP Technical Research Institute of Sweden | SWEDEN |
| Laboratorium der Urkantone | SWITZERLAND |
| Kantonale Lebensmittelkontrolle | SWITZERLAND |
| Laboratorio cantonale | SWITZERLAND |
| Harlan Laboratories Ltd | SWITZERLAND |
| SQTS | SWITZERLAND |
| Kantonales Labor Zürich | SWITZERLAND |
| LGC Limited | UNITED KINGDOM |
| Intertek | UNITED KINGDOM |
| SATRA Technology Centre | UNITED KINGDOM |
| Bureau Veritas Consumer Products Services UK Ltd | UNITED KINGDOM |

Abbreviations

| | |
|----------|---|
| AAS | Atomic Absorption Spectroscopy |
| AC | Analytical Correction |
| AMC | Analytical Methods Committee of the Royal Society of Chemistry |
| BCR | Community Bureau of Reference |
| CEN | European Committee for Standardization |
| CITAC | Co-operation for International Traceability in Analytical Chemistry |
| CRM | Certified Reference Material |
| EA | European Co-operation for Accreditation |
| EC | European Commission |
| EN | European Standard |
| EU | European Union |
| EURACHEM | A focus for Analytical Chemistry in Europe |
| FN / FP | False Negative / False Positive |
| GUM | Guide to the Expression of Uncertainty in Measurement |
| ICP-MS | Inductively-Coupled Plasma Mass Spectrometry |
| ICP-OES | Inductively-Coupled Plasma Optical Emission Spectrometry |
| ILC | Interlaboratory Comparison |
| IMEP | International Measurement Evaluation Programme |
| IRMM | Institute for Reference Materials and Measurements |
| ISO | International Organisation for Standardisation |
| IUPAC | International Union for Pure and Applied Chemistry |
| JRC | Joint Research Centre |
| NANDO | New Approach Notified and Designated Organisations |
| MoU | Memorandum of Understanding |
| RSD | Relative Standard Deviation |
| SP | Swedish National Testing and Research Institute |
| SWEDAC | Swedish Board for Accreditation and Conformity Assessment |
| TN / TP | True Negative / True Positive |

References

- [1] EN 71-3:1994, Safety of toys - Part 3: Migration of certain elements (1994), European Committee for Standardisation (CEN), ICS 97.200.50
- [2] ISO 13528, Statistical Methods for Use in Proficiency Testing by Interlaboratory Comparisons (2005), issued by International Organisation for Standardisation
- [3] Council Directive 88/378/EEC of 3 May 1988 on the approximation of the laws of the Member States concerning the safety of toys (1988), issued by European Commission, Official Journal of the European Union, L 187
- [4] Quevauviller P (2001) Certified reference materials for the quality control of inorganic analyses of manufactured products (glass, polymers, paint coatings). *TrAC - Trends in Analytical Chemistry* 20(8): 446-456
- [5] Roper P, Walker R, Quevauviller P (2000) Collaborative study for the quality control of trace element determinations in paint coatings. Part 2. Certification of alkyd resin paint reference materials for the migratable contents of trace elements (CRMs 620 and 623). *Fresenius' Journal of Analytical Chemistry* 366(3): 289-297
- [6] Thompson M, Ellison SLR, Wood R (2006) The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories: (IUPAC technical report). *Pure and Applied Chemistry* 78(1): 145-196
- [7] Pauwels J, Van Der Yeen A, Lamberty A, Schimmel H (2000) Evaluation of uncertainty of reference materials. *Accreditation and Quality Assurance* 5(3): 95-99
- [8] Pauwels J, Lamberty A, Schimmel H (1998) The determination of the uncertainty of reference materials certified by laboratory intercomparison. *Accreditation and Quality Assurance* 3(5): 180-184

- [9] Linsinger T (2005) Comparison of a measurement result with the certified value. Application Note 1. European Commission - Joint Research Centre - Institute for Reference Materials and Measurements (IRMM), www.erm-crm.org
- [10] ISO (1993) Guide to the Expression of Uncertainty in Measurement (GUM). International Organisation for Standardisation, Geneva - Switzerland
- [11] Robust statistics: a method of coping with outliers (2001). AMC Technical Brief issued by the Statistical Subcommittee of the Analytical Methods Committee (AMC) of the Royal Society of Chemistry
- [12] Representing data distributions with Kernel density estimates (2006). AMC Technical Brief issued by the Statistical Subcommittee of the Analytical Methods Committee (AMC) of the Royal Society of Chemistry
- [13] Quantifying Uncertainty in Analytical Measurement (2000). Eurachem/CITAC, <http://www.eurachem.org>

Annexes

| | |
|---|----|
| Annex 1 : Invitation to EA to nominate laboratories | 31 |
| Annex 2 : Invitation to expert laboratories..... | 32 |
| Annex 3 : Invitation to notified bodies from NANDO list..... | 33 |
| Annex 4 : Publication on IRMM website..... | 34 |
| Annex 5 : Sample accompanying letter..... | 35 |
| Annex 6 : Questionnaire | 36 |
| Annex 7 : 'Confirmation of receipt' form | 37 |
| Annex 8 : Homogeneity study | 38 |
| Annex 9 : Reference values and their associated uncertainties | 39 |
| Annex 10 : Results for Antimony | 40 |
| Annex 11 : Results for Arsenic | 42 |
| Annex 12 : Results for Barium..... | 44 |
| Annex 13 : Results for Cadmium | 46 |
| Annex 14 : Results for Chromium | 48 |
| Annex 15 : Results for Lead | 50 |
| Annex 16 : Results for Mercury | 52 |
| Annex 17 : Results for Selenium..... | 54 |
| Annex 18 : Kernel densities | 56 |
| Annex 19 : Summary of scorings and results interpretation | 57 |
| Annex 20 : Experimental details derived from questionnaire | 58 |

Annex 1 : Invitation to EA to nominate laboratories

Registration of participants is open until **22 May 2009**. Distribution of the samples is foreseen for the end of May-early June 2009. The deadline for the reporting of results is then scheduled for 30 June 2009. In order to register, laboratories must:

1. **Enter** their details online:
<https://irmm.jrc.ec.europa.eu/ile/ileRegistration.do?selComparison=239>
When accessing this page you might be confronted with a Certificate Error page, please press the continue button to proceed with the registration.
2. **Print** the completed form when the system asks to do so.
Do not forget to clearly indicate on the printed form that you have been appointed by the European Cooperation for Accreditation to take part in this exercise.
3. **Send** the printout to both the IMEP-24 and the EA-IMEP-24 coordinators:

IMEP-24 coordinator
Ms. Ines Baer
Fax +32 14 571865
E-mail jrc-irmm-imep@ec.europa.eu

EA-IMEP-24 coordinator
Mrs. Annika Norling
Fax +46 0 791 89 29
E-mail Annika.norling@swedac.se

Please contact me if you have any questions or comments. We are looking forward to our cooperation!

With kind regards



Ines Baer
IMEP-24 Coordinator

2

EUROPEAN COMMISSION
JOINT RESEARCH CENTRE

Institute for Reference Materials and Measurements



Geel, 6 April 2009
JRC.D04/TBa/ve/ARES(2009)/62600

SWEDAC
Annika Norling
Box 2231
10315 Stockholm
SWEDEN

Dear Annika,

Intercomparison for heavy metals in toys according to EN71-3:1994

The Institute for Reference Materials and Measurements (IRMM) organises an interlaboratory comparison for the determination of the eight heavy metals whose safety limits are set out by the EU Toy Safety Directive (88/578/EEC) and specified in the harmonised European Standard EN71-3:1994. The concerned elements are lead, arsenic, antimony, barium, selenium, mercury, chromium and cadmium. The sample matrix is a mild steel plate coated with alkyd resin paint.

In the frame of the EA-IRMM collaboration agreement, IRMM kindly invites EA to nominate laboratories for free participation. These laboratories must be involved in toy safety evaluation and be familiar with the above mentioned standard, since it will be the method to be applied to the sample. They also should hold (or be in the process of obtaining) an accreditation for this type of measurement.

I suggest that you forward this invitation to the national EA accreditation bodies for their consideration. There is a maximum number of 65 samples at your disposal, i.e. ca. 2-3 nominees per country.

Confidentiality of the participants and their results towards third parties is guaranteed. However, IMEP will disclose details of the participants that have been nominated by EA to the EA working group for ILCS in Testing. The EA accreditation bodies may wish to inform the nominees of this disclosure.

Refisewweg 111, B-2440 Geel - Belgium. Telephone: (32-14) 571 211. <http://irmm.jrc.ec.europa.eu>
Telephone: direct line (32-14) 571 602. Fax: (32-14) 571 605.
E-mail: jrc-irmm-imep@ec.europa.eu

Annex 2 : Invitation to expert laboratories

- Print the completed form when the system asks to do so.
- Send the printout to the IMEP-24 coordinator:

IMEP-24 coordinator
 Ms. Ines Baer
 Fax +32 14 571865
 E-mail jrc-irmm-imep@ec.europa.eu

Distribution of the samples is foreseen for the end of April 2009.

Don't hesitate to contact me if you have any questions or comments. We are looking forward to our cooperation!

With kind regards



Ines Baer
 IMEP-24 Coordinator

EUROPEAN COMMISSION
 JOINT RESEARCH CENTRE
 Institute for Reference Materials and Measurements



Geel, 25 March 2009
 JRC:D04/IBa/ve/ARES(2009/)

[Name]
 [Institution]
 [Address]
 [Address]
 [Postal Code]
 [Country]

Dear [Name],

Intercomparison for heavy metals in toys according to EN71-3:1994

You have been contacted by Johannes van de Kreeke concerning the above mentioned intercomparison project. Please take note that because his leaving our institute, the project was taken over by me, Ines Baer. You have agreed to participate in the proficiency test and that your results could be used for the establishment of the reference value. Please be reminded that in that case a high precision analysis is expected of you and not a routine analysis. It goes without saying that your participation would be free of charge.

As a reminder, the interlaboratory comparison concerns the determination of the eight heavy metals whose safety limits are set out by the EU Toy Safety Directive (88/378/EEC) and specified in the harmonised European Standard EN71-3:1994. The elements are lead, arsenic, antimony, barium, selenium, mercury, chromium and cadmium. The sample matrix is a mild steel plate coated with alkyd resin paint.

I would like to inform you that the homogeneity and stability study has been realised and the results were of the quality that it was decided to proceed with the project. The exercise was launched a few days ago and the registration interface is open now. I therefore kindly invite you to register using the following link :

<https://irmm.jrc.ec.europa.eu/jrc/ileRegistration.do?setComparison=239>

When accessing this page you might be confronted with a Certificate Error page, please press the continue button to proceed with the registration.

References 111, B 2440 Geel - Belgium, Telephone: (32-14) 571 211, <http://irmm.jrc.ec.europa.eu>
 Telephone direct line (32-14) 571 862, Fax: (32-14) 571 866.
 E-mail: jrc-irmm-imep@ec.europa.eu

Annex 3 : Invitation to notified bodies from NANDO list

IMEP-24 Interlaboratory comparison for heavy metals in toys
BAER Ines (JRC-GEEL)

This message was sent with High importance.

To: BAER Ines (JRC-GEEL)
Cc: JRC IRMM IMEP

Dear Madam or Sir,

My name is Ines Baer and I am working at the European Commission - Institute for Reference Materials and Measurements (IRMM), more specifically on the organisation of Interlaboratory comparisons (ILC) in the frame IMEP, the International Measurement Evaluation Programm.

The reason I am contacting you is that your institute is listed under the European Commission Directive 88/378/EEC – Safety of toys as being responsible for this type of examination. We are currently organising an interlaboratory comparison exercise for the determination of the eight heavy metals whose safety limits are set out by this same directive and specified in the harmonised European Standard EN71-3:1994.

We would therefore very much appreciate your participation in this ILC and invite you to register. You will find all relevant information on our website http://irmm.jrc.ec.europa.eu/html/interlaboratory_comparisons/imep/imep-24/index.htm.
The deadline for registration is the 22 May 2009.

Feel free to contact me in case of any further questions.

Thank you for considering this invitation
Looking forward to hearing from you

Kind regards
Ines Baer
IMEP-24-Coordinator

Ines Baer
EC-JRC-IRMM
☎ +32 (0)14 57 16 82
☎ +32 (0)14 57 18 65
✉ jrc-irmm-imep@ec.europa.eu

Disclaimer: The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission.

Annex 4 : Publication on IRMM website

The screenshot shows the IRMM website page for IMEP-24. The browser address bar shows the URL: http://irmm.jrc.ec.europa.eu/html/interlaboratory_comparisons/mep/imep-24/index.htm. The page header includes the European Commission logo and the text "Joint Research Centre Institute for Reference Materials and Measurements". The main content area features the IMEP logo and the title "IMEP-24 Heavy metals in toys". Below the title, there is a paragraph describing the interlaboratory comparison for the determination of eight heavy metals (lead, arsenic, antimony, barium, selenium, mercury, chromium, and cadmium) in toys, specifically a mild steel plate coated with alkyl resin paint. The text mentions that the measurement results were evaluated against traceable reference values and that participating laboratories benchmarked their performance with peers throughout Europe. A "Schedule" section follows, containing a table with three columns: Registration, Sample Dispatch, and Reporting of results. The table details the registration status (closed), the sample dispatch date (June 2009), the reporting deadline (30/06/2009), the reference values date (July 2009), the provision of individual certificates (not foreseen), and the report to participants date (first half of October 2009). The page footer indicates the last update was on 17/09/2009. The sidebar on the right contains a "News archive" section with links to "Reference materials and measurements", "Food, biotechnology and health", "Environmental analysis", and "Nuclear research". Logos for ERM, IRMM, and CRL are also visible in the sidebar.

IMEP-24 Heavy metals in toys

IMEP-24 was an interlaboratory comparison for the determination of the eight heavy metals whose safety limits are set out by the EU Toy Safety Directive (88/378/EEC) and specified in the harmonised European Standard EN71-3:1994. The concerned elements were lead, arsenic, antimony, barium, selenium, mercury, chromium and cadmium. The sample matrix was a mild steel plate coated with alkyl resin paint. The laboratories were asked to apply the sample preparation method described in EN71-3:1994, but otherwise to follow their routine procedure.

Characteristic for this study was its high metrological standard. This may be of particular interest to those who were involved in analysis of trace elements in toys. The measurement results were evaluated against traceable reference values where possible, and uncertainty statements when provided by participants were included in the evaluation. Participating in IMEP-24 also helped laboratories to benchmark their performance with peers throughout Europe.

Schedule

| Registration | Sample Dispatch | Reporting of results |
|------------------|-------------------------|----------------------------|
| closed | June 2009 | deadline 30/06/2009 |
| Reference values | Individual certificates | Report to participants |
| July 2009 | not foreseen | first half of October 2009 |

Last Update 17/09/2009

Annex 5 : Sample accompanying letter

all paint is scrapped off the plate and carefully homogenised by mixing before sieving through a 500 micron mesh, taking care not to lose the finer particles due to effects of static electricity.

Result reporting

Please report your results online at <https://irmm.jrc.ec.europa.eu/office/Reporting.do>. Your personal code for access is: «Part_Key». The website will guide you through the reporting procedure. Please enter for each parameter:

- In the spaces allocated for "measurement 1" to "measurement 3": the measurement results and the technique you used, but not the uncertainty for each individual measurement
- In the space allocated for "measurement 4": the mean of the results, the technique you used, and the uncertainty of the mean, including the expansion factor.

The results should be reported in the same form (e.g. number of significant figures) as those normally reported to the customer. After reporting your results, please complete the subsequent questionnaire. Do not forget to submit and confirm when required. Directly after completing the measurement results and questionnaire parts, you will be prompted to print the completed report form. Please do so, sign this paper version, put the stamp of your institute and return it to IRMM by fax or e-mail. Check your results carefully for any errors before submission, since this is your definitive confirmation.

Deadline for reporting the results is 3 July 2009.

Result assessment

Your results will be assessed against certified reference values and their uncertainties where available. The scores applied by IMEP are z and zeta scores cf. ISO 13528 [1]. They will be used for each of the eight heavy metals.

Your participation in this project is greatly appreciated. If you have any remaining questions, please contact me by e-mail: jrc-irmm-imep@ec.europa.eu.

Please keep in mind that collusion is contrary to professional scientific conduct and serves only to nullify the benefits of proficiency tests to customers, accreditation bodies, and analysts alike.

With kind regards



Dr. Ines Baer
IMEP-24 Coordinator

[1] ISO 13528:2005: Statistical Methods for Use in Proficiency Testing by Interlaboratory Comparisons



Geel, 27 May 2009
JRC.D04/IBa/ve/ARES(2009)/110380

«Title» «Firstname» «Surname»
«Organisation»
«Department»
«Address»
«Address2»
«Address3»
«Zip» «Town»
«Country»

Subject: IMEP-24 - Intercomparison for eight heavy metals in toys

Dear «Title» «Surname»,

Thank you very much for participating in IMEP-24, an interlaboratory comparison for the determination of the eight heavy metals whose safety limits are set out by the EU Toy Safety Directive (88/378/EEC) and specified in the harmonised European Standard EN71-3:1994. Please find further relevant information regarding your participation below.

Packages and content

This parcel contains:

- a) the test material consisting of a 10 x 15 cm coated steel plate,
- b) "Confirmation of receipt" form,
- c) this accompanying letter.

Please check whether the sample has remained undamaged during transport, and confirm undamaged receipt with the "Confirmation of receipt" form, and please send it back by fax to +32 14 57 1865 or email to the project coordinator.

Please store the sample in the dark at ≤ 18 °C until analysis.

Measurands and matrix

Measurands are the migrated concentrations of lead, arsenic, antimony, barium, selenium, mercury, chromium and cadmium to be determined as described in EN71-3:1994. The sample matrix is an alkyd resin paint deposited on a mild steel plate.

Preparing the samples for analysis

The procedures you follow for this exercise should resemble as close as possible those that you use in routine sample analysis. The amount of paint deposited on the plate is sufficient so that one gram of paint can be scrapped off for analysis. It is recommended that

Releaseweg 111, B-2440 Geel - Belgium. Telephone: (32-44) 571 211. <http://irmm.jrc.ec.europa.eu>
Telephone: direct line (32-14) 571 662. Fax: (32-14) 571 660.
E-mail: jrc-irmm-imep@ec.europa.eu

Annex 6 : Questionnaire

2. Uncertainty related questions

2.1. What is the level of confidence reflected by the coverage (k) factors stated above? (in %)

2.2. What is the basis of your uncertainty estimate (multiple answers are possible)

a) uncertainty budget according to ISO-GUM
 b) known uncertainty of the standard method
 c) uncertainty of the method as determined during in-house validation
 d) measurement of replicates (i.e. precision)
 e) expert questionnaire
 f) use of intercomparison data
 g) other

2.2.1. If other, please specify, _____

2.3. Do you usually provide an uncertainty statement to your customers for this type of analysis?

no
 yes

3. Quality related questions

3.1. Does your laboratory have a quality system in place?

no
 yes

3.1.1. If Yes, which:

a) ISO 17025
 b) ISO 9000 series
 c) Other _____

3.1.1.1. If other, please specify, _____

3.1.2. If yes, are you accredited?

No
 Yes

3.1.2.1. If yes, by which Accreditation Body have you been accredited? _____

3.2. Does your laboratory carry out this type of analysis (as regards the parameters, matrix and methods)

no
 yes

3.2.1. If Yes, please estimate the number of samples :

a) 0-50 samples per year
 b) 50-250 samples per year
 c) 250-1000 samples per year
 d) more than 1000 samples per year

3.3. Does your laboratory take part in an interlaboratory comparison for this type of analysis on a regular basis ?

no
 yes

3.3.1. If yes, which one(s): _____

3.4. Does your laboratory use a reference material for this type of analysis?

no
 yes

3.4.1. If yes, is the material used for the validation of procedures ?

no
 yes

3.4.2. If yes, is the material used for calibration of instruments ?

no
 yes

3.4.3. If yes, which one(s): _____

4. In view of your results would you accept or reject the entrance of the material on the market ?

Accept
 Reject

4.1. Why? _____

5. Do you have any comments ? Please let us know:

1. Method related questions

1.1. Have you used the standard method EN 71-3:1994 for the analysis?

No
 Yes

1.1.1. If yes, have you deviated from that method?

No
 Yes

1.1.1.1. If yes, please specify: _____

1.1.2. If not, please describe briefly the method used: _____

1.2. If you have applied EN 71-3:1994, please answer the following questions.

1.2.1. State the sample amount used: _____

1.2.2. What pore size have you used for sieving ? _____

1.2.3. What type of shilling device have you used ? _____

1.2.4. Have you pre-heated the labware to 37 °C ?

No
 Yes

1.2.5. Did you need to adjust the pH ?

No
 Yes

1.2.6. What was the final pH ? _____

1.2.7. Specify the type and porosity of the membrane filter : _____

1.2.8. Was a centrifugation step necessary ?

No
 Yes

1.2.9. Did the ratio of solid to acid extractant exceed 1:50 ? _____

1.2.10. Did you analyse the samples on the day of processing ?

No
 Yes

1.2.10.1. If not, did you adjust the acid concentration to 1 mol/HCl ?

No
 Yes

1.2.11. Have you included base material in the analysis ?

No
 Yes

1.3. What are your detection limits for :

1.3.1. Arsenic _____

1.3.2. Barium _____

1.3.3. Cadmium _____

1.3.4. Chromium _____

1.3.5. Mercury _____

1.3.6. Lead _____

1.3.7. Antimony _____

1.3.8. Selenium _____

Annex 7 : 'Confirmation of receipt' form



EUROPEAN COMMISSION
JOINT RESEARCH CENTRE
Institute for reference materials and measurements
Isotope measurements

Annex to JRC.D04/IBa/ive/ARES(2009)/110360

«TITLE» «FIRSTNAME» «SURNAME»
«ORGANISATION»
«DEPARTMENT»
«ADDRESS»
«ADDRESS2»
«ADDRESS3»
«ZIP» «TOWN»
«COUNTRY»

IMEP-24

Migrated concentrations of Pb, As, Sn, Ba, Se, Hg, Cr and Cd in toys

Confirmation of receipt of the samples

*Please return this form at your earliest convenience.
This confirms that the sample package arrived.
In case the package is damaged,
please state this on the form and contact us immediately.*

ANY REMARKS

Date of package arrival

Signature

Please return this form to:
Dr Ines Baer

IMEP-24 Coordinator
EC-JRC-IRMM
Retieseweg 111
B-2440 GEEL, Belgium

Fax : +32-14-571865
e-mail : jrc-irmm-imep@ec.europa.eu

Retieseweg 111, B-2440 Geel - Belgium. Telephone: (32-14) 571 211. <http://irmm.jrc.ec.europa.eu>
Telephone: direct line (32-14) 571 682. Fax: (32-14) 571 865.


E-mail: jrc-irmm-imep@ec.europa.eu



Annex 8 : Homogeneity study

| Sample | Sb | | As | | Ba | | Cd | | Cr | | Pb | | Hg | | Se | |
|---|--------|------|--------|------|--------|-------|--------|-------|--------|------|--------|-------|--------|-------|--------|-------|
| | R1 | R2 | R1 | R2 | R1 | R2 | R1 | R2 | R1 | R2 | R1 | R2 | R1 | R2 | R1 | R2 |
| 1801 | 75.1 | 71.6 | 19.9 | 19.0 | 436.3 | 468.8 | 137.5 | 136.6 | 55.3 | 55.2 | 120.8 | 120.3 | 523.6 | 545.7 | 219.8 | 220.7 |
| 1578 | 71.2 | 87.5 | 20.8 | 18.6 | 422.2 | 459.0 | 137.5 | 139.5 | 54.2 | 60.3 | 118.8 | 131.6 | 467.7 | 527.0 | 208.8 | 212.8 |
| 1478 | 62.0 | 63.2 | 13.9 | 13.9 | 447.6 | 438.6 | 130.5 | 137.9 | 60.4 | 58.8 | 124.9 | 130.8 | 497.7 | 594.6 | 161.5 | 173.6 |
| 1140 | 72.1 | 80.2 | 15.7 | 16.9 | 485.6 | 474.7 | 131.6 | 134.7 | 57.9 | 59.4 | 137.5 | 140.7 | 450.9 | 531.0 | 180.0 | 180.0 |
| 1138 | 69.3 | 69.1 | 17.2 | 16.7 | 475.4 | 455.7 | 134.1 | 132.6 | 59.0 | 58.4 | 142.4 | 141.1 | 478.7 | 497.8 | 183.8 | 184.3 |
| 0733 | 73.4 | 73.7 | 15.5 | 16.7 | 456.8 | 483.7 | 109.6 | 136.0 | 59.2 | 68.6 | 120.8 | 126.1 | 380.0 | 408.0 | 169.3 | 169.2 |
| 0652 | 74.9 | 80.1 | 18.3 | 18.2 | 506.2 | 483.1 | 124.6 | 147.3 | 62.6 | 65.2 | 125.9 | 138.2 | 483.5 | 514.3 | 171.2 | 168.5 |
| 0508 | 84.7 | 84.2 | 17.5 | 19.3 | 503.9 | 497.5 | 125.4 | 139.5 | 64.1 | 66.6 | 119.5 | 131.6 | 552.8 | 608.2 | 188.2 | 190.3 |
| 0164 | 88.2 | 71.2 | 18.3 | 18.2 | 430.6 | 449.2 | 123.2 | 118.2 | 64.0 | 68.9 | 127.4 | 115.5 | 382.5 | 389.8 | 175.3 | 162.5 |
| 0163 | 83.1 | 85.8 | 19.8 | 20.0 | 468.8 | 470.5 | 139.8 | 131.0 | 61.9 | 64.7 | 123.2 | 137.2 | 485.8 | 505.4 | 166.1 | 166.4 |
| Mean | 76.0 | | 17.7 | | 465.7 | | 132.4 | | 61.2 | | 128.7 | | 491.3 | | 182.6 | |
| $\hat{\sigma}$ [%] | 30 | | 30 | | 15 | | 15 | | 15 | | 15 | | 25 | | 30 | |
| $\hat{\sigma}$ [mg kg ⁻¹] | 22.8 | | 5.3 | | 69.9 | | 19.9 | | 9.2 | | 19.3 | | 122.8 | | 54.8 | |
| Homogeneity test according to the IUPAC International Harmonised Protocol (mg ² kg ⁻²) | | | | | | | | | | | | | | | | |
| S_{an}^2 | 33.44 | | 0.604 | | 232.3 | | 79.25 | | 8.780 | | 43.71 | | 1271 | | 16.96 | |
| S_{sam}^2 | 28.71 | | 3.311 | | 360.3 | | 0 | | 10.06 | | 29.41 | | 3086 | | 356.9 | |
| σ_{all}^2 | 46.82 | | 2.543 | | 439.2 | | 35.47 | | 7.593 | | 33.55 | | 1357 | | 270.1 | |
| Crit value c | 121.8 | | 5.392 | | 1060 | | 146.7 | | 23.144 | | 107.2 | | 3835 | | 525.0 | |
| $S_{sam}^2 \leq c?$ | Passed | | Passed | | Passed | | Passed | | Passed | | Passed | | Passed | | Passed | |
| Homogeneity test according to ISO 13528 (mg kg ⁻¹) | | | | | | | | | | | | | | | | |
| $0.3 \hat{\sigma}$ | 6.843 | | 1.595 | | 20.96 | | 5.956 | | 2.756 | | 5.792 | | 36.844 | | 16.435 | |
| S_x | 6.740 | | 1.901 | | 21.83 | | 5.908 | | 3.802 | | 7.160 | | 61.00 | | 19.11 | |
| S_w | 5.782 | | 0.777 | | 15.24 | | 8.902 | | 2.963 | | 6.611 | | 35.65 | | 4.118 | |
| S_s | 5.358 | | 1.820 | | 18.98 | | 0 | | 3.172 | | 5.423 | | 55.55 | | 18.89 | |
| $S_s \leq 0.3 \hat{\sigma}$ | Yes | | No | | Yes | | Yes | | No | | Yes | | No | | No | |
| Test | Passed | | Failed | | Passed | | Passed | | Failed | | Passed | | Failed | | Failed | |

Annex 9 : Reference values and their associated uncertainties



— X_{exp} — Legal limit from EN 71-3:1994 (LL)
- - - $X_{exp} \pm U_{exp}$ - - - Max experimental value giving LL after applying analytical correction from EN 71-3:1994

If not present, the respective value is above the presented range.

Annex 10 : Results for Antimony


$X_{ref} = 83.0$ and $U_{ref} = 19.0$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | x4 | U _{lab} | k* | Mean | u _{lab} | Technique | z | zeta |
|---------|--------|--------|--------|------|------------------|-----|---------|------------------|-------------------------------|------|------|
| 004 | 102 | 97.3 | 95.0 | | 24.5 | 2 | 98.1 | 12.3 | ICP-OES | 0.6 | 1.0 |
| 024 | 117 | 116.8 | | | 35.1 | 1 | 116.9 | 35.1 | ICP-OES | 1.4 | 0.9 |
| 029 | 77.5 | 80.2 | 80 | | 12.7 | 2 | 79.2 | 6.4 | ICP-OES | -0.2 | -0.3 |
| 046 | 100 | 105 | 106 | | 1.4 | √3 | 104 | 0.8 | ICP-MS | 0.8 | 2.2 |
| 058 | | | | | 20 | 2 | 97 | 10 | XRF- Indicative Quantity Test | 0.6 | 1.0 |
| 074 | | | | | | 2 | 38.3075 | | ICP-OES | -1.8 | |
| 142 | 110 | 95.1 | 96.3 | | 16 | 2 | 100.5 | 8 | ICP-OES | 0.7 | 1.4 |
| 150 | 74 | 82 | 78 | | 8 | √3 | 78 | 4.6 | ICP-OES | -0.2 | -0.5 |
| 224 | 146 | 152 | | | 29.8 | √3 | 149 | 17.2 | HG-AAS | 2.7 | 3.4 |
| 233 | | | | | 5.59 | √3 | 92.7 | 3.23 | ICP-OES | 0.4 | 1.0 |
| 239 | 48.507 | 40.340 | 42.877 | | 8.575 | 2 | 43.908 | 4.288 | ICP-OES | -1.6 | -3.8 |
| 285 | 22.67 | 21.9 | 23.9 | | 0.82 | 1 | 22.82 | 0.82 | ICP-MS | -2.4 | -6.3 |
| 332 | 133.0 | 123.5 | 119.0 | | 0.22 | 2 | 125.2 | 0.11 | FAAS | 1.7 | 4.4 |
| 371 | 77.60 | 80.12 | 77.17 | | 11.74 | 60 | 78.30 | 0.20 | ICP-OES | -0.2 | -0.4 |
| 405 | 194 | 190 | | | | | 192 | | ICP-MS | 4.4 | |
| 421 | | | | | 16 | √3 | 79 | 9 | ICP-MS | -0.2 | -0.3 |
| 422 | 39.2 | 40.1 | 37.5 | | 3.89 | 2 | 38.9 | 1.95 | ICP-MS | -1.8 | -4.5 |
| 458 | 111 | 119 | 110 | | 10 | 2 | 113 | 5 | ICP-OES | 1.2 | 2.8 |
| 489 | 83 | 82 | 76 | | 7.6 | 2 | 80 | 3.8 | ICP-OES | -0.1 | -0.3 |
| 522 | 20 | 23 | 18 | | 5.1 | 2 | 20 | 2.6 | ICP-OES | -2.5 | -6.4 |
| 557 | 119 | 117 | 115 | | 12 | 2 | 117 | 6 | ICP-OES | 1.4 | 3.0 |
| 562 | 132.5 | 93.9 | 101.6 | | 14.2 | 2 | 109.3 | 7.1 | ICP-OES | 1.1 | 2.2 |
| 590 | | | | | | | 0.0021 | | FAAS | -3.3 | |
| 593 | 92 | 95 | 107 | | 2.4 | 0.4 | 98 | 6.0 | ICP-OES | 0.6 | 1.3 |
| 697 | 14 | 14 | 14 | | 1 | √3 | 14 | 1 | ICP-MS | -2.8 | -7.2 |
| 713 | 223 | 228 | 240 | | 6.8 | 2 | 230 | 3.4 | ICP-MS | 5.9 | 14.6 |
| 758 | 128.0 | 136.0 | 126.0 | | 10.583 | 2 | 130.0 | 5.292 | ICP-OES | 1.9 | 4.3 |
| 793 | 128 | 115 | 122 | | 11.0 | 2 | 122 | 5.5 | ICP-MS | 1.6 | 3.5 |
| 842 | 114.32 | 104.02 | 100.50 | | 0.038 | 2 | 106.28 | 0.019 | ICP-OES | 0.9 | 2.5 |
| 867 | 90 | 93 | 96 | | 5 | √3 | 93 | 3 | ICP-OES | 0.4 | 1.0 |
| 920 | 79.4 | 79.4 | 79.4 | | 15 | 2 | 79.4 | 8 | ICP-OES | -0.1 | -0.3 |
| 922 | 60.9 | 57.5 | 55 | 59.5 | 3.8 | 2 | 58.2 | 1.9 | FAAS | -1.0 | -2.6 |
| 924 | 40.7 | 40.9 | 39.5 | | 10.1 | 2 | 40.4 | 5.1 | ICP-OES | -1.7 | -4.0 |
| 936 | 81 | 83 | | | 5 | 2 | 82 | 3 | ETAAS | 0.0 | -0.1 |
| 951 | 21.28 | 20.10 | 21.78 | | 1.22 | √3 | 21.05 | 0.70 | Graphite Furnace AAS | -2.5 | -6.5 |
| 975 | 75.2 | 77.4 | | | 15 | 2 | 76.3 | 8 | ICP-MS | -0.3 | -0.6 |
| 994 | 135 | 137 | 128 | | | 2 | 133 | | ICP-OES | 2.0 | |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Antimony

Certified value: $X_{ref} = 83.0 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 19.0 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 11 : Results for Arsenic


$X_{ref} = 23.0$ and $U_{ref} = 6.3$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|-------|-------|-------|--------|------------|---------|--------|-------------------------------|------|------|
| 004 | 21.7 | 21.2 | 21.6 | 5.4 | 2 | 21.5 | 3.1 | ICP-OES | -0.2 | -0.3 |
| 024 | 16.65 | 17.56 | | 5.13 | 1 | 17.11 | 5.13 | ICP-OES | -0.9 | -1.0 |
| 029 | 13.9 | 14.4 | 13.8 | 3.10 | 2 | 14.0 | 1.55 | ICP-OES | -1.3 | -2.6 |
| 046 | 13 | 14 | 15 | 0.3 | $\sqrt{3}$ | 14 | 0.2 | ICP-MS | -1.3 | -2.9 |
| 058 | | | | 20 | 2 | 156 | 10 | XRF- Indicative Quantity Test | 19.3 | 12.7 |
| 074 | | | | | | 8.01305 | | ICP-OES | -2.2 | |
| 142 | 11.8 | 10.7 | 11.5 | 1.0 | 2 | 11.3 | 0.5 | ICP-OES | -1.7 | -3.7 |
| 150 | 13 | 14 | 15 | 1 | $\sqrt{3}$ | 14 | 1 | ICP-OES | -1.3 | -2.8 |
| 224 | 27.1 | 28.5 | | 5.56 | $\sqrt{3}$ | 27.8 | 3.21 | HG-AAS | 0.7 | 1.1 |
| 233 | | | | 11.61 | $\sqrt{3}$ | 19.5 | 6.70 | ICP-OES | -0.5 | -0.5 |
| 239 | 8.340 | 5.819 | 6.522 | 1.3044 | 2 | 6.894 | 0.6522 | ICP-OES | -2.3 | -5.0 |
| 285 | 4.07 | 3.77 | 4.14 | 0.16 | 1 | 3.99 | 0.16 | ICP-MS | -2.8 | -6.0 |
| 332 | 18.0 | 21.0 | 29.5 | 0.04 | 2 | 22.8 | 0.02 | FAAS | 0.0 | -0.1 |
| 371 | 19.27 | 19.13 | 19.98 | 2.92 | 60 | 19.46 | 0.05 | ICP-OES | -0.5 | -1.1 |
| 405 | 18 | 18 | | | | 18 | | ICP-MS | -0.7 | |
| 421 | | | | 2.86 | $\sqrt{3}$ | 14.3 | 1.65 | ICP-MS | -1.3 | -2.4 |
| 422 | 5.38 | 5.20 | 5.06 | 0.521 | 2 | 5.21 | 0.261 | ICP-MS | -2.6 | -5.6 |
| 458 | 22 | 23 | 21 | 2 | $\sqrt{3}$ | 22 | 1 | ICP-OES | -0.1 | -0.3 |
| 489 | 18 | 19 | 15 | 4.2 | 2 | 17 | 2.1 | ICP-OES | -0.8 | -1.5 |
| 522 | 3.2 | 3.6 | 3 | 0.6 | 2 | 3.3 | 0.3 | ICP-OES | -2.9 | -6.2 |
| 557 | 11.2 | 10.9 | 10.1 | 1.1 | 2 | 10.7 | 0.6 | ICP-OES | -1.8 | -3.8 |
| 562 | 14.5 | 12 | 14.5 | 1.04 | $\sqrt{3}$ | 13.7 | 0.60 | ICP-OES | -1.4 | -2.9 |
| 590 | <1 | | | | | | | FAAS | | |
| 593 | 23 | 22 | 21 | 1.25 | $\sqrt{3}$ | 22 | 0.72 | ICP-OES | -0.1 | -0.3 |
| 697 | <5 | <5 | <5 | | | | | ICP-OES | | |
| 713 | | | | | | 16 | | ICP-MS | -1.0 | |
| 758 | 13.24 | 13.61 | 13.90 | 0.662 | 2 | 13.58 | 0.331 | HG-AAS | -1.4 | -3.0 |
| 793 | 16.0 | 11.0 | 14.0 | 1.2 | 2 | 13.7 | 0.6 | ICP-MS | -1.4 | -2.9 |
| 842 | 26.92 | 26.91 | 27.44 | 0.038 | $\sqrt{3}$ | 27.09 | 0.022 | ICP-OES | 0.6 | 1.3 |
| 867 | 14 | 15 | 16 | 1 | $\sqrt{3}$ | 15 | 1 | ICP-OES | -1.2 | -2.5 |
| 920 | 21.8 | 21.8 | 21.8 | 15 | 2 | 21.8 | 8 | ICP-OES | -0.2 | -0.1 |
| 922 | <22 | <22 | <22 | | | | | FAAS | | |
| 924 | <5 | <5 | <5 | | | | | ICP-OES | | |
| 936 | 28 | 28 | | 6 | 2 | 28 | 3 | ETAAS | 0.7 | 1.1 |
| 951 | 5.91 | 6.48 | 6.24 | 0.43 | $\sqrt{3}$ | 6.21 | 0.25 | Graphite Furnace AAS | -2.4 | -5.3 |
| 975 | 24 | 25 | | 10.5 | $\sqrt{3}$ | 25 | 6.1 | ICP-MS | 0.2 | 0.2 |
| 994 | 30 | 32 | 30 | 0.4 | 2 | 31 | 0.2 | ICP-OES | 1.1 | 2.4 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Arsenic

Certified value: $X_{ref} = 23.0 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 6.3 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 12 : Results for Barium


$X_{ref} = 430$ and $U_{ref} = 50$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | x4 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|---------|---------|---------|-------|--------|------------|----------|--------|-------------------------------|------|-------|
| 004 | 518 | 519 | 513 | | 129 | 2 | 517 | 65 | ICP-OES | 1.3 | 1.3 |
| 024 | 407.2 | 485.9 | | | 67 | 1 | 446.6 | 67 | ICP-OES | 0.3 | 0.2 |
| 029 | 403.0 | 411.3 | 411.9 | | 40.9 | 2 | 408.7 | 20.5 | ICP-OES | -0.3 | -0.7 |
| 046 | 429 | 422 | 415 | | 4.9 | $\sqrt{3}$ | 422 | 2.8 | ICP-MS | -0.1 | -0.3 |
| 058 | <66 | | | | | | | | XRF- Indicative Quantity Test | | |
| 074 | | | | | | 2 | 485.2105 | | ICP-OES | 0.9 | |
| 142 | 528 | 517 | 505 | | 23 | 2 | 517 | 12 | ICP-OES | 1.3 | 3.1 |
| 150 | 295 | 305 | 315 | | 30 | $\sqrt{3}$ | 305 | 17 | ICP-OES | -1.9 | -4.1 |
| 224 | 526.63 | 527.42 | | | 105.4 | $\sqrt{3}$ | 527.03 | 60.9 | ICP-OES | 1.5 | 1.5 |
| 233 | | | | | 4.66 | $\sqrt{3}$ | 364.7 | 2.69 | ICP-OES | -1.0 | -2.6 |
| 239 | 244.205 | 185.984 | 204.748 | | 40.950 | 2 | 211.646 | 20.475 | ICP-OES | -3.4 | -6.8 |
| 285 | 97.99 | 97.44 | 102.4 | | 2.22 | 1 | 99.28 | 2.22 | ICP-MS | -5.1 | -13.2 |
| 332 | 375.5 | 392 | 386.5 | | 0.72 | 2 | 384.7 | 0.36 | FAAS | -0.7 | -1.8 |
| 371 | 414.85 | 412.85 | 410.0 | | 61.9 | 30 | 412.57 | 2.1 | ICP-OES | -0.3 | -0.7 |
| 405 | 639 | 602 | | | | | 621 | | ICP-MS | 3.0 | |
| 421 | | | | | 78 | $\sqrt{3}$ | 389 | 45 | ICP-MS | -0.6 | -0.8 |
| 422 | 233 | 237 | 230 | | 23.3 | 2 | 233 | 11.7 | ICP-MS | -3.0 | -7.1 |
| 458 | 530 | 525 | 529 | | 5 | 2 | 528 | 3 | ICP-OES | 1.5 | 3.9 |
| 489 | 416 | 422 | 429 | | 13 | 2 | 422 | 7 | ICP-OES | -0.1 | -0.3 |
| 522 | 112 | 111 | 100 | 107 | 13.3 | 2 | 108 | 6.7 | ICP-OES | -5.0 | -12.5 |
| 557 | 331 | 333 | 330 | | 33 | 2 | 331 | 17 | ICP-OES | -1.5 | -3.3 |
| 562 | 492.6 | 474.2 | 561.5 | | 61.1 | 2 | 509.4 | 30.6 | ICP-OES | 1.2 | 2.0 |
| 590 | <1 | | | | | | | | FAAS | | |
| 593 | 481 | 463 | 505 | | 1.85 | $\sqrt{3}$ | 483 | 1.07 | ICP-OES | 0.8 | 2.1 |
| 697 | <100 | <100 | <100 | <100 | | | | | ICP-OES | | |
| 713 | 507 | 506 | 510 | | 3 | 2 | 508 | 2 | ICP-MS | 1.2 | 3.1 |
| 758 | 371.0 | 372.0 | 364.5 | | 8.145 | 2 | 369.2 | 4.073 | ICP-OES | -0.9 | -2.4 |
| 793 | 388 | 341 | 369 | | 32 | 2 | 366 | 16 | ICP-MS | -1.0 | -2.2 |
| 842 | 443.59 | 452.56 | 437.95 | | 0.038 | 2 | 444.70 | 0.019 | ICP-OES | 0.2 | 0.6 |
| 867 | 410 | 410 | 430 | | 21 | $\sqrt{3}$ | 417 | 12 | ICP-OES | -0.2 | -0.5 |
| 920 | 400.8 | 404.8 | 400.8 | | 15 | 2 | 402.1 | 8 | ICP-OES | -0.4 | -1.1 |
| 922 | 376.5 | 480.6 | 421.1 | 460.8 | 29 | 2 | 434.8 | 15 | FAAS | 0.1 | 0.2 |
| 924 | 181 | 187 | 185 | | 46 | 2 | 184 | 23 | ICP-OES | -3.8 | -7.2 |
| 927 | 448 | 489 | | | | | 469 | | ICP-OES | 0.6 | |
| 936 | 454 | 431 | | | 30 | 2 | 443 | 15 | ETAAS | 0.2 | 0.4 |
| 951 | 422.63 | 449.23 | 418.37 | | 33.19 | $\sqrt{3}$ | 430.08 | 19.16 | Graphite Furnace AAS | 0.0 | 0.0 |
| 975 | 495 | 488 | | | 91.5 | 2 | 492 | 45.8 | ICP-MS | 1.0 | 1.2 |
| 994 | 508 | 490 | 479 | | 15.8 | 2 | 492 | 7.9 | ICP-OES | 1.0 | 2.4 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Barium

Certified value: $X_{ref} = 430 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 50 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 13 : Results for Cadmium


$X_{ref} = 117$ and $U_{ref} = 21$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | x4 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|--------|--------|--------|-------|--------|------------|----------|-------|-------------------------------|-------|-------|
| 004 | 179 | 186 | 193 | | 46.5 | 2 | 186 | 23.3 | ICP-OES | 3.9 | 2.7 |
| 024 | 137.4 | 136.6 | | | 20.5 | 1 | 137.0 | 20.5 | ICP-OES | 1.1 | 0.9 |
| 029 | 163.0 | 165.2 | 162.2 | | 16.4 | 2 | 163.5 | 8.2 | ICP-OES | 2.6 | 3.5 |
| 046 | 120 | 118 | 119 | | 0.7 | $\sqrt{3}$ | 119 | 0.4 | ICP-MS | 0.1 | 0.2 |
| 058 | | | | | 61 | 2 | 4392 | 31 | XRF- Indicative Quantity Test | 243.6 | 132.5 |
| 074 | | | | | | | 195.7942 | | ICP-OES | 4.5 | |
| 142 | 235 | 298 | 286 | | 66 | 2 | 273 | 33 | ICP-OES | 8.9 | 4.5 |
| 150 | 120 | 110 | 130 | | 12 | $\sqrt{3}$ | 120 | 7 | ICP-OES | 0.2 | 0.2 |
| 224 | 207.49 | 199.53 | | | 40.8 | $\sqrt{3}$ | 203.51 | 23.6 | ICP-OES | 4.9 | 3.4 |
| 233 | | | | | 4.81 | $\sqrt{3}$ | 139.2 | 2.78 | ICP-OES | 1.3 | 2.0 |
| 239 | 74.374 | 50.770 | 58.558 | | 11.712 | 2 | 61.234 | 5.856 | ICP-OES | -3.2 | -4.6 |
| 285 | 21.21 | 19.63 | 22.24 | | 1.07 | 1 | 21.03 | 1.07 | ICP-MS | -5.5 | -9.1 |
| 298 | 305 | 307 | | | 2 | 2 | 306 | 1 | FAAS | 10.8 | 17.9 |
| 332 | 124.0 | 111.5 | 102.0 | | 0.22 | 2 | 112.5 | 0.11 | FAAS | -0.3 | -0.4 |
| 371 | 109.65 | 109.3 | 109.6 | | 16.43 | 30 | 109.5 | 0.55 | ICP-OES | -0.4 | -0.5 |
| 405 | 174 | 163 | | | | | 169 | | ICP-MS | 2.9 | |
| 421 | | | | | 30 | 2 | 152 | 15 | ICP-MS | 2.0 | 1.9 |
| 422 | 72.9 | 68.4 | 74.3 | | 7.19 | 2 | 71.9 | 3.60 | ICP-MS | -2.6 | -4.1 |
| 458 | 179 | 173 | 175 | | 6 | 2 | 176 | 3 | ICP-OES | 3.3 | 5.4 |
| 489 | 143 | 140 | 140 | | 3.5 | 2 | 141 | 1.8 | ICP-OES | 1.4 | 2.3 |
| 522 | 23 | 20 | 18 | | 4.6 | 2 | 20 | 2.3 | ICP-OES | -5.5 | -9.0 |
| 557 | 74.4 | 74.3 | 73.4 | | 7.4 | 2 | 74.0 | 3.7 | ICP-OES | -2.4 | -3.9 |
| 562 | 143.9 | 125.8 | 136.4 | | 4.33 | 2 | 135.4 | 2.17 | ICP-OES | 1.0 | 1.7 |
| 590 | | | | | | | 0.00238 | | FAAS | -6.7 | |
| 593 | 182 | 173 | 191 | | 1.85 | 0.7 | 182 | 2.64 | ICP-OES | 3.7 | 6.0 |
| 697 | 14 | 13 | 14 | | 1 | $\sqrt{3}$ | 14 | 1 | ICP-OES | -5.9 | -9.8 |
| 713 | 127 | 130 | 133 | | 4 | 2 | 130 | 2 | ICP-MS | 0.7 | 1.2 |
| 758 | 130.0 | 129.0 | 130.5 | | 1.528 | 2 | 129.8 | 0.764 | ICP-OES | 0.7 | 1.2 |
| 793 | 130 | 118 | 142 | | 11 | 2 | 130 | 6 | ICP-MS | 0.7 | 1.1 |
| 842 | 131.91 | 143.88 | 149.35 | | 0.038 | 2 | 141.71 | 0.019 | ICP-OES | 1.4 | 2.4 |
| 867 | 140 | 140 | 150 | | 7 | $\sqrt{3}$ | 143 | 4 | ICP-OES | 1.5 | 2.3 |
| 920 | 142.9 | 140.9 | 142.9 | | 15 | 2 | 142.2 | 8 | ICP-OES | 1.4 | 2.0 |
| 922 | 188.3 | 208.1 | 208.1 | 193.2 | 11.6 | 2 | 199.4 | 5.8 | FAAS | 4.7 | 6.9 |
| 924 | 17.4 | 17.7 | 17.2 | | 4.4 | 2 | 17.4 | 2.2 | ICP-OES | -5.7 | -9.3 |
| 927 | 139 | 112 | | | | | 126 | | ICP-OES | 0.5 | |
| 936 | 171 | 163 | | | 7.1 | 2 | 167 | 3.6 | ETAAS | 2.8 | 4.5 |
| 951 | 117.57 | 130.13 | 143.90 | | 13.19 | $\sqrt{3}$ | 130.53 | 7.62 | Graphite Furnace AAS | 0.8 | 1.0 |
| 975 | 193 | 177 | | | 27.5 | 2 | 185 | 13.8 | ICP-MS | 3.9 | 3.9 |
| 994 | 202 | 193 | 191 | | 4.4 | 2 | 195 | 2.2 | ICP-OES | 4.5 | 7.3 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Cadmium

Certified value: $X_{ref} = 117 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 21 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 14 : Results for Chromium


$X_{ref} = 64$ and $U_{ref} = 24$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | x4 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|--------|--------|--------|------|-------|------------|---------|-------|-------------------------------|------|------|
| 004 | 82.3 | 83.2 | 80.8 | | 20.5 | 2 | 82.1 | 10.3 | ICP-OES | 1.9 | 1.1 |
| 024 | 60.29 | 61.5 | | | 9.13 | 1 | 60.90 | 9.13 | ICP-OES | -0.3 | -0.2 |
| 029 | 57.6 | 58.5 | 58.2 | | 5.8 | 2 | 58.1 | 2.9 | ICP-OES | -0.6 | -0.5 |
| 046 | 62 | 66 | 63 | | 1.3 | $\sqrt{3}$ | 64 | 0.8 | ICP-MS | 0.0 | 0.0 |
| 058 | <31 | | | | | | | | XRF- Indicative Quantity Test | | |
| 074 | | | | | | | 75.1126 | | ICP-OES | 1.2 | |
| 142 | 86.6 | 87.4 | 84.1 | | 3.5 | 2 | 86.0 | 1.8 | ICP-OES | 2.3 | 1.8 |
| 150 | 40 | 37 | 43 | | 4 | $\sqrt{3}$ | 40 | 2 | ICP-OES | -2.5 | -2.0 |
| 224 | 66.11 | 65.42 | | | 13.16 | $\sqrt{3}$ | 65.77 | 7.60 | ICP-OES | 0.2 | 0.1 |
| 233 | | | | | 9.0 | $\sqrt{3}$ | 50.2 | 5.2 | ICP-OES | -1.4 | -1.1 |
| 239 | 29.772 | 21.262 | 24.045 | | 4.809 | 2 | 25.026 | 2.405 | ICP-OES | -4.1 | -3.2 |
| 285 | 9.99 | 9.53 | 10.57 | | 0.43 | 1 | 10.03 | 0.43 | ICP-MS | -5.6 | -4.5 |
| 298 | 128 | 133 | | | 8 | 2 | 131 | 4 | ETAAS | 6.9 | 5.3 |
| 332 | 66.5 | 79.0 | 71.5 | | 0.14 | 2 | 72.3 | 0.07 | FAAS | 0.9 | 0.7 |
| 371 | 56.32 | 57.11 | 56.95 | | 8.52 | 30 | 56.79 | 0.28 | ICP-OES | -0.8 | -0.6 |
| 405 | 86 | 83 | | | | | 85 | | ICP-MS | 2.1 | |
| 421 | | | | | 12 | $\sqrt{3}$ | 61 | 7 | ICP-MS | -0.3 | -0.2 |
| 422 | 42.1 | 39.4 | 38.0 | | 3.98 | 2 | 39.8 | 1.99 | ICP-MS | -2.5 | -2.0 |
| 458 | 75 | 74 | 76 | | 2 | 2 | 75 | 1 | ICP-OES | 1.1 | 0.9 |
| 489 | 55 | 53 | 55 | | 2.3 | 2 | 54 | 1.2 | ICP-OES | -1.0 | -0.8 |
| 522 | 11 | 11 | 10 | 10 | 1.4 | 2 | 11 | 0.7 | ICP-OES | -5.6 | -4.5 |
| 557 | 52.2 | 53.3 | 53.5 | | 5.3 | 2 | 53.0 | 2.7 | ICP-OES | -1.1 | -0.9 |
| 562 | 64.9 | 49.4 | 60.8 | | 2.6 | 2 | 58.4 | 1.3 | ICP-OES | -0.6 | -0.5 |
| 590 | | | | | | | 0.00049 | | FAAS | -6.7 | |
| 593 | 65 | 62 | 68 | | 1.9 | 0.7 | 65 | 2.7 | ICP-OES | 0.1 | 0.1 |
| 697 | <10 | | | | | | | | ICP-OES | | |
| 713 | 71 | 71 | 79 | | 5.8 | 2 | 74 | 2.9 | ICP-MS | 1.0 | 0.8 |
| 758 | 77.0 | 76.8 | 78.0 | | 1.286 | 2 | 77.3 | 0.643 | ICP-OES | 1.4 | 1.1 |
| 793 | 69.2 | 54.2 | 58.7 | | 5.3 | 2 | 60.7 | 2.7 | ICP-MS | -0.3 | -0.3 |
| 842 | 70.34 | 70.91 | 70.92 | | 0.038 | 2 | 70.72 | 0.019 | ICP-OES | 0.7 | 0.6 |
| 867 | 57 | 58 | 58 | | 3 | $\sqrt{3}$ | 58 | 2 | ICP-OES | -0.7 | -0.5 |
| 920 | 49.6 | 49.6 | 49.6 | | 15 | 2 | 49.6 | 8 | ICP-OES | -1.5 | -1.0 |
| 922 | 85.7 | 85.2 | 87.2 | 85.7 | 5.1 | 2 | 86.0 | 2.6 | FAAS | 2.3 | 1.8 |
| 924 | 18.7 | 18.6 | 18.1 | | 4.6 | 2 | 18.5 | 2.3 | ICP-OES | -4.7 | -3.7 |
| 927 | 69 | 81 | | | | | 75 | | ICP-OES | 1.1 | |
| 936 | 65.6 | 62.2 | 59.7 | | 3.5 | 2 | 62.5 | 1.8 | FAAS | -0.2 | -0.1 |
| 951 | 50.23 | 51.86 | 55.20 | | 4.46 | $\sqrt{3}$ | 52.43 | 2.57 | Graphite Furnace AAS | -1.2 | -0.9 |
| 975 | 71.0 | 73.7 | | | 8.8 | 2 | 72.4 | 4.4 | ICP-MS | 0.9 | 0.7 |
| 994 | 72 | 69 | 70 | | 1.6 | 2 | 70 | 0.8 | ICP-OES | 0.7 | 0.5 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Chromium

Certified value: $X_{ref} = 64 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 24 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 15 : Results for Lead


$X_{ref} = 140$ and $U_{ref} = 40$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|--------|--------|--------|--------|------------|----------|-------|-------------------------------|------|------|
| 004 | 163 | 173 | 182 | 43 | 2 | 173 | 22 | ICP-OES | 1.6 | 1.1 |
| 024 | 101.1 | 87.47 | | 14.14 | 1 | 94.29 | 14.14 | ICP-OES | -2.2 | -1.9 |
| 029 | 155.6 | 155.3 | 153.8 | 15.5 | 2 | 154.9 | 7.8 | ICP-OES | 0.7 | 0.7 |
| 046 | 81 | 75 | 85 | 3.7 | $\sqrt{3}$ | 80 | 2.1 | ICP-MS | -2.8 | -3.0 |
| 058 | | | | 30 | 2 | 275 | 15 | XRF- Indicative Quantity Test | 6.4 | 5.4 |
| 074 | | | | | | 216.0752 | | ICP-OES | 3.6 | |
| 142 | 223 | 243 | 237 | 20 | 2 | 234 | 10 | ICP-OES | 4.5 | 4.2 |
| 150 | 115 | 120 | 110 | 12 | $\sqrt{3}$ | 115 | 7 | ICP-OES | -1.2 | -1.2 |
| 224 | 244.30 | 257.81 | | 50.2 | $\sqrt{3}$ | 251 | 29.0 | ICP-OES | 5.3 | 3.2 |
| 233 | | | | 5.13 | $\sqrt{3}$ | 115.9 | 2.96 | ICP-OES | -1.1 | -1.2 |
| 239 | 66.999 | 43.918 | 51.573 | 10.315 | 2 | 54.163 | 5.158 | ICP-OES | -4.1 | -4.2 |
| 285 | 15.43 | 15 | 15.78 | 0.32 | 1 | 15.40 | 0.32 | ICP-MS | -5.9 | -6.2 |
| 298 | 276 | 322 | | 46 | 2 | 299 | 23 | FAAS | 7.6 | 5.2 |
| 332 | 66.5 | 52.5 | 49.0 | 0.1 | 2 | 56.0 | 0.1 | FAAS | -4.0 | -4.2 |
| 371 | 188.6 | 184.3 | 193.6 | 28.32 | 30 | 188.8 | 0.94 | ICP-OES | 2.3 | 2.4 |
| 405 | 113 | 95 | | | | 104 | | ICP-MS | -1.7 | |
| 421 | | | | 39 | $\sqrt{3}$ | 195 | 23 | FAAS | 2.6 | 1.8 |
| 422 | 22.8 | 37.8 | 22 | 2.75 | 2 | 27.5 | 1.38 | ICP-MS | -5.4 | -5.6 |
| 458 | 134 | 123 | 109 | 25 | 2 | 122 | 13 | ICP-OES | -0.9 | -0.8 |
| 489 | 113 | 120 | 121 | 8.7 | 2 | 118 | 4.4 | ICP-OES | -1.0 | -1.1 |
| 522 | 20 | 18 | 17 | 3.1 | 2 | 18 | 1.6 | ICP-OES | -5.8 | -6.1 |
| 557 | 32.8 | 30.5 | 33.0 | 3.2 | 2 | 32.1 | 1.6 | ICP-OES | -5.1 | -5.4 |
| 562 | 134.1 | 77.6 | 132.7 | 5.4 | 2 | 114.8 | 2.7 | ICP-OES | -1.2 | -1.2 |
| 590 | | | | | | 0.00495 | | FAAS | -6.7 | |
| 593 | 189 | 182 | 174 | 2.1 | 0.70 | 182 | 3.0 | ICP-OES | 2.0 | 2.1 |
| 697 | <20 | | | | | | | ICP-OES | | |
| 713 | 49 | 41 | 41 | 5.0 | 2 | 44 | 2.5 | ICP-MS | -4.6 | -4.8 |
| 758 | 117.0 | 119.0 | 121.5 | 4.509 | 2 | 119.2 | 2.255 | ICP-OES | -1.0 | -1.0 |
| 793 | 72.3 | 79.3 | 128 | 8.1 | 2 | 93.2 | 4.1 | ICP-MS | -2.2 | -2.3 |
| 842 | 73.70 | 94.76 | 114.82 | 0.038 | 2 | 94.43 | 0.019 | ICP-OES | -2.2 | -2.3 |
| 867 | 110 | 120 | 120 | 6 | $\sqrt{3}$ | 117 | 3 | ICP-OES | -1.1 | -1.1 |
| 920 | 150.8 | 150.8 | 152.8 | 15 | 2 | 151.5 | 8 | ICP-OES | 0.5 | 0.5 |
| 922 | 173.4 | 153.6 | 163.5 | 9.8 | 2 | 163.5 | 4.9 | FAAS | 1.1 | 1.1 |
| 924 | 14.6 | 14.7 | 14.9 | 3.7 | 2 | 14.7 | 1.9 | ICP-OES | -6.0 | -6.2 |
| 936 | 151 | 160 | | 14 | 2 | 156 | 7 | ETAAS | 0.7 | 0.7 |
| 951 | 91.14 | 120.73 | 121.30 | 18.10 | $\sqrt{3}$ | 111.06 | 10.45 | Graphite Furnace AAS | -1.4 | -1.3 |
| 975 | 153 | 146 | | 26.5 | 2 | 150 | 13.3 | ICP-MS | 0.5 | 0.4 |
| 994 | 214 | 205 | 203 | 4.7 | 2 | 207 | 2.4 | ICP-OES | 3.2 | 3.3 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Lead

Certified value: $X_{ref} = 140 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 40 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 16 : Results for Mercury


$X_{ref} = 370$ and $U_{ref} = 127$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | x4 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|--------|--------|--------|-------|-------|------------|----------|-------|---------------------------------|------|------|
| 004 | 469 | 435 | 453 | | 181 | 2 | 452 | 90 | ICP-OES | 0.9 | 0.7 |
| 024 | 278.9 | 256.3 | | | 66.9 | 1 | 267.6 | 66.9 | ICP-OES | -1.1 | -1.1 |
| 029 | 278.9 | 291.2 | 293.4 | | 46.0 | 2 | 287.8 | 23.0 | ICP-OES | -0.9 | -1.2 |
| 046 | 215 | 208 | 196 | | 5.1 | $\sqrt{3}$ | 206 | 2.9 | ICP-MS | -1.8 | -2.6 |
| 058 | | | | | 95 | 2 | 4623 | 48 | XRF- Indicative Quantity Test | 46.0 | 53.5 |
| 074 | | | | | | | 140.7179 | | ICP-OES | -2.5 | |
| 142 | 25.2 | 23.2 | 22.0 | | 3.3 | 2 | 23.5 | 1.7 | CV-AFS | -3.7 | -5.4 |
| 150 | 290 | 270 | 250 | | 10 | $\sqrt{3}$ | 270 | 6 | ICP-OES | -1.1 | -1.6 |
| 224 | 970 | 1070 | 1020 | | 204 | $\sqrt{3}$ | 1020 | 118 | AMA analyzer | 7.0 | 4.9 |
| 233 | | | | | 6.45 | $\sqrt{3}$ | 434.9 | 3.72 | ICP-OES | 0.7 | 1.0 |
| 239 | 18.046 | 20.293 | | | 5.751 | 2 | 19.170 | 2.876 | CV-AAS | -3.8 | -5.5 |
| 285 | 92.63 | 108 | 99 | | 6.31 | 1 | 99.88 | 6.31 | ICP-MS | -2.9 | -4.2 |
| 298 | 2564 | 2612 | | | 68 | 2 | 2588 | 34 | Hg-Analyzer (AAS by Combustion) | 24.0 | 30.7 |
| 332 | 126.9 | 130.5 | 118.5 | | 0.22 | 2 | 125.3 | 0.11 | CV-AAS | -2.6 | -3.8 |
| 371 | 270.3 | 265.25 | 269.45 | | 40.25 | 50 | 268.33 | 0.81 | ICP-OES | -1.1 | -1.6 |
| 405 | 164 | 185 | | | | | 175 | | ICP-MS | -2.1 | |
| 421 | | | | | 5.6 | $\sqrt{3}$ | 28 | 3.2 | CV-AAS | -3.7 | -5.4 |
| 422 | 10.8 | 5.50 | 13.9 | | 1.01 | 2 | 10.1 | 0.51 | CV-AAS | -3.9 | -5.6 |
| 458 | 188 | 158 | 189 | | 35 | 2 | 178 | 18 | ICP-OES | -2.1 | -2.9 |
| 489 | 480 | 422 | 342 | | 138.6 | 2 | 415 | 69.3 | ICP-OES | 0.5 | 0.5 |
| 522 | 19 | 22 | 21 | | 2.6 | 2 | 21 | 1.3 | ICP-OES | -3.8 | -5.5 |
| 557 | 150 | 130 | 140 | | 21 | 2 | 140 | 11 | CV-AAS | -2.5 | -3.6 |
| 562 | 98.7 | 88 | 102.3 | | 15.4 | 2 | 96.3 | 7.7 | ICP-OES | -3.0 | -4.3 |
| 590 | | | | | | | 0.00086 | | FAAS | -4.0 | |
| 593 | 312 | 349 | 325 | | 5.05 | 0.5 | 329 | 10.10 | ICP-OES | -0.4 | -0.6 |
| 697 | 55 | 50 | 43 | | 1 | $\sqrt{3}$ | 49 | 1 | HG-AAS | -3.5 | -5.0 |
| 713 | 708 | 668 | 602 | | 54 | 2 | 659 | 27 | ICP-MS | 3.1 | 4.2 |
| 758 | 50.0 | 47.3 | 55.67 | | 8.569 | 2 | 51.0 | 4.285 | CV-AAS | -3.4 | -5.0 |
| 793 | 105 | 135 | 176.0 | | 20 | 2 | 139 | 10 | ICP-MS | -2.5 | -3.6 |
| 842 | 194.62 | 229.74 | 288.52 | | 0.038 | 2 | 237.63 | 0.019 | ICP-OES | -1.4 | -2.1 |
| 867 | 330 | 320 | 320 | | 17 | $\sqrt{3}$ | 323 | 10 | ICP-OES | -0.5 | -0.7 |
| 920 | 666.7 | 682.5 | 682.5 | | 15 | 2 | 677.2 | 8 | ICP-OES | 3.3 | 4.8 |
| 922 | 530.1 | 485.5 | 465.7 | 495.4 | 30.2 | 2 | 494.2 | 15.1 | FAAS | 1.3 | 1.9 |
| 924 | 46.9 | 46.3 | 45.9 | | 10.2 | 2 | 46.4 | 5.1 | FIMS | -3.5 | -5.1 |
| 936 | 150 | 134 | | | 38 | 2 | 142 | 19 | CV-AAS | -2.5 | -3.4 |
| 951 | 444.37 | 485.40 | 512.50 | | 78.03 | $\sqrt{3}$ | 480.76 | 45.05 | FIAS MHS | 1.2 | 1.4 |
| 975 | 222 | 231 | | | 60.5 | 2 | 227 | 30.3 | ICP-MS | -1.6 | -2.0 |
| 994 | 206 | 226 | 250 | | 3.6 | 2 | 227 | 1.8 | ICP-OES | -1.5 | -2.2 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Mercury

Reference value: $X_{ref} = 370 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 127 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 17 : Results for Selenium


$X_{ref} = 240$ and $U_{ref} = 60$; all values are given in ($mg\ kg^{-1}$)

| Part Nr | x1 | x2 | x3 | x4 | Ulab | k* | Mean | ulab | Technique | z | zeta |
|---------|---------|--------|--------|-------|--------|------------|----------|-------|-------------------------------|------|------|
| 004 | 206 | 216 | 226 | | 54 | 2 | 216 | 27 | ICP-OES | -0.3 | -0.6 |
| 024 | 150.9 | 137.2 | | | 43.2 | 1 | 144.1 | 43.2 | ICP-OES | -1.3 | -1.8 |
| 029 | 181.6 | 182.1 | 176.4 | | 36.0 | 2 | 180.0 | 18.0 | ICP-OES | -0.8 | -1.7 |
| 046 | 115 | 103 | 114 | | 2.7 | $\sqrt{3}$ | 111 | 1.6 | ICP-MS | -1.8 | -4.3 |
| 058 | | | | | 34 | 2 | 1416 | 17 | XRF- Indicative Quantity Test | 16.3 | 34.1 |
| 074 | | | | | | | 169.7634 | | ICP-OES | -1.0 | |
| 142 | 296 | 344 | 358 | | 65 | 2 | 333 | 33 | ICP-OES | 1.3 | 2.1 |
| 150 | 160 | 140 | 150 | | 15 | $\sqrt{3}$ | 150 | 9 | ICP-OES | -1.3 | -2.9 |
| 224 | 338 | 337 | | | 67.6 | $\sqrt{3}$ | 338 | 39.0 | HG-AAS | 1.4 | 2.0 |
| 233 | | | | | 4.78 | $\sqrt{3}$ | 174.2 | 2.76 | ICP-OES | -0.9 | -2.2 |
| 239 | 108.370 | 70.574 | 83.517 | | 16.703 | 2 | 87.487 | 8.352 | ICP-OES | -2.1 | -4.9 |
| 285 | 31.69 | 31 | 36 | | 2.21 | 1 | 32.9 | 2.21 | ICP-MS | -2.9 | -6.9 |
| 332 | 37.0 | 48.0 | 42.5 | | 0.08 | 2 | 42.5 | 0.04 | FAAS | -2.7 | -6.6 |
| 371 | 190.38 | 187.55 | 193.92 | | 28.59 | 60 | 190.62 | 0.48 | ICP-OES | -0.7 | -1.6 |
| 405 | 103 | 117 | | | | | 110 | | ICP-MS | -1.8 | |
| 421 | | | | | 33 | $\sqrt{3}$ | 167 | 19 | ICP-MS | -1.0 | -2.1 |
| 422 | 35.7 | 34.1 | 30.5 | | 3.34 | 2 | 33.4 | 1.67 | ICP-MS | -2.9 | -6.9 |
| 458 | 144 | 135 | 120 | | 24 | 2 | 133 | 12 | ICP-OES | -1.5 | -3.3 |
| 489 | 165 | 161 | 166 | | 5.3 | 2 | 164 | 2.7 | ICP-OES | -1.1 | -2.5 |
| 522 | 34 | 29 | 31 | | 4.3 | 2 | 31 | 2.2 | ICP-OES | -2.9 | -6.9 |
| 557 | 55.5 | 53.0 | 50.9 | | 5.3 | 2 | 53.1 | 2.7 | ICP-OES | -2.6 | -6.2 |
| 562 | 161 | 123.5 | 164.7 | | 19.9 | 2 | 149.7 | 10.0 | ICP-OES | -1.3 | -2.9 |
| 590 | | | | | | | 0.00033 | | FAAS | -3.3 | |
| 593 | 262 | 268 | 241 | | 2.2 | $\sqrt{3}$ | 257 | 1.3 | ICP-OES | 0.2 | 0.6 |
| 697 | <100 | | | | | | | | ICP-MS | | |
| 713 | 245 | 180 | 180 | | 31.6 | 2 | 202 | 15.8 | ICP-MS | -0.5 | -1.1 |
| 758 | 191.0 | 193.0 | 205.5 | | 15.716 | 2 | 196.5 | 7.858 | ICP-OES | -0.6 | -1.4 |
| 793 | 126 | 109 | 162 | | 11 | 2 | 132 | 6 | ICP-MS | -1.5 | -3.5 |
| 842 | 90.42 | 110.43 | 128.78 | | 0.038 | 2 | 109.88 | 0.019 | ICP-OES | -1.8 | -4.3 |
| 867 | 160 | 180 | 170 | | 9 | $\sqrt{3}$ | 170 | 5 | ICP-OES | -1.0 | -2.3 |
| 920 | 222.2 | 224.2 | 224.2 | | 15 | 2 | 223.5 | 8 | ICP-OES | -0.2 | -0.5 |
| 922 | 172.9 | 155.6 | 187.8 | 173.4 | 10.8 | 2 | 172.4 | 5.4 | FAAS | -0.9 | -2.2 |
| 924 | 25.6 | 25.9 | 25.2 | | 6.4 | 2 | 25.6 | 3.2 | ICP-OES | -3.0 | -7.1 |
| 951 | 67.69 | 88.48 | 92.61 | | 13.56 | $\sqrt{3}$ | 82.93 | 7.83 | Graphite Furnace AAS | -2.2 | -5.1 |
| 975 | 160 | 145 | | | 21.5 | 2 | 153 | 10.8 | ICP-MS | -1.2 | -2.7 |
| 994 | 336 | 318 | 325 | | 10.4 | 2 | 326 | 5.2 | ICP-OES | 1.2 | 2.8 |

* $\sqrt{3}$ is set by the ILC coordinator when no expansion factor k is reported. The reported uncertainty was assumed to have a rectangular distribution with $k=\sqrt{3}$. For explanation see Ch 9.3.1

IMEP-24 (heavy metals in toys): Selenium


Certified value: $X_{ref} = 240 \text{ mg}\cdot\text{kg}^{-1}$; $U_{ref} = 60 \text{ mg}\cdot\text{kg}^{-1}$ ($k=2$)



This graph displays all revised measurement results and their associated uncertainties. The uncertainties are shown as reported. The thick black line corresponds to X_{ref} , the blue lines mark the boundary of the reference interval ($X_{ref} \pm 2U_{ref}$), and the orange lines that of the target interval ($X_{ref} \pm 2\sigma$).



Annex 18 : Kernel densities



Annex 19 : Summary of scorings and results interpretation

| Part Nr | Sb | | | As | | | Ba | | | Cd | | | Cr | | | Pb | | | Hg | | | Se | | | Questionnaire Q4: Would you accept or reject material on market? |
|------------|------|------|------|------|------|------|------|-------|------|-------|-------|------|------|------|------|------|------|------|------|------|------|------|------|------|--|
| | z | zeta | Int* | z | zeta | Int* | z | zeta | Int* | Z | zeta | Int* | z | zeta | Int* | z | zeta | Int* | z | zeta | Int* | z | zeta | Int* | |
| 004 | 0.6 | 1.0 | TN | -0.2 | -0.3 | TN | 1.3 | 1.3 | TN | 3.9 | 2.7 | TP | 1.9 | 1.1 | TN | 1.6 | 1.1 | TP | 0.9 | 0.7 | TP | -0.3 | -0.6 | TN | Reject |
| 024 | 1.4 | 0.9 | TN | -0.9 | -1.0 | TN | 0.3 | 0.2 | TN | 1.1 | 0.9 | TP | -0.3 | -0.2 | TN | -2.2 | -1.9 | FN | -1.1 | -1.1 | TP | -1.3 | -1.8 | TN | Reject |
| 029 | -0.2 | -0.3 | TN | -1.3 | -2.6 | TN | -0.3 | -0.7 | TN | 2.6 | 3.5 | TP | -0.6 | -0.5 | TN | 0.7 | 0.7 | TP | -0.9 | -1.2 | TP | -0.8 | -1.7 | TN | Reject |
| 046 | 0.8 | 2.2 | TN | -1.3 | -2.9 | TN | -0.1 | -0.3 | TN | 0.1 | 0.2 | TP | 0.0 | 0.0 | TN | -2.8 | -3.0 | FN | -1.8 | -2.6 | TP | -1.8 | -4.3 | TN | Reject |
| 058 | 0.6 | 1.0 | TN | 19.3 | 12.7 | FP | | | | 243.6 | 132.5 | TP | | | | 6.4 | 5.4 | TP | 46.0 | 53.5 | TP | 16.3 | 34.1 | FP | Reject |
| 074 | -1.8 | | TN | -2.2 | | TN | 0.9 | | TN | 4.5 | | TP | 1.2 | | TN | 3.6 | | TP | -2.5 | | TP | -1.0 | | TN | Reject |
| 142 | 0.7 | 1.4 | TN | -1.7 | -3.7 | TN | 1.3 | 3.1 | TN | 8.9 | 4.5 | TP | 2.3 | 1.8 | FP | 4.5 | 4.2 | TP | -3.7 | -5.4 | FN | 1.3 | 2.1 | TN | Reject |
| 150 | -0.2 | -0.5 | TN | -1.3 | -2.8 | TN | -1.9 | -4.1 | TN | 0.2 | 0.2 | TP | -2.5 | -2.0 | TN | -1.2 | -1.2 | FN | -1.1 | -1.6 | TP | -1.3 | -2.9 | TN | Reject |
| 224 | 2.7 | 3.4 | TN | 0.7 | 1.1 | TN | 1.5 | 1.5 | TN | 4.9 | 3.4 | TP | 0.2 | 0.1 | TN | 5.3 | 3.2 | TP | 7.0 | 4.9 | TP | 1.4 | 2.0 | TN | Reject |
| 233 | 0.4 | 1.0 | TN | -0.5 | -0.5 | TN | -1.0 | -2.6 | TN | 1.3 | 2.0 | TP | -1.4 | -1.1 | TN | -1.1 | -1.2 | FN | 0.7 | 1.0 | TP | -0.9 | -2.2 | TN | Reject |
| 239 | -1.6 | -3.8 | TN | -2.3 | -5.0 | TN | -3.4 | -6.8 | TN | -3.2 | -4.6 | FN | -4.1 | -3.2 | TN | -4.1 | -4.2 | FN | -3.8 | -5.5 | FN | -2.1 | -4.9 | TN | Accept |
| 285 | -2.4 | -6.3 | TN | -2.8 | -6.0 | TN | -5.1 | -13.2 | TN | -5.5 | -9.1 | FN | -5.6 | -4.5 | TN | -5.9 | -6.2 | FN | -2.9 | -4.2 | FN | -2.9 | -6.9 | TN | Reject |
| 298 | | | | | | | | | | 10.8 | 17.9 | TP | 6.9 | 5.3 | FP | 7.6 | 5.2 | TP | 24.0 | 30.7 | TP | | | | Reject |
| 332 | 1.7 | 4.4 | TN | 0.0 | -0.1 | TN | -0.7 | -1.8 | TN | -0.3 | -0.4 | TP | 0.9 | 0.7 | TN | -4.0 | -4.2 | FN | -2.6 | -3.8 | TP | -2.7 | -6.6 | TN | Reject |
| 371 | -0.2 | -0.4 | TN | -0.5 | -1.1 | TN | -0.3 | -0.7 | TN | -0.4 | -0.5 | TP | -0.8 | -0.6 | TN | 2.3 | 2.4 | TP | -1.1 | -1.6 | TP | -0.7 | -1.6 | TN | Reject |
| 405 | 4.4 | | FP | -0.7 | | TN | 3.0 | | TN | 2.9 | | TP | 2.1 | | TN | -1.7 | | FN | -2.1 | | TP | -1.8 | | TN | Reject |
| 421 | -0.2 | -0.3 | TN | -1.3 | -2.4 | TN | -0.6 | -0.8 | TN | 2.0 | 1.9 | TP | -0.3 | -0.2 | TN | 2.6 | 1.8 | TP | -3.7 | -5.4 | FN | -1.0 | -2.1 | TN | Reject |
| 422 | -1.8 | -4.5 | TN | -2.6 | -5.6 | TN | -3.0 | -7.1 | TN | -2.6 | -4.1 | FN | -2.5 | -2.0 | TN | -5.4 | -5.6 | FN | -3.9 | -5.6 | FN | -2.9 | -6.9 | TN | Reject |
| 458 | 1.2 | 2.8 | TN | -0.1 | -0.3 | TN | 1.5 | 3.9 | TN | 3.3 | 5.4 | TP | 1.1 | 0.9 | TN | -0.9 | -0.8 | FN | -2.1 | -2.9 | TP | -1.5 | -3.3 | TN | Reject |
| 489 | -0.1 | -0.3 | TN | -0.8 | -1.5 | TN | -0.1 | -0.3 | TN | 1.4 | 2.3 | TP | -1.0 | -0.8 | TN | -1.0 | -1.1 | FN | 0.5 | 0.5 | TP | -1.1 | -2.5 | TN | Reject |
| 522 | -2.5 | -6.4 | TN | -2.9 | -6.2 | TN | -5.0 | -12.5 | TN | -5.5 | -9.0 | FN | -5.6 | -4.5 | TN | -5.8 | -6.1 | FN | -3.8 | -5.5 | FN | -2.9 | -6.9 | TN | Accept |
| 557 | 1.4 | 3.0 | TN | -1.8 | -3.8 | TN | -1.5 | -3.3 | TN | -2.4 | -3.9 | FN | -1.1 | -0.9 | TN | -5.1 | -5.4 | FN | -2.5 | -3.6 | TP | -2.6 | -6.2 | TN | Reject |
| 562 | 1.1 | 2.2 | TN | -1.4 | -2.9 | TN | 1.2 | 2.0 | TN | 1.0 | 1.7 | TP | -0.6 | -0.5 | TN | -1.2 | -1.2 | FN | -3.0 | -4.3 | FN | -1.3 | -2.9 | TN | Reject |
| 590 | -3.3 | | TN | | | | | | | -6.7 | | FN | -6.7 | | TN | -6.7 | | FN | -4.0 | | FN | -3.3 | | TN | Accept |
| 593 | 0.6 | 1.3 | TN | -0.1 | -0.3 | TN | 0.8 | 2.1 | TN | 3.7 | 6.0 | TP | 0.1 | 0.1 | TN | 2.0 | 2.1 | TP | -0.4 | -0.6 | TP | 0.2 | 0.6 | TN | Reject |
| 697 | -2.8 | -7.2 | TN | | | | | | | -5.9 | -9.8 | FN | | | | | | | -3.5 | -5.0 | FN | | | | Accept |
| 713 | 5.9 | 14.6 | FP | -1.0 | | TN | 1.2 | 3.1 | TN | 0.7 | 1.2 | TP | 1.0 | 0.8 | TN | -4.6 | -4.8 | FN | 3.1 | 4.2 | TP | -0.5 | -1.1 | TN | Reject |
| 758 | 1.9 | 4.3 | TN | -1.4 | -3.0 | TN | -0.9 | -2.4 | TN | 0.7 | 1.2 | TP | 1.4 | 1.1 | TN | -1.0 | -1.0 | FN | -3.4 | -5.0 | FN | -0.6 | -1.4 | TN | Reject |
| 793 | 1.6 | 3.5 | TN | -1.4 | -2.9 | TN | -1.0 | -2.2 | TN | 0.7 | 1.1 | TP | -0.3 | -0.3 | TN | -2.2 | -2.3 | FN | -2.5 | -3.6 | TP | -1.5 | -3.5 | TN | Reject |
| 842 | 0.9 | 2.5 | TN | 0.6 | 1.3 | TN | 0.2 | 0.6 | TN | 1.4 | 2.4 | TP | 0.7 | 0.6 | TN | -2.2 | -2.3 | FN | -1.4 | -2.1 | TP | -1.8 | -4.3 | TN | Reject |
| 867 | 0.4 | 1.0 | TN | -1.2 | -2.5 | TN | -0.2 | -0.5 | TN | 1.5 | 2.3 | TP | -0.7 | -0.5 | TN | -1.1 | -1.1 | FN | -0.5 | -0.7 | TP | -1.0 | -2.3 | TN | Accept |
| 920 | -0.1 | -0.3 | TN | -0.2 | -0.1 | TN | -0.4 | -1.1 | TN | 1.4 | 2.0 | TP | -1.5 | -1.0 | TN | 0.5 | 0.5 | TP | 3.3 | 4.8 | TP | -0.2 | -0.5 | TN | Reject |
| 922 | -1.0 | -2.6 | TN | | | TN | 0.1 | 0.2 | TN | 4.7 | 6.9 | TP | 2.3 | 1.8 | FP | 1.1 | 1.1 | TP | 1.3 | 1.9 | TP | -0.9 | -2.2 | TN | Reject |
| 924 | -1.7 | -4.0 | TN | | | TN | -3.8 | -7.2 | TN | -5.7 | -9.3 | FN | -4.7 | -3.7 | TN | -6.0 | -6.2 | FN | -3.5 | -5.1 | FN | -3.0 | -7.1 | TN | Accept |
| 927 | | | | | | | 0.6 | | TN | 0.5 | | TP | 1.1 | | TN | | | | | | | | | | Reject |
| 936 | 0.0 | -0.1 | TN | 0.7 | 1.1 | TN | 0.2 | 0.4 | TN | 2.8 | 4.5 | TP | -0.2 | -0.1 | TN | 0.7 | 0.7 | TP | -2.5 | -3.4 | TP | | | | Reject |
| 951 | -2.5 | -6.5 | TN | -2.4 | -5.3 | TN | 0.0 | 0.0 | TN | 0.8 | 1.0 | TP | -1.2 | -0.9 | TN | -1.4 | -1.3 | FN | 1.2 | 1.4 | TP | -2.2 | -5.1 | TN | Accept |
| 975 | -0.3 | -0.6 | TN | 0.2 | 0.2 | TN | 1.0 | 1.2 | TN | 3.9 | 3.9 | TP | 0.9 | 0.7 | TN | 0.5 | 0.4 | TP | -1.6 | -2.0 | TP | -1.2 | -2.7 | TN | Reject |
| 994 | 2.0 | | TN | 1.1 | 2.4 | TN | 1.0 | 2.4 | TN | 4.5 | 7.3 | TP | 0.7 | 0.5 | TN | 3.2 | 3.3 | TP | -1.5 | -2.2 | TP | 1.2 | 2.8 | TN | Accept, Reject |

* Int is the interpretation whether the reported result is a true negative (TN), true positive (TP), false negative (FN) or false positive (FP) result. Highlighted in bold red are participants whose conclusion is either *Accept* as this conclusion is considered wrong or *Reject* when the reported results do not give rise to this conclusion. See Ch 9.4.2 for further details.

Annex 20 : Experimental details derived from questionnaire

| Part Nr | Official Method | Modification / Method description | Sample amount | Sieve size | Shaking device |
|---------|-----------------|---|--|------------------|---|
| 004 | EN71-3 | | meas 1: 300 mg, meas 2: 200 mg, meas 3: 280 mg | 0.5 mm | Grant OLS 200 |
| 024 | EN71-3 | | 1 g | 500 micron | None |
| 029 | EN71-3 | | | | |
| 046 | EN71-3 | Paint was scraped / homogenised / sieved through 500µm mesh | | 500 µm | MS2 Minishaker IKA |
| 058 | No | XRF screening test for content (no migration) | | | |
| 074 | EN71-3 | EN 71-3, paragraph 8.1.1 | 0.4115 g | 0.5 mm | linitest |
| 082 | EN71-3 | | 0.13-0.18 g | 0.5 mm | shaking water bath |
| 142 | EN71-3 | | 100 mg | 0.5 mm | horizontal shaker |
| 150 | EN71-3 modified | No sieving; Weight to Volume 1:50, add 0.07 mol/l HCl (pH 1.4), 1 h shaken, 1 h at 37° C, filtrated, measurement with ICP | 0.2 - 0.3 g | n.a. | typical lab shaker |
| 176 | EN71-3 | | 0.31 | 500 micron | Horizontal shaking water bath 200 strokes/min |
| 224 | EN71-3 | | made-ground 0.1 g / 5 ml | pore size 0.5 mm | horizontal shaker LT2 |
| 233 | EN71-3 modified | we removed 0.24 g of sample and diluted it with 25 ml of HCl 0.07N | 0.24 g | whatman 1 | maxi shake Heto |
| 239 | EN71-3 | | 0.3 g, 0.4 g | 0.5 mm | normal stirrer |
| 285 | EN71-3 | | 200 mg | 0.5 mm | Kreisschüttler |
| 298 | EN71-3 modified | No Sieve Filtration | 0.105 g; 0.068 g | none | by hand |
| 332 | EN71-3 | The test were carried out according to EN 71-3 point 8.1.2 | 1 g | 500 micron | water bath with shaking device |
| 371 | EN71-3 | We have scrap the paint ? pass through the pore size ? weight the powder obtained add a volume equal to 50x mass : skake one hour at 37°C without light rest one hour and after filter | good | 500 mesh | rotation magnet stired |
| 405 | EN71-3 | | 0.2 g | 0.5 mm | Shaking water bath |
| 421 | EN71-3 | | 0.1 g for ICP-MS 0.3 for FAAS and CVAAS (Pb & Hg) | none | magnetic stirrer |
| 422 | EN71-3 | | | | |
| 458 | EN71-3 | 0.15 g of sievedpaint was extracted with 7.5 ml of 0.07M hydrochloric acid in a plastic test tube. The paint was shaken at 37°C in the dark in a water bath before allowing to stand for another hour at 37°C. The solution was filtered through a 0.45 µm PVDF syringe filter and analsed using a VARIAN VISTA PRO ICP-OES with axial torch. | 0.15 g | 0.5 mm | grant shaking water bath |
| 489 | EN71-3 | | 0.2 g | 300 - 500 um | mechanical shaker |
| 522 | EN71-3 | | 0.2 g | 300 - 500 um | mechanical shaker |

IMEP-24: Analysis of eight heavy metals in toys according to EN 71-3:1994

| Part Nr | Official Method | Modification / Method description | Sample amount | Sieve size | Shaking device |
|---------|-----------------|---|--------------------------------|---------------------------------|---|
| 557 | EN71-3 | Part 8.1 of SN EN 71-3 | 1 g | 0.5 mm | horizontal shaking device |
| 562 | EN71-3 | | 0.1 g | 0.5 mm | |
| 590 | EN71-3 | | | 0.5 mm | Shaker with ideal swivel motion IKA KS 130 basic |
| 593 | EN71-3 | the sample with 0.07 N of HCl is agited into a water bath at 37°C during 1 hour, and without agitation during 1 hour. The sample is then filtered and analyzed into ICP-OES | 0.1 g | 500 µm | Water bath |
| 661 | EN71-3 | | 600 mg | 500 µm | mechanical shaking |
| 697 | EN71-3 | 0.5 g/25 ml; 1h shaking, 1h standing (37±2°C), 9.9 ml + 0.1 ml internal standard, Hg dilution 1:10 | 0.5 g | 520 | horizontal shaker with circular motion |
| 713 | EN71-3 | EN 71-3:1994 + A1 2000 + AC 2002/ HCl Extraction 2h/ Filtration/ ICP-MS | 50 mg each | non | magnetic stirrer, water bath |
| 758 | EN71-3 | | 0.5 g | 500 micron | magnetic stirring |
| 779 | EN71-3 | | 200 mg | 500 um | orbital shaker |
| 793 | EN71-3 | Paint is scraped off and sieved on 0.5 mm sieve. Subsample is the weighed into extraction tube, and extraction liqiud is added (L/S=50). After extraction liqiud is analyzed on ICPMS. | 0.2 g | 0.5 mm | Shaking table |
| 842 | EN71-3 | Sample scraped from plate using scalpel and sieved. Weighed out 0.3g in triplicate into conical flask. 0.07 HCl solution preheated to 37C. Pipetted 15ml of 0.07 HCl in to each flask. Flask shaken and pH measured. Sample placed in water bath at 37C and shaken in the dark for 1h, then left to stand for 1 h. Solution filtered and diluted to 25ml prior to analysis by ICP-OES | 0.3 g | 0.5 mm | Shaken by hand |
| 867 | EN71-3 | | 0.2 g | 0.5 mm | horizontal shaker |
| 920 | EN71-3 modified | 5 ml extraction volume diluted 4x before analysis | 0.1 g | 0.5 mm | orbital |
| 922 | EN71-3 | migration of heavy metals (Sb, As, Ba, Cd, Cr, Pb, Hg and Se) from the toy to the hydrochloric acid solution 0.07M, determination of the metals migrated using FAAS | 1.0092 | 0.05 | a water-bath with a shaking device |
| 924 | EN71-3 | The test portion was 200mg and mixed with 10 ml of 0.07 mol/l HCl and diluted to 25 ml. | 200 mg | 0.5 mm | Thermostatic inkubator |
| 927 | EN71-3 | | 1 g | 500 micron | none |
| 936 | EN71-3 modified | The sample wasn't sieved because of the available amount | 0.2 g | none | water bath with orbital saking |
| 951 | EN71-3 modified | Sieving was not possible, because the static electricity was to high | about 100 mg for one sample | | Roller Mixer SRT 2 - Stuart Scientific, CAT No SRT2; Nr: R000 101 155 |
| 975 | EN71-3 | | 200 mg | 500 micron | Waterbath with agitation |
| 994 | EN71-3 | | 0.1033. 0.1027. 0.1005 gram | nominal aperture size 0.5 mm | water shaker bath 150rpm |

IMEP-24: Analysis of eight heavy metals in toys according to EN 71-3:1994

| Part Nr | Preheated labware | adjusted pH ? | Final pH | Membrane filter | Centrifugation | Solid/Acid ratio > 1:50 | Analysis on day of preparation? | If not, adjusted acid concentration to 1 mol/l HCl ? | Inclusion of base material |
|---------|-------------------|---------------|------------------|--|----------------|---|---------------------------------|--|----------------------------|
| 004 | No | No | 1.0-1.5 | PALL, Bulk GHP Acrodisc with 0.45 µm GHP membrane | No | no | Yes | | No |
| 024 | No | No | | | No | No | No | No | No |
| 029 | No | No | about 1 | 0.45 µm | No | no | No | Yes | No |
| 046 | No | No | 1 | 0.45 µm Minsart (Sartorius) | No | no | Yes | | No |
| 058 | | | | | | | | | |
| 074 | No | No | | 0.45 um, sartorius | No | no | Yes | | No |
| 082 | No | No | <1.5 | Acrodisc Syringe Filters, 0.45 µm | No | no | Yes | | No |
| 142 | No | Yes | < 1.5 | 0.45 µm cellulose acetate | No | no | No | Yes | No |
| 150 | Yes | No | 1.4 | 0.45 um | No | no | Yes | | No |
| 176 | Yes | No | | millipore 0.45µm aqueous syringe filter | No | No | Yes | No | No |
| 224 | Yes | Yes | 1.1 | Whatman ME 25/21 St, Membrane Filtres 0.45 µm | No | the ratio of solid to acid was exactly 1:50 | Yes | No | No |
| 233 | Yes | No | 1.4 | | No | no, it was 1:104 | Yes | | Yes |
| 239 | Yes | No | 1 | 0.45 um | No | No | Yes | | |
| 285 | Yes | No | 1.3 | 0.43 µm | No | No | Yes | | Yes |
| 298 | No | No | | 0.45 um | No | no | No | Yes | No |
| 332 | Yes | No | 1.22 | 0.45 µm | No | no | Yes | No | No |
| 371 | Yes | No | 1.5 | 0.45 µm | No | no | Yes | | No |
| 405 | No | No | 1.5 | 0.45 um | No | No | Yes | Yes | No |
| 421 | Yes | No | 1.25 | 0.45 um membrane filter | No | no | Yes | No | No |
| 422 | | | | | | | | | |
| 458 | No | No | 1.0 to 1.5 | PVDF 0.45 µm | No | No | Yes | No | No |
| 489 | Yes | No | Acidic | 0.45 um | No | no | Yes | | No |
| 522 | Yes | No | Acidic | 0.45 um | No | No | Yes | | No |
| 557 | Yes | No | < 1.5 | CM 0.45 µm | No | No | Yes | No | Yes |
| 562 | Yes | No | 1.0~1.3 | 0.45 um | No | No | Yes | | No |
| 590 | No | No | 1.4 | membrane filter (Mixed cellulosa ester) 0.45 µm diameter 0.47 mm | No | no | Yes | No | Yes |
| 593 | Yes | No | between 1 to 1.5 | acrodisc 0.45 µm | No | no | Yes | Yes | No |
| 661 | No | No | 1.3 | 0.45 µm | No | no | Yes | | No |
| 697 | No | Yes | 1.1 - 1.4 | cellulosemischester, 0.45 µm | No | no | Yes | No | No |
| 713 | No | No | 1.3 | Regenerated cellulose 0.45 µm | No | no | Yes | | No |

IMEP-24: Analysis of eight heavy metals in toys according to EN 71-3:1994

| Part Nr | Preheated labware | adjusted pH ? | Final pH | Membrane filter | Centrifugation | Solid/Acid ratio > 1:50 | Analysis on day of preparation? | If not, adjusted acid concentration to 1 mol/l HCl ? | Inclusion of base material |
|---------|-------------------|---------------|-----------|---------------------------------------|----------------|-------------------------|---------------------------------|--|----------------------------|
| 758 | Yes | No | 1.5 | Schleicher and Schuell, 595, 4 - 7 µm | No | no | Yes | | No |
| 779 | Yes | No | 1.5 | 0.45 µm | No | no | Yes | | No |
| 793 | Yes | No | below 1.5 | 0.45 µm polycarbonate | No | no | No | No | No |
| 842 | Yes | No | 1.15 | 0.45 µm Nylon | No | No | Yes | | No |
| 867 | No | No | | 0.45 | No | | No | No | No |
| 920 | No | No | 1.3 | millipore 0.45 µm | No | no | Yes | | No |
| 922 | Yes | Yes | 1.12 | whatman 1001 125 | No | no | Yes | No | No |
| 924 | No | No | ~1.3 | Glass microfibre filters GF/A | No | Yes | No | Yes | No |
| 927 | Yes | No | 1.28 | whatman 50 | No | no, it didn't | No | No | No |
| 936 | No | No | 1.3 | 0.45 µm; cellulose nitrate | No | no | No | Yes | No |
| 951 | Yes | No | 1.3 | Nylon - 0.45 µm | No | no | No | | Yes |
| 975 | Yes | No | 1-1.5 | 45 µm | No | No | Yes | | No |
| 994 | Yes | No | 1.3 | pore size of 0.45 µm | No | no | Yes | | No |

European Commission

EUR 24094 EN – Joint Research Centre – Institute for Reference Materials and Measurements

Title: IMEP-24: Analysis of eight heavy metals in toys according to EN 71-3:1994

Author(s): Ines Baer, Johannes van de Kreeke, Inge Verbist, Danny Vendelbo, Thomas Linsinger, Piotr Robouch, Philip Taylor, Beatriz de la Calle

Luxembourg: Office for Official Publications of the European Communities

2009 – 61 pp. – 21 x 29.7 cm

EUR – Scientific and Technical Research series – ISSN 1018-5593

ISBN 978-92-79-14273-4

DOI 10.2787/191

Abstract

The Institute for Reference Materials and Measurements (IRMM) of the Joint Research Centre (JRC), a Directorate-General of the European Commission, operates the International Measurement Evaluation Programme (IMEP). It organises interlaboratory comparisons (ILC's) in support to EU policies. This report presents the results of an ILC which focussed on the determination of soluble antimony, arsenic, barium, cadmium, chromium, lead, mercury, and selenium according to European Standard EN 71-3:1994.

The principle of the procedure in EN 71-3:1994 consists in the extraction of soluble elements from toy material under the conditions simulating the material remaining in contact with stomach acid for a period of time after swallowing.

Forty participants from eighteen countries registered to the exercise, of which 33 reported results for As, 35 for Ba and Se, 37 for Cr, Pb, and Sb, 38 for Hg, and 39 for Cd. For seven measurands the test material had already been certified in the past. The validity of the certificate was reconfirmed and the certified values were taken as the reference values for this ILC. As no certified value was available for Hg, the mean value of the results provided by four expert laboratories was used together with the corresponding uncertainty. Participants were invited to report the uncertainty on their measurements. This was done by 35 of the 39 laboratories having submitted results in this exercise.

Laboratory results were rated with z and zeta scores in accordance with ISO 13528. The standard deviations for proficiency assessment were based on the analytical correction laid down in EN 71-3:1994.

How to obtain EU publications

Our priced publications are available from EU Bookshop (<http://bookshop.europa.eu>), where you can place an order with the sales agent of your choice.

The Publications Office has a worldwide network of sales agents. You can obtain their contact details by sending a fax to (352) 29 29-42758.

The mission of the JRC is to provide customer-driven scientific and technical support for the conception, development, implementation and monitoring of EU policies. As a service of the European Commission, the JRC functions as a reference centre of science and technology for the Union. Close to the policy-making process, it serves the common interest of the Member States, while being independent of special interests, whether private or national.

