

# JRC TECHNICAL REPORTS

Report on the 2016 Proficiency Test of the European Union Reference Laboratory for Mycotoxins for the network of National Reference Laboratories

> Determination of aflatoxin B1 in defatted peanut powder

Katrien Bouten Elena Cubero-Leon Andreas Breidbach Carsten Mischke Joerg Stroka

2017



This publication is a Technical report by the Joint Research Centre (JRC), the European Commission's science and knowledge service. It aims to provide evidence-based scientific support to the European policymaking process. The scientific output expressed does not imply a policy position of the European Commission. Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use that might be made of this publication.

#### **Contact information**

Name: Joerg Stroka Address: Retieseweg 111 2440 Geel Email: Joerg.Stroka@ec.europa.eu Tel.: 014571229

JRC Science Hub

https://ec.europa.eu/jrc

JRC107711

EUR 28740 EN

PDF ISBN 978-92-79-72215-8 ISSN 1831-9424 doi:10.2760/56278

Luxembourg: Publications Office of the European Union, 2017

#### © European Union, 2017

The reuse of the document is authorised, provided the source is acknowledged and the original meaning or message of the texts are not distorted. The European Commission shall not be held liable for any consequences stemming from the reuse.

How to cite this report: Bouten, K., Cubero-Leon, E., Breidbach, A., Mischke, C., Stroka, J., *Report on the 2016 Proficiency Test of the European Union Reference Laboratory for Mycotoxins for the network of National Reference Laboratories: Determination of aflatoxin B1 in defatted peanut powder*, EUR 28740 EN, Publications Office of the European Union, Luxembourg, 2017, ISBN 978-92-79-72215-8, doi:10.2760/56278, JRC107711.

All images © European Union 2017, except: https://pixabay.com/en/peanuts-shell-food-delicious-1735672/

# Contents

| Ac  | knowledgements1                                  |
|-----|--|
| AŁ  | ostract3   |
| 1   | Introduction4                                    |
| 2   | Scope  |
|     | 2.1 Confidentiality                              |
| 3   | Time Frame5                                      |
| 4   | Material5  |
|     | 4.1 Preparation5                                 |
|     | 4.2 Homogeneity5                                 |
|     | 4.3 Stability5                                   |
|     | 4.4 Distribution                                 |
| 5   | Instruction to participants                      |
| 6   | Reference values and their uncertainties6        |
| 7   | Evaluation of the results7                       |
|     | 7.1 General observations7                        |
|     | 7.2 Scores and evaluation criteria7              |
|     | 7.3 Laboratory results and scoring8              |
|     | 7.4 Evaluation of questionnaire13                |
| 8   | Conclusions13                                    |
| Re  | eferences14                                      |
| Lis | st of abbreviations and definitions15            |
| Lis | st of figures                                    |
| Lis | st of tables16                                   |
| Ar  | inexes   |
|     | Annex 1. Announcement- opening of registration18 |
|     | Annex 2. Homogeneity study                       |
|     | Annex 3. Stability study                         |
|     | Annex 4. Sample accompanying letter              |
|     | Annex 5. Material receipt form                   |
|     | Annex 6. Questionnaire                           |
|     | Annex 7. Aflatoxin B1 Z- & zeta scores graphs    |
|     | Annex 8. Experimental details                    |
|     | Annex 9. Recoveries                              |

# Acknowledgements

The organizers of the study would like to thank the colleagues involved in the project for their support. The laboratories participating in this exercise, listed in Table 1, are also kindly acknowledged.

| Organisation  | Country        |
|---|----------------|
| AGES GmbH   | Austria        |
| LVA GmbH  | Austria        |
| CODA-CERVA  | Belgium        |
| Central Laboratory for Chemical Testing and Control, BFSA                                   | Bulgaria       |
| E.C. Inspekt d.o.o.   | Croatia        |
| Andrija Stampar Teaching Institute for Public Health  | Croatia        |
| State General Laboratory  | Cyprus         |
| UKZUZ - Central Institute for Supervising and Testing in Agriculture                        | Czech Republic |
| Czech Agriculture and Food Inspection Authority (CAFIA)                                     | Czech Republic |
| DTU Food  | Denmark        |
| Danish Veterinary and Food Administration   | Denmark        |
| Agricultural Research Centre, laboratory for Residues and Contaminants                      | Estonia        |
| Central Chemistry Laboratory of Health Board  | Estonia        |
| Finnish Customs Laboratory  | Finland        |
| Laboratoire SCL de Rennes   | France         |
| LUFA-ITL GmbH   | Germany        |
| CVUA Westfalen - Standort Hamm -  | Germany        |
| Federal Institute for Risk Assessment (BfR)   | Germany        |
| Feedstuffs control laboratory of Athens – Ministry of rural development and food            | Greece         |
| Chemical State Laboratory, Division of Piraeus and Aegean                                   | Greece         |
| National Food Chain Safety Office, Food And Feed Safety Directorate, Food Toxicological NRL | Hungary        |
| National Food Chain Safety Office, Food and Feed Safety Directorate, Feed Investigation NRL | Hungary        |
| The State Laboratory  | Ireland        |
| Public Analyst's Laboratory   | Ireland        |
| Istituto Superiore di Sanita'   | Italy          |
| Institute of Food Safety, Animal Health and Environment "BIOR"                              | Latvia         |
| National Food And Veterinary Risk Assessment Institute                                      | Lithuania      |
| Laboratoire national de santé   | Luxembourg     |
| Public Health Laboratory  | Malta          |
| RIKILT - Wageningen UR  | Netherlands    |
| The Norwegian Veterinary Institute  | Norway         |
| National Institute of Public Health - National Institute of Hygiene                         | Poland         |
| National Veterinary Research Institute  | Poland         |
| ASAE - DRAL - LFQ   | Portugal       |
| Sanitary Veterinary and Food Safety Directorate Bucharest                                   | Romania        |
| DSVSA GALATI - LSVSA  | Romania        |
| Hygiene and Veterinary Public Health Institute  | Romania        |
| SP Laboratorija A.D.  | Serbia         |
| Regional Public Health Authority - RUVZ so sídlom v Poprade                                 | Slovakia       |
| State veterinary and food institute Dolný Kubín, Veterinary and food institute in Košice    | Slovakia       |
| University of Ljubljana, Veterinary Faculty, National Veterinary Institute                  | Slovenia       |
| National laboratory for health, environment and food  | Slovenia       |
| Laboratorio de Salud Pública de Albacete ( Junta de Comunidades de Castilla-La Mancha)      | Spain          |
| National Centre for food – Spanish consuming, food safety and nutrition agency              | Spain          |
| National Food Agency  | Sweden         |
| National Veterinary Institute, SVA  | Sweden         |

## Table 1: Participating laboratories

| Organisation                                     | Country |
|--|---------|
| Worcestershire Scientific Services               | UK      |
| Edinburgh Scientific Services                    | UK      |
| Kent Scientific Services                         | UK      |
| Tayside Scientific Services                      | UK      |
| West Yorkshire Joint Services                    | UK      |
| Public Analyst Scientific Services Ltd (Norwich) | UK      |
| Hampshire Scientific Services                    | UK      |
| Fera Science Ltd.                                | UK      |

# Abstract

The Joint Research Centre (JRC), a Directorate-General of the European Commission, operates the European Union Reference Laboratory (EURL) for Mycotoxins. One of its core tasks is to organise proficiency tests (PTs) among appointed National Reference Laboratories (NRLs).

This report presents the results of the PT on the determination of aflatoxins in defatted peanut powder.

The test items for this PT were two contaminated defatted peanut powder samples. The materials were produced by the JRC and dispatched to the participants mid October 2016. Each participant received one bottle per test material containing approximately 55 g each.

Fifty-six participants from thirty countries (among them 40 NRLs and 16 official food control laboratories) registered for the exercise and 54 sets (Sample A and B) of results were reported.

The assigned values, established by an exact-matching double isotope dilution mass spectrometric technique, were 2.80  $\mu$ g/kg (± 0.19  $\mu$ g/kg) aflatoxin B1 in sample A and 3.20  $\mu$ g/kg (± 0.20  $\mu$ g/kg) in sample B.

Participants' results were rated with z-scores and zeta-scores for aflatoxin B1 in accordance with ISO 13528:2015. The z-score compares the participant's deviation from the reference value with the target standard deviation accepted for the PT, whereas the zeta-score provides an indication of whether the participant's estimate of uncertainty is consistent with the observed deviation from the assigned value.

Only z-scores were used for the evaluation whether an individual laboratory underperformed. In total, 96 % of the attributed z-scores were below an absolute value of two for sample A and and 92 % for sample B. This indicates that most of the participants performed satisfactorily. The few participants that had z-scores above an absolute value of 2 will have to investigate the reasons for the deviation (root-cause analysis) and report the planned corrective actions to the EURL.

# **1** Introduction

Aflatoxins are mycotoxins produced by strains of *Aspergillus flavus*, *A. parasiticus* and *A. nomius*. *A. flavus* produces aflatoxins B only, while the other species grow both aflatoxin B and aflatoxin G. They contaminate mainly maize and peanuts, but also other food and feed materials like nuts and spices.

Due to their carcinogenic, hepatotoxic, teratogenic and mutagenic effect in humans and farm animals (Wild and Turner, 2002) aflatoxins are a threat to public health. Aflatoxin B1 (Figure 1) is the most potent hepatocarcinogen known in mammals and it is classified by the International Agency of Research on Cancer (IARC, 1993) as a Group 1 carcinogen.




Figure1: Formula of Aflatoxin B1, Aflatoxin G1, Aflatoxin B2 and Aflatoxin G2

The most frequently used method for aflatoxin determination is high performance liquid chromatography (HPLC) with Kobra Cell derivatisation and fluorescence detection (Yao *et al.*, 2015; EN 14123:2003). Other successfully employed methods for the determination of aflatoxins are liquid chromatography coupled to mass spectrometry (LC-MS) or tandem mass spectrometry (LC-MS/MS). Screening methods, like enzyme-linked immuno sorbent assay (ELISA) (Ardic *et al.*, 2008) and immunochromatography (Shim *et al.*, 2007) can be used for the determination of total aflatoxins.

Peanuts are a valuable source of vegetable proteins and often the oil is used for human consumption while the protein remainder is fed to animals. The maximum levels of aflatoxin B1 in animal feed are laid down in Directive 2002/32/EC.

# 2 Scope

As stated in Article 32 of Regulation (EC) No 882/2004, one of the core duties of the EURL is to organise proficiency tests (PTs) for the benefit of staff of NRLs. The scope of this PT was to test the competence of the appointed NRLs to determine the amount of aflatoxin B1 in defatted peanut powder. All invited laboratories were allowed to use their method of choice.

The EURL Mycotoxins performed the assessment of the measurement results on the basis of requirements laid down in legislation and followed the administrative and logistic procedures of ISO/IEC 17043:2010. The PT activities of the EURL are performed under ISO/IEC 17043:2010 accreditation.

# **2.1 Confidentiality**

Confidentiality of the participant's results towards third parties is guaranteed by transmission of data through a dedicated web-based interface and a secure databank hosted by the JRC. European Commission rules on data protection were strictly followed as well.

# 3 Time Frame

The PT was opened for registration on  $1^{th}$  of July 2016 and the deadline for registration was  $19^{th}$  of July 2016. (Annex 1) The samples were dispatched to the participants on  $18^{th}$  and  $19^{th}$  of October 2016. The deadline for reporting the results was the  $9^{th}$  of December 2016.

# 4 Material

# 4.1 Preparation

The test materials used in this study were one naturally contaminated peanut powder (sample A) and one spiked peanut powder (sample B). The material for sample A was available at the JRC as fine powder. It was homogenised in a tumble mixer prior designation as test material.

The material for sample B was prepared by fortifying a small portion of peanut powder with an aliquot of a solution of all 4 aflatoxins. This fortified portion was blended with blank material. This blend was subsequently ground with a Retsch ZM 200 ultra-centrifugal mill, using sieves size of 1.0 mm and 0.5 mm.

Both materials were packed in amber plastic containers, taking portions from different places of the lot at random. The total sample size was ca. 55 g.

# 4.2 Homogeneity

To verify the homogeneity of the test materials, 10 units per material (Sample A and Sample B) were randomly selected. Two independent determinations per bottle were performed using a liquid chromatography isotope dilution tandem mass spectrometry detection (LC-ID-MS/MS) based method. The order of measurements of the batch was randomised. Homogeneity was evaluated according to the International Harmonised Protocol for the Proficiency Testing of Analytical Chemistry Laboratories (Thompson *et al.*, 2006). The materials proved to be adequately homogeneous (Annex 2).

# 4.3 Stability

The stability study was conducted following an isochronous experimental design (Lamberty *et al.*, 1995); -18 °C was chosen as reference temperature for sample storage. Stability was evaluated according to the International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories (Thompson *et al.*, 2006). The materials proved to be adequately stable at 25 °C, 4 °C and -18 °C for 6 weeks. The results of the study are listed in Annex 3.

# 4.4 Distribution

The test materials were dispatched at ambient temperature on 18<sup>th</sup> and 19<sup>th</sup> of October 2016. The samples were mostly received within 24 hours after dispatch.

Each participant received:

- *a)* two units containing approximately 55 g of each test material
- b) an accompanying letter with instructions on sample handling and reporting (Annex 4)
- c) a material receipt form (Annex 5)
- *d*) laboratory specific reporting files with a lab code by email

# **5** Instruction to participants

The laboratories were asked to report the recovery corrected levels as well as their expanded measurement uncertainty in  $\mu$ g/kg (coverage factor k=2) for the aflatoxins they are analysing in their laboratory on a routine basis.

Results were reported by the participants using the RingDat software, which is part of the ProLab software. Laboratory specific files generated by the ProLab software were sent to each laboratory individually (personal files) by email. A specific questionnaire was also included. The questionnaire was intended to provide further information on method and laboratory details to allow conclusions on possible individual and common effects observed for possible follow-up procedures. A copy of the questionnaire is shown in Annex 6.

Participants received the information that the materials were shipped at ambient temperature and that upon arrival the materials needed to be stored immediately at -18  $^{\circ}$ C until the analysis is performed.

# 6 Reference values and their uncertainties

The assigned values of aflatoxin B1 and their uncertainties for the test samples were established by "Exact-Matching Double Isotope Dilution Mass Spectrometry" at the JRC. This methodology is considered to provide the highest degree of accuracy of the assigned value and traceability to SI units.(Mackay *et al.* 2003). The assigned values were 2.80  $\mu$ g/kg (Sample A) and 3.20  $\mu$ g/kg (Sample B) for aflatoxin B1. The expanded uncertainties (k=2) of the respective assigned values were 0.19  $\mu$ g/kg and 0.20  $\mu$ g/kg. Other aflatoxins are not regulated for animal feed. As a result the reported values for other aflatoxins than aflatoxin B1 are reported in Tables 4 and 5 for information only.

# **7** Evaluation of the results

## 7.1 General observations

Fifty-six participants from 27 countries (among them 40 NRLs and 16 official food control laboratories) registered to the exercise [Table 3] and 54 sets of results were reported.

The laboratories were free to use their method of choice. Forty-one laboratories used HPLC-FLD, eleven analysed the samples with LCMS and two used ELISA.

## 7.2 Scores and evaluation criteria

Individual laboratory performance was assessed in terms of z and zeta ( $\zeta$ ) scores in accordance with ISO 13528:2015 and the IUPAC International Harmonised Protocol.

$$z = \frac{x_{lab} - x_{ref}}{\sigma_p}$$
 Equation 1  
$$\zeta = \frac{x_{Lab} - x_{ref}}{\sqrt{u_{lab}^2 + u_{ref}^2}}$$
 Equation 2

where:

 $x_{lab}$  is the measurement result reported by a participant

 $x_{ref}$  is the reference value (assigned value)

 $u_{lab}$  is the standard uncertainty reported by a participant

 $u_{ref}$  is the standard uncertainty of the reference value

 $\sigma_p$  is the standard deviation for proficiency assessment (target standard deviation)

 $\sigma_p$  was calculated using the Horwitz equation, modified by Thompson [10] for analyte concentrations < 120  $\mu g/kg$ :

- for analyte concentration < 120  $\mu$ g/kg

$$\sigma_p = 0.22 \ C$$

**Equation 3** 

where:

 ${\it C}$  is the concentration of the measurand (assigned value,  $x_{ref}$ , ) expressed as a dimensionless mass ratio, e.g. 1  $\mu g/kg$  = 10<sup>-9</sup>, 1 mg/kg = 10<sup>-6</sup>

The z-score compares the participant's deviation from the reference value with the target standard deviation accepted for the proficiency test,  $\sigma_p$ . The z-score is interpreted as:

| z  ≤ 2      | indicates satisfactory performance |
|-------------|------------------------------------|
| 2 <  z  < 3 | indicates questionable performance |

 $|z| \ge 3$  indicates unsatisfactory performance

The zeta ( $\zeta$ )-score provides an indication of whether the participant's estimate of uncertainty is consistent with the observed deviation from the assigned value. The  $\zeta$ -score is the most relevant evaluation parameter, as it includes all parts of a measurement result, namely the expected value, its uncertainty as well as the uncertainty of the assigned values.

The interpretation of the  $\zeta$ -score is similar to the interpretation of the z-score:

| ζ  ≤ 2      | indicates satisfactory performance   |
|-------------|--------------------------------------|
| 2 <  ζ  < 3 | indicates questionable performance   |
| ζ ≥3        | indicates unsatisfactory performance |

An unsatisfactory performance based on a  $|\zeta|$ -score  $\geq 3$  might be due to an underestimation of the uncertainty, or to a large deviation from the reference value, or to a combination of the two factors.

# 7.3 Laboratory results and scoring

Statistical evaluation of the results was performed using MS Excel and ProLab software 8. The robust mean values and robust standard deviations were computed according to Algorithm A of ISO 13528:2015

Summaries of the statistical evaluation for Aflatoxin B1 are presented in Table 2. The calculated z-and  $\zeta$ -scores for aflatoxin B1 in sample A and sample B are listed in Table 3. Graphs with the z-and  $\zeta$ -scores are in Annex 7.

| Aflatoxin B1  |       | Sample A   | Sample B    |
|---|-------|------------|-------------|
| Number of results   |       | 54         | 54          |
| Range of results  | µg/kg | 0.19 - 3.9 | 0.22 - 4.36 |
| Robust mean of results of participants                            | µg/kg | 2.37       | 2.85        |
| Assigned value  | µg/kg | 2.80       | 3.20        |
| Expanded uncertainty (k=2) of the assigned value                  | µg/kg | 0.19       | 0.20        |
| standard deviation for proficiency assessment $\sigma_{\text{p}}$ | µg/kg | 0.62       | 0.70        |
| Number (percentage) of results $ z  > 2.0$                        |       | 4 (7 %)    | 2 (4 %)     |

**Table 2.** Summary statistics of the results for aflatoxin B1 in sample A & sample B.

**Table 3.** Analytical results, z-and  $\zeta$ -scores for aflatoxin B1 in sample A and sample B. Colour code: green- satisfactory performance, yellow – questionable performance, red – unsatisfactory performance. Laboratories that didn't report any results are marked as "No result"

|          | Sample A<br>Aflatoxin B1 |                |           |           |           | Sample B<br>Aflatoxin B1 |           |           |           |           |
|----------|--------------------------|----------------|-----------|-----------|-----------|--------------------------|-----------|-----------|-----------|-----------|
| Lab code | Sample<br>A<br>µg/ kg    | U lab<br>(k=2) | u lab     | Z-Score   | ζ-score   | Sample<br>B<br>µg/ kg    | U lab     | u lab     | Z-Score   | ζ-score   |
| LC0002   | 2.11                     | 0.3            | 0.15      | -1.1      | -3.9      | 2.51                     | 0.30      | 0.15      | -1.0      | -3.8      |
| LC0003   | 2.47                     | 0.49           | 0.24      | -0.5      | -1.3      | 2.82                     | 0.56      | 0.28      | -0.5      | -1.3      |
| LC0004   | 2.93                     | 0.88           | 0.44      | 0.2       | 0.3       | 3.30                     | 0.99      | 0.49      | 0.1       | 0.2       |
| LC0005   | 2.90                     | 0.87           | 0.43      | 0.2       | 0.2       | 3.5                      | 1.10      | 0.55      | 0.4       | 0.5       |
| LC0006   | 2.36                     | 0.6            | 0.3       | -0.7      | -1.4      | 2.88                     | 0.73      | 0.36      | -0.5      | -0.8      |
| LC0007   | 1.68                     | 0.3            | 0.15      | -1.8      | -6.3      | 2.68                     | 0.48      | 0.24      | -0.7      | -2.0      |
| LC0008   | 3.07                     | 0.98           | 0.49      | 0.4       | 0.5       | 3.66                     | 1.20      | 0.60      | 0.7       | 0.8       |
| LC0009   | 3.90                     | 0.78           | 0.39      | 1.8       | 2.7       | 3.3                      | 0.67      | 0.33      | 0.1       | 0.3       |
| LC0010   | 1.21                     | 0.5            | 0.25      | -2.6      | -5.9      | 1.94                     | 0.80      | 0.40      | -1.8      | -3.1      |
| LC0011   | 2.38                     | 0.38           | 0.19      | -0.7      | -2        | 2.89                     | 0.46      | 0.23      | -0.4      | -1.2      |
| LC0012   | 3.00                     | 1.2            | 0.6       | 0.3       | 0.3       | 3.4                      | 1.40      | 0.70      | 0.3       | 0.3       |
| LC0013   | No result                | No result      | No result | No result | No result | No result                | No result | No result | No result | No result |
| LC0014   | 2.57                     | 0.07           | 0.04      | -0.4      | -2.3      | 3.06                     | 0.12      | 0.06      | -0.2      | -1.2      |
| LC0015   | 3.17                     | 1.27           | 0.64      | 0.6       | 0.6       | 3.49                     | 1.39      | 0.70      | 0.4       | 0.4       |
| LC0016   | 2.23                     | 0.59           | 0.29      | -0.9      | -1.8      | 2.83                     | 0.72      | 0.36      | -0.5      | -1.0      |
| LC0017   | 1.84                     | 0.5            | 0.25      | -1.6      | -3.6      | 2.68                     | 0.72      | 0.36      | -0.7      | -1.4      |
| LC0018   | 3.27                     | 1.9            | 0.95      | 0.8       | 0.5       | 3.97                     | 2.30      | 1.15      | 1.1       | 0.7       |
| LC0019   | 2.63                     | 0.7            | 0.35      | -0.3      | -0.5      | 2.9                      | 0.77      | 0.39      | -0.4      | -0.8      |
| LC0020   | No result                | No result      | No result | No result | No result | No result                | No result | No result | No result | No result |
| LC0022   | 2.23                     | 0.85           | 0.42      | -0.9      | -1.3      | 2.72                     | 1.04      | 0.52      | -0.7      | -0.9      |
| LC0023   | 1.83                     | 0.32           | 0.16      | -1.6      | -5.2      | 2.25                     | 0.40      | 0.20      | -1.3      | -4.2      |
| LC0024   | 1.67                     | 0.42           | 0.21      | -1.8      | -4.9      | 1.83                     | 0.46      | 0.23      | -1.9      | -5.5      |
| LC0025   | 3.41                     | 0.51           | 0.25      | 1         | 2.2       | 2.82                     | 0.42      | 0.21      | -0.5      | -1.6      |
| LC0026   | 2.01                     | 0.4            | 0.2       | -1.3      | -3.6      | 2.64                     | 0.53      | 0.27      | -0.8      | -2.0      |
| LC0027   | 1.68                     | 0.67           | 0.34      | -1.8      | -3.2      | 2.05                     | 0.82      | 0.41      | -1.6      | -2.7      |
| LC0028   | 3.84                     | 0.36           | 0.18      | 1.7       | 5.1       | 4.36                     | 0.36      | 0.18      | 1.6       | 5.6       |
| LC0029   | 1.97                     | 0.77           | 0.39      | -1.3      | -2.1      | 2.68                     | 1.07      | 0.54      | -0.7      | -1.0      |
| LC0030   | 1.80                     | 0.36           | 0.18      | -1.6      | -4.9      | 2.15                     | 0.43      | 0.21      | -1.5      | -4.4      |
| LC0031   | 0.19                     | 0.06           | 0.03      | -4.2      | -26.3     | 0.22                     | 0.07      | 0.03      | -4.2      | -28.3     |
| LC0032   | 1.80                     | 0.58           | 0.29      | -1.6      | -3.3      | 3.8                      | 1.21      | 0.60      | 0.9       | 1.0       |
| LC0033   | 2.65                     | 0.7            | 0.35      | -0.3      | -0.4      | 3.09                     | 0.82      | 0.41      | -0.2      | -0.3      |
| LC0035   | 2.29                     | 0.71           | 0.35      | -0.8      | -1.4      | 3.06                     | 0.95      | 0.47      | -0.2      | -0.3      |
| LC0036   | 2.90                     | 0.5            | 0.25      | 0.2       | 0.4       | 3.48                     | 0.63      | 0.31      | 0.4       | 0.8       |
| LC0037   | 2.62                     | 1.05           | 0.52      | -0.3      | -0.3      | 3.14                     | 1.26      | 0.63      | -0.1      | -0.1      |
| LC0038   | 2.65                     | 0.8            | 0.4       | -0.2      | -0.4      | 3.22                     | 1.00      | 0.50      | 0.0       | 0.0       |
| LC0039   | 1.12                     | 0.49           | 0.24      | -2.7      | -6.4      | 1.25                     | 0.55      | 0.28      | -2.8      | -6.7      |
| LC0040   | 2.56                     | 0.38           | 0.19      | -0.4      | -1.1      | 2.91                     | 0.44      | 0.22      | -0.4      | -1.2      |
| LC0041   | 2.04                     | 0.14           | 0.07      | -1.2      | -6.4      | 2.42                     | 0.11      | 0.06      | -1.1      | -6.8      |
| LC0042   | 2.79                     | 0.61           | 0.3       | 0         | 0         | 3.43                     | 0.75      | 0.38      | 0.3       | 0.6       |
| LC0043   | 2.09                     | 0.63           | 0.32      | -1.2      | -2.2      | 1.92                     | 0.58      | 0.29      | -1.8      | -4.2      |
| LC0044   | 2.78                     | 0.67           | 0.34      | 0         | -0.1      | 3.63                     | 0.87      | 0.43      | 0.6       | 1.0       |
| LC0045   | 2.10                     | 0.8            | 0.4       | -1.1      | -1.7      | 2.5                      | 1.00      | 0.50      | -1.0      | -1.4      |
| LC0046   | 3.20                     | 0.3            | 0.15      | 0.6       | 2.3       | 4.2                      | 0.40      | 0.20      | 1.4       | 4.5       |
| LC0047   | 2.11                     | 0.63           | 0.32      | -1.1      | -2.1      | 2.70                     | 0.81      | 0.41      | -0.7      | -1.2      |
| LC0048   | 1.87                     | 0.09           | 0.05      | -1.5      | -8.8      | 2.30                     | 0.12      | 0.06      | -1.3      | -7.8      |
| LC0049   | 2.30                     | 0.23           | 0.12      | -0.8      | -3.4      | 4.3                      | 0.65      | 0.32      | 1.6       | 3.3       |
| LC0050   | 2.50                     | 0.4            | 0.2       | -0.5      | -0.6      | 2.4                      | 0.40      | 0.20      | -1.1      | -1.6      |
| 0        | 1.50                     | 0.75           | 0.37      | -2.1      | -3.4      | 1.77                     | 0.89      | 0.44      | -2.0      | -3.1      |

|          | Sample A<br>Aflatoxin B1 |                |       |         |         |                       |       | Sample B<br>Aflatoxin B1 | L       |         |
|----------|--------------------------|----------------|-------|---------|---------|-----------------------|-------|--------------------------|---------|---------|
| Lab code | Sample<br>A<br>µg/ kg    | U lab<br>(k=2) | u lab | Z-Score | ζ-score | Sample<br>B<br>µg/ kg | U lab | u lab                    | Z-Score | ζ-score |
| LC0052   | 2.35                     | 0.73           | 0.36  | -0.7    | -1.2    | 2.95                  | 0.92  | 0.46                     | -0.4    | -0.5    |
| LC0053   | 3.00                     | 1.2            | 0.6   | 0.3     | 0.3     | 3.2                   | 1.30  | 0.65                     | 0.0     | 0.0     |
| LC0054   | 2.00                     | 0.27           | 0.14  | -1.3    | -4.8    | 2.2                   | 0.30  | 0.15                     | -1.4    | -5.5    |
| LC0055   | 1.80                     | 0.13           | 0.07  | -1.6    | -8.7    | 2.19                  | 0.16  | 0.08                     | -1.4    | -7.9    |
| LC0056   | 2.31                     | 0.35           | 0.17  | -0.8    | -2.5    | 2.37                  | 0.36  | 0.18                     | -1.2    | -4.0    |
| LC0057   | 2.81                     | 1.52           | 0.76  | 0       | 0       | 3.36                  | 1.77  | 0.89                     | 0.2     | 0.2     |

**Table 4.** Summary of the results for aflatoxin B2 in sample A & aflatoxin B2, aflatoxin G1 and aflatoxin G2 in sample B.

|  |       | Sample A<br>Aflatoxin B2 | Sample B<br>Aflatoxin B2 | Sample B<br>Aflatoxin G1 | Sample B<br>Aflatoxin G2 |
|--|-------|--------------------------|--------------------------|--------------------------|--------------------------|
| Number of results                      |       | 54                       | 54                       | 54                       | 54                       |
| Range of results                       | µg/kg | 0.09 - 0.63              | 0 - 1.9                  | 0 - 3                    | 0 - 2.1                  |
| Robust mean of results of participants | µg/kg | 0.18                     | 0.67                     | 1.56                     | 0.63                     |

**Table 5.** Analytical results for aflatoxins B2 in sample A and aflatoxin B2, aflatoxin G1 and aflatoxin G2 in sample B. Laboratories that didn't report any results are marked as "No result", all other results are the figures that the laboratories reported.

|          | Sample A<br>Aflatoxin B2 | Sample B<br>Aflatoxin B2 | Sample B<br>Aflatoxin G1 | Sample B<br>Aflatoxin G2 |
|----------|--------------------------|--------------------------|--------------------------|--------------------------|
| Lab code | µg/kg                    | µg/kg                    | µg/kg                    | µg/kg                    |
| LC0002   | 0.15                     | 0.61                     | 1.31                     | 0.57                     |
| LC0003   | 0.22                     | 0.66                     | 1.62                     | 0.91                     |
| LC0004   | 0                        | 0                        | 0                        | 0                        |
| LC0005   | < 0.50                   | 0.9                      | 1.8                      | 0.7                      |
| LC0006   | < 0.50                   | 0.72                     | 1.37                     | 0.72                     |
| LC0007   | No result                | No result                | No result                | No result                |
| LC0008   | < 0.20                   | 0.78                     | 2.04                     | 0.73                     |
| LC0009   | < 2.50                   | < 2.50                   | < 2.50                   | < 2.50                   |
| LC0010   | < 0.10                   | 0.48                     | 1.44                     | 0.56                     |
| LC0011   | < 0.20                   | 0.66                     | 1.59                     | 0.53                     |
| LC0012   | < 0.50                   | 0.8                      | 1.7                      | 0.8                      |
| LC0013   | No result                | No result                | No result                | No result                |
| LC0014   | 0.17                     | 0.79                     | 1.62                     | 0.7                      |
| LC0015   | < 5.00                   | < 5.00                   | 1.57                     | < 5.00                   |
| LC0016   | 0.16                     | 0.7                      | 1.33                     | 0.65                     |
| LC0017   | < 0.25                   | 0.6                      | 1.44                     | 0.58                     |
| LC0018   | 0.63                     | 1.28                     | 2.17                     | 0.7                      |
| LC0019   | 0.17                     | 0.75                     | 1.67                     | 0.7                      |
| LC0020   | No result                | No result                | No result                | No result                |
| LC0022   | 0.17                     | 0.64                     | 1.56                     | 0.6                      |
| LC0023   | 0.09                     | 0.48                     | 1.32                     | 0.39                     |
| LC0024   | 0.16                     | 0.66                     | 1.39                     | 0.61                     |
| LC0025   | 0.19                     | 0.65                     | 1.46                     | 0.58                     |
| LC0026   | 0.12                     | 0.56                     | 1.69                     | 0.51                     |
| LC0027   | 0.11                     | 0.47                     | 1.1                      | 0.3                      |

|          | Sample A<br>Aflatoxin B2 | Sample B<br>Aflatoxin B2 | Sample B<br>Aflatoxin G1 | Sample B<br>Aflatoxin G2 |
|----------|--------------------------|--------------------------|--------------------------|--------------------------|
| Lab code | µg/kg                    | µg/kg                    | µg/kg                    | µg/kg                    |
| LC0028   | 0.51                     | 1.88                     | 2.22                     | 0.55                     |
| LC0029   | 0.19                     | 0.67                     | 1.19                     | 0.37                     |
| LC0030   | 0.11                     | 0.44                     | 1.23                     | 0.38                     |
| LC0031   | 0.13                     | 0.06                     | 0.11                     | 0.02                     |
| LC0032   | 0.2                      | 0.9                      | 1.9                      | 0.7                      |
| LC0033   | < 0.20                   | 0.49                     | 1.28                     | 0.48                     |
| LC0035   | 0.19                     | 0.74                     | 1.49                     | 0.61                     |
| LC0036   | 0.27                     | 0.81                     | 1.7                      | 0.65                     |
| LC0037   | 0.19                     | 0.63                     | 1.49                     | 0.42                     |
| LC0038   | 0.28                     | 0.88                     | 1.88                     | 0.67                     |
| LC0039   | 0.1                      | 0.36                     | 0.85                     | 0.38                     |
| LC0040   | 0.35                     | 0.99                     | 2.56                     | 1.77                     |
| LC0041   | 0.14                     | 0.54                     | 1.23                     | 0.44                     |
| LC0042   | 0.16                     | 0.82                     | 1.88                     | 0.77                     |
| LC0043   | < 0.30                   | 1.92                     | 2.23                     | 2.06                     |
| LC0044   | 0.19                     | 0.82                     | 2.08                     | 0.77                     |
| LC0045   | < 0.20                   | 0.6                      | 1.3                      | 0.5                      |
| LC0046   | 0.29                     | 1                        | 2                        | 1.1                      |
| LC0047   | < 0.30                   | 0.66                     | 1.47                     | 0.58                     |
| LC0048   | 0.15                     | 0.54                     | 1.35                     | 0.57                     |
| LC0049   | 0.12                     | 0.83                     | 3                        | 1.2                      |
| LC0050   | 0.22                     | 0.62                     | 1.73                     | 0.52                     |
| LC0051   | < 0.20                   | 0.43                     | 0.87                     | 0.33                     |
| LC0052   | 0.16                     | 0.69                     | 1.75                     | 0.6                      |
| LC0053   | < 1.00                   | < 1.00                   | 1.8                      | < 2.50                   |
| LC0054   | 0.14                     | 0.49                     | 1.25                     | 0.47                     |
| LC0055   | 0.11                     | 0.52                     | 1.06                     | 0.39                     |
| LC0056   | 0.2                      | 0.74                     | 1.24                     | 0.64                     |
| LC0057   | 0.27                     | 0.77                     | 1.64                     | < 1.00                   |

For the 2 laboratories using ELISA the z-scores for total aflatoxins were calculated with the assigned value for aflatoxin B1 and the robust mean from all the other participants for aflatoxin B2, aflatoxin G1 and aflatoxin G2 for indicative purposes.

|             | Sample A<br>AFLA total |         | Sample<br>AFLA to | e B<br>Stal |  |
|-------------|------------------------|---------|-------------------|-------------|--|
| Lab<br>code | µg/kg                  | Z-Score | µg/kg             | Z-Score     |  |
| LC0021      | 4.14                   | 1.8     | 5.10              | -0.7        |  |
| LC0034      | 3.15                   | 0.3     | 4.52              | 1.1         |  |





Figure 1. Sigmoidal plot of individual results reported for aflatoxin B1 in sample A

Figure 2. Sigmoidal plot of individual results reported for aflatoxin B1 in sample B



Laboratory

# 7.4 Evaluation of questionnaire

All laboratories that reported result filled in the questionnaire. The summary of the answers is presented in Annex 8.

The main technique used to determine aflatoxins in defatted peanut powder is HPLC-FLD, followed by LC-MS. Two laboratories used ELISA and reported only total aflatoxins.

# 8 Conclusions

Fifty-four laboratories participated in this study and, for the determination of aflatoxin B1, the performance of most of the participants based on their z-scores was satisfactory (> 92 %).

Most laboratories reported acceptable recoveries for all aflatoxins. (Annex 9)

In line with observations of previous PTs organised by the EURL for Mycotoxins, performance of the laboratory based on their  $\zeta$ -scores were not as satisfactory as when based on the z-scores, which indicates that the respective participants should review their uncertainty estimation.

## References

AMC, Analytical Methods Committee. Robust statistics: a method of coping with outliers. Technical brief, No 6, Apr 2001. Available from: http://www.rsc.org/pdf/amc/brief6.pdf

Ardic, M., Karakaya, Y., Atasever, M., Durmaz, H. Determination of aflatoxin B(1) levels in deep-red ground pepper (isot) using immunoaffinity column combined with ELISA. Food Chemistry and Toxicology, 2008; 46:1596-1599. doi: 10.1016/j.fct.2007.12.025

EN 14123:2003 Determination of aflatoxin B1 and the sum of aflatoxin B1, B2, G1 and G2 in peanuts, pistachios, figs and paprika powder. European Committee for Standardization.

IARC. Monographs on the evaluation of carcinogenic risks to humans. Some naturally occurring substances: food items and constituents, heterocyclic aromatic amines and mycotoxins. 1993; 56: 1-599.

ISO 13528:2015. Statistical methods for use in proficiency testing by interlaboratory comparisons.

ISO/IEC 17043:2010. Conformity assessment - General requirements for proficiency testing.

Commission Regulation (EC) No 882/2004 of the European Parliament and of the council of 29 April 2004 on official controls performed to ensure the verification of compliance with feed and food law, animal health and animal welfare rules. L191/1. http://eur-

lex.europa.eu/LexUriServ.do?uri=CONSLEG:2004R0882:20060525:EN:PDF

Directive 2002/32/EC of the European Parliament and of the Council of 7 May 2002 of undesirable substances in animal feed. L140/10 http://eur-lex.europa.eu/resource.html?uri=cellar:aca28b8c-bf9d-444f-b470-268f71df28fb.0004.02/DOC 1&format=PDF

Lamberty A., Schimmel H., Pauwels J. The study of the stability of reference materials by isochronous measurements, Fresenius Journal of Analytical Chemistry, 1998; 360, 359-361. doi: 10.1007/s002160050711

Mackay, L.G., Taylor, P., Myros, R.B., Hearn, R., King, B. High accuracy analysis by isotope dilution mass spectrometry using an iterative exact matching technique. Accreditation and Quality Assurance, 2003; 8: 191-194. doi: 10.1007/s00769-003-0622-z

Shim, W.B., Yang, Z.Y., Kim, J.S., Kang, S.J., Woo, G.J., Chung, Y.C., Eremin, S.A., Chung D.H. Journal of Microbiology and Biotechnology, 2007; 17: 1629-1637.

Thompson, M. Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing. Analyst, 2000; 125: 385-386. doi: 10.1039/B000282H

Thompson, M., Ellison, S.L.R., and Wood, R. The International Harmonised Protocol for the Proficiency Testing of Analytical Chemistry Laboratories. Pure and Applied Chemistry, 2006; 78: 145–196.

doi: 10.1351/pac200678010145

Wild, C.P., Turner, P.C. The toxicology of aflatoxins as a basis for public health decisions. Mutagenesis, 2002; 17: 471-481. doi: <u>10.1093/mutage/17.6.471</u>

Yao, H., Hruska, z., Di Mavungu, J.D. Developments in detection and determination of aflatoxins. World Mycotoxin Journal, 2015; 8: 181-191. doi: <u>10.3920/WMJ2014.1797</u>

# List of abbreviations and definitions

| AFLA     | Aflatoxin                                      |
|----------|--|
| ELISA    | Enzyme-Linked Immuno Sorbent Assay             |
| EC       | European Commission                            |
| EN       | European Standard                              |
| EU       | European Union                                 |
| EURL     | European Union Reference Laboratory            |
| FLD      | Fluorescence detector                          |
| HPLC     | High-Performance Liquid Chromatography         |
| IARC     | International Agency for Research on Cancer    |
| ISO      | International Organization for Standardization |
| JRC      | Joint Research Centre                          |
| LC-MS/MS | Liquid chromatography-tandem mass spectrometry |
| LOD      | Limit of Detection                             |
| LOQ      | Limit of Quantification                        |
| NRL      | National Reference Laboratory                  |
| OCL      | Official Control Laboratory                    |
| РТ       | Proficiency Test                               |
|          |  |

# List of figures

| Figure 1. | Sigmoidal | plot of | individual | results | reported | for | aflatoxin | Β1 | in sample | Α | . 12 |
|-----------|-----------|---------|------------|---------|----------|-----|-----------|----|-----------|---|------|
| Figure 2. | Sigmoidal | plot of | individual | results | reported | for | aflatoxin | Β1 | in sample | В | . 12 |

# List of tables

| Table 1: Participating laboratories  |
|--|
| Table 2.         Summary statistics of the results for aflatoxin B1 in sample A & sample B   |
| <b>Table 3.</b> Analytical results, z-and zeta scores for aflatoxin B1 in sample A and sample B.Colour code: green- satisfactory performance, yellow – questionable performance, red –unsatisfactory performance |
| <b>Table 4.</b> Summary of the results for aflatoxin B2 in sample A & aflatoxin B2, aflatoxin G1 andaflatoxin G2 in sample B   |
| <b>Table 5.</b> Analytical results for aflatoxins B2 in sample A and aflatoxin B2, aflatoxin G1 andaflatoxin G2 in sample B  |
| <b>Table 6.</b> Analytical results and z-scores for aflatoxins in sample A and in sample B. Colourcode: yellow – questionable peformance, red – unsatisfactory performance                                       |

## Annexes

#### Annex 1. Announcement- opening of registration

From: JRC IRMM EURL MYCOTOX Sent: Friday, July 01, 2016 2:22 PM To: JRC IRMM EURL MYCOTOX Cc: STROKA Joerg (JRC-GEEL); CUBERO LEON Elena (JRC-GEEL); BOUTEN Katrien (JRC-GEEL) Subject: PT 2016 Aflatoxins in defatted peanut powder

#### Dear colleagues

On behalf of the European Union Reference Laboratory for Mycotoxins (EURL Mycotoxins), I have the pleasure to announce the opening for registration to the inter-laboratory comparison/proficiency test on the determination of aflatoxins in defatted peanut powder.

According to Regulation (EC) No 882/2004 it is obligatory for EU National Reference Laboratories to participate. The main PT samples will be dispatched in October 2016, the mixed aflatoxins solution will be dispatched directly after the registration period.

The deadline for registration is 19<sup>th</sup> July 2016.

Technical details on the PT design:

Prior to the main PT (aflatoxins in defatted peanut powder) you will receive a vial with a mixed aflatoxins solution.

For all laboratories we will ask to measure the exact content of aflatoxins in the solution against their standards.

For laboratories that use LC with fluorescence detection you might remember that we asked some time ago to report fluorescence response factors for aflatoxins. We have gathered sufficient data that supports the assumption that a single calibration with AfB1 is sufficient for compliance judgement (under given conditions). To further substantiate this finding we ask you to use the calibrants in the vial (a dilution scheme will be supplied) to determine response factors prior the main PT. This means injecting the solution at least in duplicate around 2 and 1 month before.

When reporting your results we will – in addition – recalculate individual results and compare the overall performance of results (not your individual) for both scenarios. We hope that these findings might trigger some discussion if a single point as well as a single aflatoxin calibration is "fit-for-purpose".

For NRLs the participation is free of charge.

The participation fee for official food control laboratories is 270 Euro per participant. The full participation fee is payable upon dispatch of the test samples. Enrolled control laboratories will be contacted for payment details upon registration.

Confidentiality of results is guaranteed.

Please register at the following link: https://ec.europa.eu/eusurvey/runner/MYCO-PT-2016-Aflatoxins

Thank you in advance for your consideration.

EURL Mycotoxins Operating Manager

#### **EURL Mycotoxins**



European Commission Joint Research Centre Retieseweg 111 B-2440 Geel, Belgium Phone: +32 (0)14 571231 Fax: +32 (0)14 571343

"The views expressed are purely those of the writer and may not in any circumstances be regarded as stating an official position of the European Commission."

Interlaboratory comparison for the determination of aflatoxins in defatted peanut powder.

Fields marked with \* are mandatory.

On behalf of the European Union Reference Laboratory for Mycotoxins (EURL Mycotoxins), I have the pleasure to announce the opening for registration to the inter-laboratory comparison/proficiency test on the determination of aflatoxins in peanut powder.

According to Regulation (EC) No 882/2004 it is obligatory for EU National Reference Laboratories to participate. The main PT samples will be dispatched in October 2016, the mixed aflatoxins solution will be dispatched directly after the registration period.

The deadline for registration is 19th July 2016.

Technical details on the PT design:

Prior to the main PT (aflatoxins in defatted peanut powder) you will receive a vial with a mixed aflato ns solution

For all laboratories we will ask to measure the exact content of aflatoxins in the solution against their standards.

For laboratories that use LC with fluorescence detection you might remember that we asked some time ago to report fluorescence response factors for atlatorins. We have gathered sufficient data that supports the assumption that a single calibration with ABI is sufficient for compliance judgement (under given conditions). To further substantiate this finding we ask you to use the calibrants in the vial (a distingtion scheme will be supplied) to determine response factors prior the main PT. This means injecting the solution at least in duplicate around 2 and 1 month before.

When reporting your results we will – in addition – receivulate individual results and compare the overall performance of results (incl your individual) for both scenarios. We hope that these findings might trigger some discussion if a single point as well as a single affatoxin calibration is '19 for-purpose'.

For NRLs the participation is free of charge.

The participation fee for official food control laboratories is 270 Euro per participant. The full participation fee is payable upon dispatch of the test samples. Enrolled control laboratories will be contacted for payment details upon registration.

Confidentiality of results is guaranteed.

Thank you in advance for your consideration.

EURL Mycotoxins Operating Manager

\*Contact person

Additional comments

1

3

2

4

| econd contact pe | rson |  |  |
|------------------|------|--|--|
|                  |      |  |  |
| rganisation      |      |  |  |
|                  |      |  |  |
| dress            |      |  |  |
|                  |      |  |  |
| stoole           |      |  |  |
|                  |      |  |  |
| a.               |      |  |  |
|                  |      |  |  |
| suntry           |      |  |  |
| lephone number   | ,    |  |  |
|                  |      |  |  |
|                  |      |  |  |
|                  |      |  |  |
| nail address     |      |  |  |
|                  |      |  |  |

HPLC-FLD LCMS cther

### Annex 2. Homogeneity study

#### Sample A

AFB1:

Content: 2.11 ug/kg Analysis of Variance Table Variance estimates for repeatability and reproducibility Response: AM\_obs Mean Sq replicates Variance sd UNIT 0.042502 4 0.001927 0.044 Repeatability 0.034793 0.190 Reproducibility 0.034793 0.036721 0.190 Test for sufficient homogeneity acc. to Harmonized Protocol PTs s\_sam^2 s\_an^2 s\_all^2 F1 F2 critical Homogenous s\_an/s\_p<0.5 Results 0.001927 0.034793 0.01939335 1.88 1.01 0.0716 TRUE TRUE AFB2: Content: 0.38 ug/kg Analysis of Variance Table Variance estimates for repeatability and reproducibility Response: AM\_obs Mean Sq replicates Variance sd UNIT 0.024864 4 0.0024964 0.05 Repeatability 0.014878 0.0148780 0.12 Reproducibility 0.014878 0.0006290064 1.88 1.01 0.0162 TRUE FALSE Sample B

```
AFB1:
Content: 3.10 ug/kg
Analysis of Variance Table
Variance estimates for repeatability and reproducibility
Response: AM obs
                        Mean Sq replicates Variance
                                                                    sd

        Mean Sq replicates variance
        sd

        UNIT
        0.267033
        4
        0.057147
        0.24

        Repeatability
        0.038444
        0.038444
        0.20

                                                  0.038444 0.20
Reproducibility
                                                    0.095591 0.31
Test for sufficient homogeneity acc. to Harmonized Protocol PTs
s_sam^2 s_an^2 s_all^2 F1 F2 critical Homogenous s_an/s_p<0.5
Results 0.057147 0.038444 0.04186116 1.88 1.01 0.118 TRUE TRUE
                                                                                         TRUE
AFB2:
Content: 0.81 ug/kg
Analysis of Variance Table
Variance estimates for repeatability and reproducibility
Response: AM_obs

        Mean Sq replicates
        Variance
        sd

        UNIT
        0.0211136
        4
        0.0033609
        0.058

        Repeatability
        0.0076702
        0.0076702
        0.0088

Reproducibility
                                                      0.0110310 0.110
Test for sufficient homogeneity acc. to Harmonized Protocol PTs
                                              s_all^2 F1 F2 critical Homogenous s_an/s_p<0.5
              s_sam^2 s_an^2
Results 0.0033609 0.0076702 0.002857972 1.88 1.01 0.0131
                                                                                            TRUE
                                                                                                          TRUE
AFG1:
Content: 1.11 ug/kg
Analysis of Variance Table
Variance estimates for repeatability and reproducibility
Response: AM_obs
                        Mean Sq replicates Variance
                                                                       sd

        Mean Sq replicates
        Variance
        sd


        UNIT
        0.039226
        4
        0.0075762
        0.087

        Repeatability
        0.008922
        0.0089218
        0.094

        Reproducibility
        0.0164980
        0.130

Test for sufficient homogeneity acc. to Harmonized Protocol PTs
s_sam^2 s_an^2 s_all^2 F1 F2 critical 
Results 0.0075762 0.0089218 0.005367028 1.88 1.01 0.0191
                                                                   F2 critical Homogenous s_an/s_p<0.5
                                                                                       TRUE
                                                                                                                  TRUE
AFG2:
Content: 0.33 ug/kg
Analysis of Variance Table
Variance estimates for repeatability and reproducibility
Response: AM obs
                         Mean Sq replicates Variance
Mean Sq
UNIT 0.0111001
Repeatability 0.0092896
                                                                         sd
                      0.0111001 4 0.0004526 0.021
0.0092896 0.096
Reproducibility
                                                      0.0097422 0.099
Test for sufficient homogeneity acc. to Harmonized Protocol PTs
s_sam^2 s_an^2 s_all^2 F1 F2 critical Homogenous s_an/s_p<0.5
Results 0.0004526 0.0092896 0.0004743684 1.88 1.01 0.0103 TRUE FALSE
```

## Annex 3. Stability study



#### Annex 4. Sample accompanying letter

EUROPEAN COMMISSION DIRECTORATE-GENERAL JOINT RESEARCH CENTRE Directorate F - Health, Consumers and Reference Materials Food & Feed Compliance

Geel, 18 October 2016

Ref. Ares(2016)5963147 - 18/10/2016

2016 Proficiency testing of National Reference Laboratories (NRLs) and Official Control Laboratories (OCLs) on the determination of aflatoxins in defatted peanut powder

Dear Participant,

Please read the following information carefully before starting any analysis. The 2016 PT on aflatoxins in peanut powder aims to assess the content in two naturally contaminated test samples ("Sample A" and "Sample B").

The materials are shipped at ambient temperature. After receipt freeze the samples immediately at -18 °C until the analysis is performed.

Please confirm the parcel's receipt by e-mail upon arrival, by using the "Materials Receipt Form". If any material is damaged, request new material immediately.

Report the recovery corrected value ( $\mu$ g/kg), including your recovery (%) and measurement uncertainty ( $\mu$ g/kg) for a coverage factor of 2 (k=2).

Participants that indicated they will use HPLC-FLD detection and have received the standard solution at the end of July should inject the diluted standard solution 3 times when analysing the samples (Set 3). Report the area of all injections of the standards in the results table (3 sets of 3 injections).

For participants only using LC-MS there is a 1 mL standard solution included in the package. The concentrations of the aflatoxins are: Aflatoxin B1: 80.1 ng/mL; Aflatoxin B2: 27.2 ng/mL; Aflatoxin G1: 84.3 ng/mL and Aflatoxin G2: 28.1 ng/mL.

Data generated by the participants will be collected by using the software RingDat, supplementary to ProLab software, that has been used for professional data handling and statistical analyses of interlaboratory tests results. You should have received two files attached to this email for reporting the results. The instructions on how to use the software RingDat can be found in the Annex 1 at the end of this document.

The deadline for reporting the PT results is the 9th December 2016.

Please keep in mind that collusion is contrary to professional scientific conduct and serves only to nullify the benefits of proficiency tests to costumers, accreditation bodies and analysts alike.

Should you need any further assistance, please do not hesitate to contact us.

Katrien Bouten (on behalf of the Operating Manager of the EURL Mycotoxins)

E-mail: <u>JRC-EURL-MYCOTOX@ec.europa.eu</u> Tel: +32-14-571231

Cc: Frans Verstraete, Hendrik Emons, Joerg Stroka

Retleseweg 111, B-2440 Geel - Belgium. Telephone: (32-14) 571 229. https://ec.europa.eu/iro/ E-mail: JRC-EURL-MYCOTOX@ec.europa.eu

#### Annex 1: Instructions for reporting the results using RingDat.

1. Download the updated version of the data entry program (called RingDat) free from the QuoData web page using following link: <u>http://quodata.de/ringdat\_en.php</u>

User: ringdat

Password: prolabdata

Alternatively, in case you already have Ringdat, you can update it via the "Programm-update" button.

2. Save the two lab specific files with the extension "\*.Lab" and "\*.LA2" attached to this email in the same folder as RingData.exe.

The name of each laboratory and the samples are codified by the software, so that each participant will receive samples with unique codified numbers (i.e., LC0001).

- The "\*.LA2" file contains information about the participant laboratory name and laboratory code.
- The "\*.LAB" file is unique to each laboratory and contains information about the samples and measurands that have to be analysed and reported.

3. Start the RingDat.exe program and open "\*.LAB" file to access your workspace.

- The first tab contains detailed information about the laboratory (Lab details).
- The second tab contains a table for entering the results for every measurand/ sample combination (Measured values).
- The third tab contains a general questionnaire (Questions and Answers).

4. Fill in the results table ("Measured values") with your data. Please find below some captures of the RingDat pages that have been configured for this PT.

#### Figure 1 - Capture of the "Measured Values" page

| Ow: Tarwell  | 9 🗸 🗸 🖓 🗖           | vi er  | 1 | FV        | <b>e</b> | 🔌 Transa | e dista |
|--|---------------------|--------|---|-----------|----------|----------|---------|
|  | 1.                  |        |   |           | _        |          |         |
| ala da Baser I al e  | , Janqua e          | • ~ ~  |   |           |          |          |         |
| ing test   | канат               | COLOR! |   |           |          |          |         |
| ingraat treet  | * 1                 |        |   |           |          |          |         |
| 5. gal -   |                     |        |   |           |          |          |         |
| Sauza -  | Hanselle<br>(F.) El |        |   | the set y |          |          | ún a    |
| SELVA  | 426                 | 8.1 U  |   |           |          |          |         |
| Sale / A   | 4.24                | 1.9    |   |           |          |          |         |
|  |                     |        |   |           |          |          |         |
| Jane : J   | 4.511               | 12.2   |   |           |          |          |         |
| January J  | 42815               | 121/2  |   |           |          |          |         |
| Jane e J   | 415.61              | 121/2  |   |           |          |          |         |
| Service 1  | 4.562               | 10160  |   |           |          |          |         |
| All descent percent  |                     |        |   |           |          |          |         |
| Verlament (with)   |                     |        |   |           |          |          |         |
| Net desire tiped of  |                     |        |   |           |          |          |         |
| And shows the part of  |                     |        |   |           |          |          |         |
| State geber  |                     |        |   |           |          |          |         |
| Cristian distant   |                     |        |   |           |          |          |         |
| Cristian Gerrar  |                     |        |   |           |          |          |         |
| SU MURLI MALAZ   |                     |        |   |           |          |          |         |
| SU Vakor gears   |                     |        |   |           |          |          |         |
| <ul> <li>SUT Market Styring</li> <li>SUT Market Styring</li> </ul> |                     |        |   |           |          |          |         |
| (d) Ankoo kemin  |                     |        |   |           |          |          |         |
| Add denote property  |                     |        |   |           |          |          |         |
|  |                     |        |   |           |          |          |         |
| 340 denotes permit-  |                     |        |   |           |          |          |         |

2

5. Afterwards, please fill in the questionnaire on the next tab.

Figure 2 - Capture of the "Questions and Answers" page

| (et 🚺 544              | 144 💙 Tride i na 🖾 Tratecel 🤤 Help   | <b>ХРаднон кин</b>  |
|------------------------|--|---|
| azik   Yeasured va     | <sub>1866</sub> - Ценке во ала бисконо   |   |
| c. Qe 🛛                | Ouestion   | Answei  |
| * Ул торома            | Harvanshysiqa caraciyyin khana ayarakwa inalah -aran<br>wenye                            | <ul> <li>The sent the inprogram</li> <li>def the approximation of the sent set of the</li></ul> |
| C Notero vez           | Places specify the reference of the analytical method used (e.g. noulified 2014) 22;     |   |
| 2. In demented         | Fuch when pixthes net an open ensurement of your feb destroy?                            |   |
| d Zalona tail          | leµna ratinta aranin l°  | © ∖ı<br>⊖ ∽⊷  |
| C Standa Jo            | Place specify that source of the standard used for calibration                           | 51200   |
| if. Fatra di u kuloren | a Natural areas dhaalaan laddii too baarah lasaad?                                       | Coste / ne nen d?<br>Coste / ater and?<br>Coste / ater ater ater  |
| 7 It other I           | It other presserves ()   | 1900a403  |
| C Water                | Whet was the percentage water used for extraction  |   |
| C Carolaiato           | What was the e-traction to vertite sample ratio used during entraction $(m_{\rm e}/g)^2$ |   |
| IC Extraction          | Minet Hastilleren, astronolde?   | Culending<br>Ostering<br>Cultationex<br>Cultationex   |
| 1" It other 2          | if odret, please specile   |   |
| 12 Extraction ine      | Sensitives the exception time?   |   |
| 15 Светкр              | Nenet goe of sender dearrup did (dur se?   | C 42<br>SFE<br>SHEDOOD<br>C dias  |
| 13 H JI. 5 9           | 11 A   |   |

6. After finishing the input, save the file using the button on the top of the window menu. You can change the inputs after saving the file as long as you haven't pushed "Finish input" button. At the end finalise the data entry by pressing the "Finish input" button.

7. Send both the "\*.LAB" and "\*.LA2" files back to us by e-mail to our functional mail box: JRC-EURL-MYCOTOX@ec.europa.eu

8. Should you want to correct some of your entries after finishing the input, you must use the original \*.LAB file downloaded from the email and introduce all the information again (results and answers to the questionnaire).

3

#### Annex 5. Material receipt form



EUROPEAN COMMISSION DIRECTORATE-GENERAL JOINT RESEARCH CENTRE Directorate F - Health, Consumers and Reference Materials Food & Feed Compliance

Geel, 18 October 2016

## PROFICIENCY TESTING MATERIALS RECEIPT FORM

Name:

Institute:

Member State:

## NOTE: STORE MATERIAL IN A FREEZER AT -18 °C!

Please ensure that the items listed below have been received undamaged, and then check the relevant statement:

| Date of receipt                        |            |
|--|------------|
| All items have been received undamaged | YESŤ / NOŤ |
| If NO, please list damaged items:      |            |

Contents of the parcel:

a) 2 test materials (defatted peanut powder) for analysis:

- Sample A

- Sample B
- b) Material receipt form

Please e-mail the completed form to:

Katrien Bouten

JRC-EURL-MYCOTOX@ec.europa.eu

Your Signature / Stamp here:


Retleseweg 111, B-2440 Geel - Belgium. Telephone: (32-14) 571 231. https://ec.europa.eu/iro/ E-mail: JRC-EURL-MYCOTOX/dec.europa.eu


# Annex 6. Questionnaire

| No | Cue                | Question   | Answer   |
|----|--------------------|--|--|
| 1  | Samples per year   | How many samples does your laboratory analyse for aflatoxins per year?             | Less than 50 samples per year<br>50-250 samples per year<br>250-1000 samples per year<br>More than 1000 samples per year |
| 17 | Reference          | Please specify the reference of the analytical method used (e.g. modified EN14123) |  |
| 31 | Implemented        | For how long is this method implemented in your laboratory?                        |  |
| 2  | Accredited         | Is your method accredited?   | No<br>Yes  |
| 3  | Standards          | Please specify the source of the standard used for calibration:                    |  |
| 4  | Extraction solvent | What was the extration solvent used?   | water/ methanol?<br>water/ acetonitrile?<br>other?   |
| 32 | If other 1         | If other, please specify   |  |
| 41 | Water              | What was the percentage water used for extraction                                  |  |
| 5  | Sample ratio       | What was the extraction solvent to sample ratio used during extraction (mL/g)?     |  |
| 6  | Extraction         | What was the extraction mode?  | blending<br>shaking<br>ultraturrax   |
| 33 | If other 2         | if other, please specify   |  |
| 7  | Extraction time    | What was the extraction time?  |  |
| 8  | Clean-up           | What type of sample clean-up did you use?  | IAC<br>SPE<br>Mycosep<br>other   |
| 34 | If other 3         | If other, please specify   |  |
| 9  | IAC                | If you used immuno affinity columns , please specify the brand:                    | R-Biopharm<br>Vicam<br>Neogen<br>other   |
| 35 | If other 4         | If other, please specify   |  |
| 10 | Overnight stop 1   | During the analysis did you need to include an overnight stop?                     | No<br>Yes  |
| 11 | Overnight stop 2   | If yes, please state for which samples and at what state of the analysis:          |  |
| 37 | MP                 | What mobile phase did you use?   |  |

| No | Cue                             | Question  | Answer   |
|----|---------------------------------|---|--|
| 12 | Detection                       | What type of detection method did you use?  | HPLC_FLD<br>LC-MS<br>ELISA<br>other  |
| 36 | If other 5                      | If other, please specify  |  |
| 21 | Ex                              | If you used HPLC-FLD, please specify the detector settings - Extinction                                   |  |
| 38 | Em                              | If you used HPLC-FLD, please specify the detector settings - Emission                                     |  |
| 39 | Bandwidth                       | If you used HPLC-FLD, please specify the detector settings - Bandwidth                                    |  |
| 16 | LCMS                            | If you used LCMS, please state the ionisation mode and transitions used for the analysis                  |  |
| 14 | ISTD (1)                        | If you used LCMS, did you used a stable istope labelled internal standard?                                | for aflatoxin B1<br>for aflatoxin B2<br>for aflatoxin G1<br>for aflatoxin G1<br>No |
| 15 | ISTD (2)                        | At what stage was the internal standard added?<br>  | before extraction<br>after extraction  |
| 23 | ELISA                           | If you used ELISA, please specify the brand<br>   |  |
| 13 | Derivatisation                  | Which derivatisation method was applied?  | KOBRA<br>PBPB<br>UV<br>other   |
| 40 | If other 6                      | If other, please specify  |  |
| 18 | Glassware                       | Did you used acid glassware?  | No<br>Yes  |
| 19 | Daylight                        | Was protection against daylight applied?  | No<br>Yes  |
| 20 | Problems                        | Did you encounter any problems during the analysis?   | No<br>Yes  |
| 25 | If yes, what were the specified | If yes, what were the specified problems and to which samples they apply?                                 |  |
| 26 | Observations                    | Did you notice any unsual observations which, however,<br>did not seem to have any effect on the results? | No<br>Yes  |
| 27 | If yes, describe                | If yes, what were the observations and to which samples do they apply?                                    |  |
| 28 | Instructions                    | Did you find the instructions distributed for this PT adequate?   | No<br>Yes  |
| 29 | If no, describe                 | If no, which parts do you think can be improved?  |  |
| 30 | Comments                        | Any other comments you wish to address?   |  |

## Annex 7. Aflatoxin B1 Z- & ζ-score graphs





Sample

## Annex 8. Experimental details

| Lab<br>Code | How many samples does your<br>laboratory analyses for<br>aflatoxins per year? | Reference of the analytical method used                           | For how long is this method<br>implemented in your<br>laboratory? | Is your<br>method<br>accredited? | Source of the standard used for calibration |
|-------------|---|---|---|----------------------------------|---|
| LC0002      | 250-1000 samples per year   | F/0329 (VEMS)   | 21 years  | yes                              | Sigma                                       |
| LC0003      | 250-1000 samples per year   | modified MSZ EN ISO 17375:2006                                    | 6 years   | yes                              | Biopur MIX1 aflatoxin<br>ROMERLABS          |
| LC0004      | 50-250 samples per year   | internal document SOP 10575.1                                     | since 2012  | yes                              | Sigma Aldrich                               |
| LC0005      | More than 1000 samples per<br>year  | in house LC method  |   | yes                              | Coring                                      |
| LC0006      | 250-1000 samples per year   | Internal method   |   | yes                              | Sigma-Aldrich                               |
| LC0007      | 50-250 samples per year   | inhouse method based on ISO17375:2006                             | since 2012  | yes                              | SIGMA                                       |
| LC0008      | 250-1000 samples per year   | In-house method   | more than 10 years  | yes                              | SigmaAldrich                                |
| LC0009      | 250-1000 samples per year   | In house method   |   | yes                              | Biopure                                     |
| LC0010      | 50-250 samples per year   | modified EN14123  | 12 years  | no                               | Sigma                                       |
| LC0011      | 50-250 samples per year   | ISEN14123-2007  | 9   | yes                              | IRMM  |
| LC0012      | 250-1000 samples per year   | in house method   | one year  | no                               |   |
| LC0014      | Less than 50 samples per year   | according EN 14123  | 8 years   | yes                              | Romer Labs "Biopure"                        |
| LC0015      | 50-250 samples per year   | modified AB Sciex Technical Note                                  | 3 years   | yes                              | Romer                                       |
| LC0016      | 50-250 samples per year   | in-house method   | since 2003  | yes                              |   |
| LC0017      | 250-1000 samples per year   | Modified EN 14123   | 10 years +  | no                               | Sigma-Aldrich                               |
| LC0018      | Less than 50 samples per year   | In-house validated metod  | 2007  | yes                              | Biopure                                     |
| LC0019      | 250-1000 samples per year   | AOAC 991.31 (2010) Final Action)                                  | Since 1995  | yes                              | Sigma Aldrich                               |
| LC0021      | Less than 50 samples per year   | manufacturer`s (Neogen) instruction for use of ELISA kit          | 5 years   | yes                              |   |
| LC0022      | 50-250 samples per year   | N/A   | 10 Years  | yes                              | R-Biopharm Rhone                            |
| LC0023      | 250-1000 samples per year   | In house method   | 8 years   | yes                              | Trilogy                                     |
| LC0024      | 50-250 samples per year   | R-Biopharm Rhône, Immunoaffinity columns, Instructions for<br>use | 14 years (since 2002)   | yes                              | Romer Labs (Biopure)                        |
| LC0025      | 50-250 samples per year   | EN 14123:2007   | 5 years   | yes                              | Biopure                                     |
| LC0026      | 50-250 samples per year   | modified EN14123  | 2 year  | no                               | R-Biopharm                                  |
| LC0027      | 250-1000 samples per year   | modified EN14123  | 20 years  | yes                              | Romer Labs                                  |
| LC0028      | 50-250 samples per year   | In house method   | 20 years  | yes                              | Sigma                                       |
| LC0029      | 50-250 samples per year   |   |   | yes                              |   |
| LC0030      | 50-250 samples per year   | AOAC Official method 991.31                                       | over 10 years   | yes                              | Biopure, Romer labs                         |
| LC0031      | 50-250 samples per year   | R-Biopharm Easi-extract aflatoxin                                 | 10 years  | yes                              | R-Biopharm                                  |

| Lab<br>Code | How many samples does your<br>laboratory analyses for<br>aflatoxins per year? | Reference of the analytical method used  | For how long is this method<br>implemented in your<br>laboratory? | Is your<br>method<br>accredited? | Source of the standard used for calibration  |
|-------------|---|--|---|----------------------------------|--|
| LC0032      | 250-1000 samples per year   | Inhouse method, modified EN14123   | >25 years   | yes                              |  |
| LC0033      | Less than 50 samples per year   | inhouse method   | >10 years   | yes                              | Sigma  |
| LC0034      | 250-1000 samples per year   | Protocol Elisa of R-Biopharm   | 8 years   | yes                              | Kit Elisa of R-Biopharm  |
| LC0035      | 50-250 samples per year   | EN 14123.  | 10 years  | yes                              | Sigma-Aldrich  |
| LC0036      | 50-250 samples per year   |  | 18 years  | yes                              | SIGMA  |
| LC0037      | 50-250 samples per year   | Sulyok M et al. 2006. Rapid Communication in Mass<br>Spectrometry 20, 2649-2659  | 2010  | yes                              | biopure, Romer Labs Diagnostic<br>GmbH, Tulln - Austria; office-<br>europe@romerlabs.com |
| LC0038      | 50-250 samples per year   | Internal method  |   | yes                              | Sigma Aldrich  |
| LC0039      | 50-250 samples per year   | UNI EN 14123:2008  | 20  | yes                              | BIOPURE  |
| LC0040      | 50-250 samples per year   |  |   | yes                              |  |
| LC0041      | 250-1000 samples per year   | own method   | since 2013  | yes                              | Supelco standard mix   |
| LC0042      | Less than 50 samples per year   | modified EN14123 with LC-MS method   | 2   | yes                              | romer  |
| LC0043      | 250-1000 samples per year   | EN 14123   | 15 years  | yes                              | SIGMA  |
| LC0044      | 50-250 samples per year   | modified EN ISO 16050  | 6 years   | yes                              | LGC Standards  |
| LC0045      | 250-1000 samples per year   | EN 14123   | >10 years   | yes                              | Biopure, aflatoxin mix   |
| LC0046      | Less than 50 samples per year   |  | 6 years   | yes                              |  |
| LC0047      | 50-250 samples per year   | modified EN 14123  | 6 years   | yes                              | Biopure Romer labs   |
| LC0048      | 250-1000 samples per year   | § 64 LFGB L23.05-2 and § 64 LFGB L15.00-2  | 20 years  | yes                              | Romer Labs Diagnostic GmbH   |
| LC0049      | Less than 50 samples per year   |  |   | no                               |  |
| LC0050      | Less than 50 samples per year   | Application notes for quantitative determination of aflatoxins in various foods by immunoaffinity chromatography, Rhone-<br>diagnostic Technologies Ltd. | 15 Years  | no                               | R-biopharm   |
| LC0051      | 50-250 samples per year   |  | 13 years  | yes                              | Romer  |
| LC0052      | 250-1000 samples per year   | modified EN14123   | 10 years  | yes                              | SUPELCO  |
| LC0053      | 50-250 samples per year   | /  | 5 years   | yes                              | Biopure  |
| LC0054      | 50-250 samples per year   | EN 14123   | 2007  | yes                              | LGC Standards  |
| LC0055      | 50-250 samples per year   | In house.  | 10 years  | yes                              | R-Biopharm   |
| LC0056      | More than 1000 samples per<br>year  | Modified EN ISO 16050  | 12  | yes                              | Sigma Aldrich (A6636; A9887;<br>A0138; A0263)  |
| LC0057      | 250-1000 samples per year   | Zachariasova M, Analytica Chimica Acta, 2010, 662, 51-61   | 4 years   | yes                              | Sigma Aldrich  |

| Lab<br>Code | Extraction solvent                                | % Water in extraction solvent | What was the solvent to sample ratio used during extraction (mL/g) | Extraction mode      | Extraction time                           |
|-------------|---|-------------------------------|--|----------------------|---|
| LC0002      | water/ methanol                                   | 70                            | 250ml/50g (5ml/1g)   | blending             | 1 minute                                  |
| LC0003      | water/acetone                                     |                               | 5  | shaking              | 60 min                                    |
| LC0004      | 0,1 % HCOOH in Water+ACN                          | 50                            | 5  | shaking              | 20 min                                    |
| LC0005      | water/ acetonitrile                               | 20                            | 5  | shaking              | 90 min                                    |
| LC0006      | water/ methanol                                   | 20                            | 1.6  | blending             | 2 min                                     |
| LC0007      | acetone/water                                     | 15                            | 250ml/50g  | shaking              | 30min                                     |
| LC0008      | water/ acetonitrile                               | 20                            | 5 mL/g   | shaking              | 45 minutes                                |
| LC0009      | 1 % acetic acid in acetonitril                    |                               | 10   | shaking              | 30 min                                    |
| LC0010      | water/ methanol                                   | 20                            | 4  | ultraturrax          | 3 minutes                                 |
| LC0011      | water/ methanol                                   | 20                            | 80   | ultraturrax          | 2minutes                                  |
| LC0012      | water/ acetonitrile                               | 30                            | 5  | shaking              | 2 h                                       |
| LC0014      | water/methanol/hexane containing<br>16.7 g/l NaCl | 13                            | 6  | shaking              | 45 min                                    |
| LC0015      | water/ acetonitrile                               | 20                            | 8/2  | shaking              | 20 mins                                   |
| LC0016      | water/ methanol                                   | 20                            |  | ultraturrax          | 3 minutes + 15' shaking                   |
| LC0017      | water/methanol/acetonitrile                       | 40                            | 10mL/1g  | ultraturrax          | 6 minutes                                 |
| LC0018      | water/ acetonitrile                               | 16                            | 100 ml/25 g  | shaking              | 30 min                                    |
| LC0019      | water/ methanol                                   | 30                            | 10   | blending             | 2 min                                     |
| LC0021      | water/ methanol                                   | 30                            | 5/1  | shaking              | approx. 50 minutes (3 minutes of shaking) |
| LC0022      | water/ methanol                                   | 30                            | 125/25   | blending             | 1 minute                                  |
| LC0023      | water/ methanol                                   | 40                            | 5/1  | shaking              |   |
| LC0024      | water/ methanol                                   | 40                            |  | shaking              | 1 hour                                    |
| LC0025      | water/ methanol                                   | 20                            | 4  | shaking              | 20  |
| LC0026      | water/ methanol                                   | 30                            | 125/25   | shaking              | 15  |
| LC0027      | water/ methanol                                   | 20                            | 4  | shaking              | 30 min                                    |
| LC0028      | water/ acetonitrile                               | 40                            | 0.4g of sample per 1ml of extraction solution                      | Blending and shaking | 2min blending 30min shaking               |
| LC0029      | water/ acetonitrile                               | 40                            | 5  | shaking              | 1 hour                                    |
| LC0030      | water/ methanol                                   | 37,5                          | 4  | shaking              | 20 minutes                                |
| LC0031      | water/ methanol                                   | 20                            | 2  | ultraturrax          | 2 minutes                                 |
| LC0032      | water/ acetonitrile                               |                               |  | ultraturrax          | 3 minutes                                 |
| LC0033      | water/acetone                                     | 25                            | 4,2  | blending             | 5 min                                     |
| LC0034      | water/ methanol                                   | 30                            | 10 ml methanol/distilled water 70 % with 2 g sample                | shaking              | 15 minut                                  |
| LC0035      | water/ methanol                                   | 20                            | 4:1  | blending             | 3 min                                     |

| Lab<br>Code | Extraction solvent                                  | % Water in extraction solvent | What was the solvent to sample ratio used during extraction (mL/g) | Extraction mode                | Extraction time |
|-------------|---|-------------------------------|--|--------------------------------|-----------------|
| LC0036      | water/ methanol                                     | 30                            | 6.7  | ultraturrax                    | 3 min           |
| LC0037      | acetonitrile/water/formic acid<br>(79/20/1) (v:v:v) | 20                            | 2 (20 ml extraction solvent, 10 g sample weight)                   | shaking                        | 0.5 h           |
| LC0038      | water/ methanol                                     | 64                            | 25:5   | First Whirlemixer then shaking | 5 sek + 15 min  |
| LC0039      | water/ methanol                                     | 20                            | 4/1  | blending                       | 3 min           |
| LC0040      | water/ methanol                                     | 58                            |  | shaking                        |                 |
| LC0041      | water/ methanol                                     | 30                            | 4 ml/g   | shaking                        | 30 min          |
| LC0042      | water/ methanol                                     | 20                            | 4  | shaking                        | 60 min          |
| LC0043      | water/methanol/hexane                               | 20                            | 4 ml/g   | blending                       | 3 minutes       |
| LC0044      | water/ methanol                                     | 30                            | 5  | shaking                        | 5 minutes       |
| LC0045      | methanol/water/cyclohexane                          | 16                            | 5  | blending                       | 3 min           |
| LC0046      | water/ acetonitrile                                 | 40                            | 4  | shaking                        | 60 min          |
| LC0047      | water/ methanol                                     | 20                            | 10ml/1g  | blending                       | 3min            |
| LC0048      | water/ methanol                                     | 40                            | 100 mL /25 g   | shaking                        | 45 min          |
| LC0049      | water/ methanol                                     | 20                            | 12,5   | blending                       | 2 min.          |
| LC0050      | water/ methanol                                     | 40                            | 250 ml/ 50 g   | ultraturrax                    | 3 minutes       |
| LC0051      | water/ methanol                                     | 20                            | 8  | shaking                        | 45min           |
| LC0052      | water/ methanol                                     | 20                            | 20 ml / 5 g  | shaking                        | 2 hours         |
| LC0053      | acetone/isopropanol/water/acetic<br>acid            | 40                            | 4.5  | shaking                        | 60 min          |
| LC0054      | water/ methanol                                     | 40                            | 5  | shaking                        | 60 min          |
| LC0055      | water/ methanol                                     | 30                            | 62.5/12.5  | blending                       | 2 minutes       |
| LC0056      | water/ methanol                                     | 20                            | 4ml/g  | shaking                        | 60min           |
| LC0057      | acetonitrile/formic acid                            |                               | 5 mL/g   | shaking                        | 30 min.         |

| Lab<br>Code | What type of<br>sample<br>clean-up did<br>you use? | Brand immuno affinity<br>column | Did you need to include an overnight stop?  | What mobile phase did you used?   |
|-------------|--|---------------------------------|---|---|
| LC0002      | IAC  | R-Biopharm                      | Νο  | water:acetonitrile:methanol (60:10:30) containing 119mg of KBr and 350 $\mu l$ of 4M nitric acid                          |
| LC0003      | IAC  | ROMERLABS                       | No  | Water/Acn/MeOH 68/16/16<br>+119mg KBr/l   |
| LC0004      | QuEChERS   | -                               | No  | A=0,1 % HCOOH in H2O; B=0,1 % HCOOH + 1mMHCOONH4 in MeOH  |
| LC0005      | none   |                                 | No  |   |
| LC0006      | IAC  | R-Biopharm                      | No  |   |
| LC0007      | IAC  | LC-Tech                         | No  | water/methanol/acetonitrile   |
| LC0008      | IAC  | Vicam                           | No  | KBr, HNO3, methanol, acetonitrile, water  |
| LC0009      |  |                                 | No  | 1 % formic acid/5 mM ammonium formate in water/ methanol  |
| LC0010      | IAC  | R-Biopharm                      | No  | Water:Methanol:Acetonitrile (6:3:2)   |
| LC0011      | IAC  | Vicam                           | All samples. Extracts were stored in the freezer prior to dilution with water and IAC cleanup                           | 45:55 Methanol : Water  |
| LC0012      | IAC  | Vicam                           | Νο  | 0.1 % Formic acid in water and 0.1 % Formic acid in acetonitrile  |
| LC0014      | IAC  | LC-Tech                         | All samples were stored at 4 °C after extraction for 18 hours.  | mixture of water/methanol/acetonitrile 54/29/17 (v/v/v) containing 0.0119 % (w/v) KBr and 0.01 % (v/v) nitric acid (65 %) |
| LC0015      | none   |                                 | Νο  | 5mM Ammonium Acetate containing 0.5 % acetic acid in water/methanol   |
| LC0016      | IAC  | R-Biopharm                      | after extraction for all samples  | water/ACN/methanol (62/19/19) + HNO3 + KBr  |
| LC0017      | IAC  | R-Biopharm                      | Νο  | $27$ % Acetonitrile/58 % water/15 % methanol +120mg potassium bromide and 350 $\mu L$ 4N Nitric acid per litre            |
| LC0018      | MultiSep   |                                 | For both samples: Sample extraction and clean-up day 1. Frozen extracts overnight. Evaporation, resolution, HPLC day 2. | Water:Acetonitrile:Methanol (900:180:240). 119 mg KBr+350 µl 4M HNO3 per liter.   |
| LC0019      | IAC  | Vicam                           | No  | Acetonitrile (20 %) :H2O (60 %) : MEOH (20 %)   |
| LC0021      |  |                                 | No  |   |
| LC0022      | IAC  | R-Biopharm                      | No  | Water/Methanol/Acetonitrile 65/17.5/17.5 + 119mg KBr + 350µl 4M<br>HNO3/litre   |
| LC0023      | IAC  | Romer                           | Νο  | Water 57:acetonitrile 5:methanol 38   |
| LC0024      | IAC  | R-Biopharm                      | Νο  | Methanol-acetonitrile-water with addition of HNO3 and KBr   |
| LC0025      | IAC  | Romer Labs AlfaStar R           | No  | Water/MeOH/HNO3/KBr   |

| Lab<br>Code | What type of<br>sample<br>clean-up did<br>you use? | Brand immuno affinity<br>column | Did you need to include an overnight stop?  | What mobile phase did you used?   |
|-------------|--|---------------------------------|---|---|
| LC0026      | IAC  | R-Biopharm                      | Νο  | Water-Methanol-Acetonitrile<br>(62:22:16)+KBr+HNO3  |
| LC0027      | IAC  | R-Biopharm                      | Νο  | acetonitril/methanol/aceticacid2 % (18/27/55; v/v/v), isocratic   |
| LC0028      | IAC  | R-Biopharm                      | Νο  | water acetonitrile methanol mix acidifed with acetic acid   |
| LC0029      | IAC  | R-Biopharm                      | Νο  | Acetonitrile + Methanol + water (18 + 27 + 55)<br>120 mg KBr, 350 µL 4 N salpetersyre / pr. litre mobilfase                                   |
| LC0030      | IAC  | Vicam                           | No  | Tetrahydrofuran - Water 21:79   |
| LC0031      | IAC  | R-Biopharm                      | No  | 60:40 % v/v Water:Methanol 119mg/L KBr & 350uL/L 4M Nitric acid   |
| LC0032      | IAC  | R-Biopharm                      | No  | Water : acetonitrile : methanol, 56 : 30: 14  |
| LC0033      | GPC, SX-3  |                                 | No  | water/methanol/acetonitrile : 65/17/18  |
| LC0034      |  |                                 | No  |   |
| LC0035      | IAC  | R-Biopharm                      | No  | Water:acetonitrile:methanol   |
| LC0036      | IAC  | Vicam                           | No  | Water/Methanol/Acetonitrile   |
| LC0037      | Mycosep  |                                 | Νο  | mobile A: water, 0.1 % HCOOH, 5mM NH4OOCH<br>mobile B: methanol, 0.1 % HCOOH, 5mM NH4OOCH   |
| LC0038      |  | R-Biopharm                      | No  | Water, methanol, acetonitrile, acetic acid  |
| LC0039      | IAC  | R-Biopharm                      | No  | H2O:MeOH:ACN 54:29:17   |
| LC0040      | IAC  | R-Biopharm                      | No  |   |
| LC0041      | IAC  | Romer                           | No  | MeOH/ACN/Water/HNO3/KBr   |
| LC0042      | IAC  | Vicam                           | No  | methanol/water with ammonium acetate and acetic acid  |
| LC0043      | IAC  | Мусоѕер                         | Νο  | 410 grammes de méthanol + 265 grammes d'acétonitrile + 1200 grammes d'eau + 700 $\mu$ l d'acide nitrique 4M + 240 mg de bromure de potassium. |
| LC0044      | IAC  | R-Biopharm                      | No  | water:acetonitrile:methanol 60:20:20  |
| LC0045      | IAC  | R-Biopharm                      | No  | water/ACN/MeOH 570:140:290  |
| LC0046      | IAC  | Vicam                           | No  | acetonitrile:water  |
| LC0047      | IAC  | LC Tech                         | Νο  | 560ml water, 25mg KBr, 280ml methanol, 160ml acetonitrile, 100ul conc. nitric acid  |
| LC0048      | IAC  | Romer Labs Diagnostic<br>GmbH   | Νο  | Methanole/Acetonitrile/Water with KBr and HNO3  |
| LC0049      | IAC  | Vicam                           | After elution of aflatoxins from IAC , samples were blown to dryness and stored in the freezer overnight. | Solvent A:Methanol: Water 20:80 and Solvent B: Methanol: Water 80:20  |
| LC0050      | IAC  | R-Biopharm                      | All samples, stored in fridge overnight prior to running on HPLC  | 62:19:19 Water:Methanol:Acetonitrile + 119mg KBr + 350ul 4M nitric acid   |
| LC0051      | IAC  | R-Biopharm                      | No  | 0.2 % acetic acid/ : MeOH/AcN(260/110)  |

| Lab<br>Code | What type of<br>sample<br>clean-up did<br>you use? | Brand immuno affinity<br>column | Did you need to include an overnight stop?   | What mobile phase did you used?  |
|-------------|--|---------------------------------|--|--|
| LC0052      | IAC  | R-Biopharm                      | For the all samples: one day extraction and the second day<br>passing through IAC and HPLC-FLD | KBr/HNO3 solution/ACN/MeOH: 6/2/3  |
| LC0053      | QuEChERS   |                                 | Νο   | water / MeOH<br>10mMammonium formate / 0.01 % (v/v) formic acid          |
| LC0054      | IAC  | R-Biopharm                      | No   | Water, Methanol, Acetonitrile, Nitric Acide, Potassium bromide,          |
| LC0055      | IAC  | R-Biopharm                      | No   | Water/Acetonitrile/Methanol with KBr                                     |
| LC0056      | IAC  | Protealmmun                     | No   | Water:MeCN:MeOH=3:1:1  |
| LC0057      | Bondesil C18                                       |                                 | No   | A: 5mM ammonium formate/0,2 %formic acid<br>B: Methanol/0,2 %formic acid |

| Lab<br>Code | What type of detection method did you use? | Emission<br>(nm) | Excitation<br>(nm) | Bandwidth | At what stage was the internal standard<br>added | Which derivatisation method was applied? |
|-------------|--|------------------|--------------------|-----------|--|--|
| LC0002      | HPLC-FLD                                   | 455nm            | 365nm              | 5nm       |  | KOBRA                                    |
| LC0003      | HPLC-FLD                                   | 425 nm           | 362 nm             | 18 nm     |  | KOBRA                                    |
| LC0006      | HPLC-FLD                                   | 435 nm           | 350 nm             |           |  | Iode                                     |
| LC0007      | HPLC-FLD                                   | 435 nm           | 365 nm             | 4nm       |  | PBPB                                     |
| LC0008      | HPLC-FLD                                   | 430 nm           | 365 nm             |           |  | KOBRA                                    |
| LC0010      | HPLC-FLD                                   | 435 nm           | 365 nm             | 15nm      |  | PBPB                                     |
| LC0011      | HPLC-FLD                                   | 440 nm           | 360 nm             | 18        |  | UV                                       |
| LC0014      | HPLC-FLD                                   | 448 nm           | 362 nm             | 15 nm     |  | KOBRA                                    |
| LC0016      | HPLC-FLD                                   | 440 nm           | 362 nm             |           |  | KOBRA                                    |
| LC0017      | HPLC-FLD                                   | 434 nm           | 364 nm             | 20nm      |  | KOBRA                                    |
| LC0018      | HPLC-FLD                                   | 450 nm           | 365 nm             |           |  |  |
| LC0019      | HPLC-FLD                                   | 420 nm           | 360 nm             | 20nm      |  | Saturated Iodine Solution                |
| LC0022      | HPLC-FLD                                   | 430 nm           | 365 nm             |           |  | KOBRA                                    |
| LC0023      | HPLC-FLD                                   | 430              | 360                | 15        |  | KOBRA                                    |
| LC0024      | HPLC-FLD                                   | 425              | 362                | 18        |  | KOBRA                                    |
| LC0025      | HPLC-FLD                                   | 440              | 365                | 18        |  | KOBRA                                    |
| LC0026      | HPLC-FLD                                   | 430 nm           | 360 nm             |           |  | KOBRA                                    |
| LC0027      | HPLC-FLD                                   |                  |                    |           |  | РВРВ                                     |
| LC0028      | HPLC-FLD                                   | 455 nm           | 365 nm             | 1         |  | KOBRA                                    |

| Lab<br>Code | What type of detection method did you use? | Emission<br>(nm) | Excitation<br>(nm) | Bandwidth             | At what stage was the internal standard<br>added | Which derivatisation method was applied? |
|-------------|--|------------------|--------------------|-----------------------|--|--|
| LC0029      | HPLC-FLD                                   | 435 nm           | 365 nm             | 15nm                  |  | KOBRA                                    |
| LC0030      | HPLC-FLD                                   | 454              | 362                |                       |  | Saturated iodine solution                |
| LC0031      | HPLC-FLD                                   | 455              | 362                | peakwidth?<br>>0.2min |  | KOBRA                                    |
| LC0032      | HPLC-FLD                                   | 434 nm           | 364 nm             | 2nm                   |  |  |
| LC0033      | HPLC-FLD                                   | 425 nm           | 365 nm             | 4 nm                  |  | online derivatisation Pickering          |
| LC0035      | HPLC-FLD                                   | 435 nm           | 362 nm             | Agilent 1100          |  | KOBRA                                    |
| LC0036      | HPLC-FLD                                   | 435              | 365                | -                     |  | РВРВ                                     |
| LC0038      | HPLC-FLD                                   | 440 nm           | 360 nm             | 20 nm                 |  | РВРВ                                     |
| LC0039      | HPLC-FLD                                   | 435              | 365                | 18                    |  | РВРВ                                     |
| LC0040      | HPLC-FLD                                   | 420              | 360                |                       |  |  |
| LC0041      | HPLC-FLD                                   | 440 nm           | 362 nm             | 10                    |  | KOBRA                                    |
| LC0043      | HPLC-FLD                                   | 435 nm           | 365 nm             | 15 nm                 |  | KOBRA                                    |
| LC0044      | HPLC-FLD                                   | 435 nm           | 365 nm             | 15 nm                 |  | TFA                                      |
| LC0045      | HPLC-FLD                                   | 435              | 365                | 18                    |  | PBPB                                     |
| LC0046      | HPLC-FLD                                   | 450 nm           | 365 nm             |                       |  | precolumn derivatisation with TFA        |
| LC0047      | HPLC-FLD                                   | 435              | 366                | gain 4                |  | KOBRA                                    |
| LC0048      | HPLC-FLD                                   | 438 nm           | 368 nm             |                       | before extraction                                | KOBRA                                    |
| LC0050      | HPLC-FLD                                   | 425nm            | 362nm              |                       |  | KOBRA                                    |
| LC0052      | HPLC-FLD                                   | 435 nm           | 365 nm             | >0.2 min              |  | KOBRA                                    |
| LC0054      | HPLC-FLD                                   | 435 nm           | 465 nm             | -                     |  | KOBRA                                    |
| LC0055      | HPLC-FLD                                   | 430              | 365                |                       |  | KOBRA                                    |
| LC0056      | HPLC-FLD                                   | 435nm            | 365nm              |                       | before extraction                                |  |

| Lab<br>Code | What type of detection method did you use? | If you used ELISA, please specify the brand | At what stage was the internal standard added |
|-------------|--|---|---|
| LC0021      | ELISA                                      | Neogen Veratox Aflatoxin HS                 | before extraction                             |
| LC0034      | ELISA                                      | R-Biopharm                                  |   |

| Lab<br>Code | What type of detection method did you use? | Ionisation mode & transitions   | Did you use a stable isotope labelled<br>internal standard?  | At what stage was the internal standard added |
|-------------|--|---|--|---|
| LC0004      | LC-MS                                      | ESI+. AFG2: 331/201, AFG1: 329/243, AFB2:<br>315/259, AFB1: 313/241   | no   | before extraction                             |
| LC0005      | LC-MS                                      |   | aflatoxin B1<br>aflatoxin B2<br>aflatoxin G1<br>aflatoxin G2 | after extraction                              |
| LC0009      | LC-MS                                      | Aflatoxin B1 (ql)- ESI+ m/z 313>128.1<br>Aflatoxin B1 (qn)- ESI+ m/z 313>285.2<br>Aflatoxin B2 (ql)- ESI+ m/z 315.1>259.2<br>Aflatoxin B2 (qn)- ESI+ m/z 315.1>287.2<br>Aflatoxin G1 (ql)- ESI+ m/z 329>200<br>Aflatoxin G1 (qn)- ESI+ m/z 329>243.2<br>Aflatoxin G2 (ql)- ESI+ m/z 331.1>245.2<br>Aflatoxin G2 (qn)- ESI+ m/z 331.1>313.2  | no   |   |
| LC0012      | LC-MS                                      | B1 (ESI +) 313.1>285.0 and 313.1>241.0<br>B2 (ESI +) 315.1>287.0 and 315.1>259.0<br>G1 (ESI +) 329.1>243.1 and 329.1>311.0<br>G2 (ESI +) 331.1>313.1 and 331.1>245.1  | no   |   |
| LC0015      | LC MSMS                                    | Precursor Product 1 Product2 Mode<br>AFB1 313.0 285.2 128.1 Po<br>AFB2 315.1 287.2 259.5 Po<br>AFG1 329.0 243.2 200.0 Po<br>AFG2 331.1 313.2 245.2 Po   | aflatoxin B1<br>aflatoxin B2<br>aflatoxin G1<br>aflatoxin G2 | after extraction                              |
| LC0037      | LC-MS                                      | AFB1 312.2 (Mw) 313.1 (Precursor) [M+H]+<br>(adduct) 285.0/241.0 (Products) 21/41 (CE) 3/3<br>(CAV) positiv (polarity) 5.80 (retention)<br>AFB2 314.2 (Mw) 315.1 (Precursor) [M+H]+<br>(adduct) 287.0/258.9 (Products) 21/29 (CE) 3/3<br>(CAV) positiv (polarity) 5.60 (retention)<br>AFG1 328.2 (Mw) 329.1 (Precursor) [M+H]+<br>(adduct) 243.0/200.1 (Products) 25/41 (CE) 3/3<br>(CAV) positiv (polarity) 5.40 (retention)<br>AFG2 330.2 (Mw) 331.1 (Precursor) [M+H]+<br>(adduct) 313.0/245.1 (Products) (Products) 21/25<br>(CE) 3/3 (CAV) positiv (polarity) 5.20 (retention) | aflatoxin B1<br>aflatoxin B2<br>aflatoxin G1<br>aflatoxin G2 | after extraction                              |
| LC0042      | LC-MS                                      | ESI positive<br>AfB1: 313.0>285.2<br>AfB2: 315.0>287.1<br>AfG1: 329.0>311.1<br>AfG2: 331.0>313.1<br>13CAfB1: 330.1>301.1  | aflatoxin B1   | after extraction                              |

| Lab<br>Code | What type of detection method did you use? | Ionisation mode & transitions  | Did you use a stable isotope labelled<br>internal standard?  | At what stage was the internal standard added |
|-------------|--|--|--|---|
| LC0049      | LC-MS                                      |  | no   |   |
| LC0051      | LC-MS                                      | ESI+<br>313; 315; 329;330  | no   |   |
| LC0053      | LC-MS                                      | ESI+<br>AfB1: 313 > 241<br>AfG1: 329.1 > 242.9<br>AfB2: 315.1 > 258.9<br>AfG2: 331.1 > 217                   | aflatoxin B1<br>aflatoxin B2<br>aflatoxin G1<br>aflatoxin G2 | after extraction                              |
| LC0057      | LC-MS                                      | ESI+<br>AFB1: 313/285, 313/241<br>AFB2: 315/287, 315/259<br>AFG1: 329/311, 329/115<br>AFG2: 331/313, 331/115 | aflatoxin B1<br>aflatoxin B2<br>aflatoxin G1<br>aflatoxin G2 | after extraction                              |

| Lab<br>Code | Did you use acid washed glassware? | Was protection against daylight applied? | Did you encounter any problems during the analysis?                              |
|-------------|------------------------------------|--|--|
| LC0002      | Yes                                | Yes                                      | No   |
| LC0003      | Yes                                | Yes                                      | No   |
| LC0004      | No                                 | Yes                                      | No   |
| LC0005      | No                                 | Yes                                      | No   |
| LC0006      | No                                 | Yes                                      | No   |
| LC0007      | Yes                                | Yes                                      | No   |
| LC0008      | Yes                                | Yes                                      | No   |
| LC0009      | No                                 | No                                       | No   |
| LC0010      | No                                 | Yes                                      | No   |
| LC0011      | Yes                                | Yes                                      | No   |
| LC0012      | Yes                                | Yes                                      | No   |
| LC0014      | Yes                                | Yes                                      | No   |
| LC0015      | No                                 | No                                       | No   |
| LC0016      | Yes                                | No                                       | when adding PBS, the solution got cloudy, so we had to dilute with water instead |

| Lab<br>Code | Did you use acid washed glassware? | Was protection against daylight applied? | Did you encounter any problems during the analysis?   |
|-------------|------------------------------------|--|---|
| LC0017      | Νο                                 | Yes                                      | We were unable to use our accredited method extraction of 60 % acetonitrile as two<br>layers were formed on centrifuging. Further portions of the samples (A and B) were<br>obtained and a modified procedure used (extraction solvents as above). It was then<br>necessary to filter the extract again before IA column. |
| LC0018      | Yes                                | Yes                                      | No  |
| LC0019      | No                                 | Yes                                      | No  |
| LC0021      |                                    | Yes                                      | No  |
| LC0022      | Yes                                | Yes                                      | No  |
| LC0023      | No                                 | Yes                                      | No  |
| LC0024      | No                                 | Yes                                      | No  |
| LC0025      | No                                 | Yes                                      | sample extracts were turbid and after had problems with SPE   |
| LC0026      | No                                 | Yes                                      | No  |
| LC0027      | No                                 | Yes                                      |   |
| LC0028      | Yes                                | Yes                                      | No  |
| LC0029      | No                                 | Yes                                      | No  |
| LC0030      | No                                 | Yes                                      | No  |
| LC0031      | Yes                                | Yes                                      | Peak shape of stored standard deteriorated badly.<br>Kobra cell membrane was used up during analysis  |
| LC0032      | Yes                                | Yes                                      | No  |
| LC0033      | No                                 | Yes                                      | No  |
| LC0034      | No                                 | Yes                                      | No  |
| LC0035      | No                                 | Yes                                      | No  |
| LC0036      | No                                 | Yes                                      | No  |
| LC0037      | No                                 | Yes                                      | No  |
| LC0038      | No                                 | Yes                                      | Set 3 standard: injection of the standard at the 3. round gave of unknown reason splitting peaks. No other samples are affected. Reanalysis of the standard did not solve the problem. No results are therefore entered.  |
| LC0039      | Yes                                | Yes                                      | No  |
| LC0040      |                                    |  |   |
| LC0041      | Yes                                | Yes                                      | No  |
| LC0042      | No                                 | Yes                                      | No  |
| LC0043      | Yes                                | Yes                                      | No  |
| LC0044      | No                                 | Yes                                      | No  |
| LC0045      | Yes                                | Yes                                      | No  |
| LC0046      | No                                 | No                                       | No  |
| LC0047      | No                                 | Yes                                      | No  |

| Lab<br>Code | Did you use acid washed glassware? | Was protection against daylight applied? | Did you encounter any problems during the analysis?  |
|-------------|------------------------------------|--|--|
| LC0048      | No                                 | Yes                                      | No   |
| LC0049      | Νο                                 | No                                       | We have had exceptionally low recoveries for all four aflatoxins. we have repeated 3 times extraction method. Usually our acceptable recoveries between 70-110 percent. But this time they have been recoveries between 50 to 30 percent on all three occasions. The reported results are corrected for recovery |
| LC0050      | No                                 | Yes                                      | No   |
| LC0051      | No                                 | No                                       | Unusual behaviour of that kind of samplewe have got very dirty extract   |
| LC0052      | Yes                                | Yes                                      | No   |
| LC0053      | No                                 | Yes                                      | No   |
| LC0054      | No                                 | Yes                                      | No   |
| LC0055      | No                                 | Yes                                      | No   |
| LC0056      | No                                 | Yes                                      | No   |
| LC0057      | Yes                                | Yes                                      | No   |

| Lab<br>Code | Did you notice any unusual observations which, however,<br>did not seem to have any effect on the results? | Did you find the instructions distributed for this PT adequate? Which parts can be improved?   | Comments   |
|-------------|--|--|--|
| LC0002      | No   | yes  | No   |
| LC0003      | No   | yes  |  |
| LC0004      | No   | Instructions for handling with 1 mL standard solution in case of LC-MS.  | Thank you for the opportunity to participate   |
| LC0005      | No   | It was not clear from the instruction sheet, what to do<br>with the standard solution measuring LC-MS. I had to<br>check this separately by e-mail.<br>Furthermore it is unclear if to enter the data with dot or<br>comma. We used comma! |  |
| LC0006      | No   | yes  | set 1 analysed 08/09/2016<br>set 2 analysed 07/10/2016<br>set 3 analysed 21/11/2016<br>The reported data for set 1, 2 and 3 were multiplied by<br>5 to take in account an extra dilution 1/5 to get our<br>calibration range |
| LC0007      | It took longer than usual for both extracted solutions to pass through the IAC columns                     | yes  |  |

| Lab<br>Code | Did you notice any unusual observations which, however,<br>did not seem to have any effect on the results?   | Did you find the instructions distributed for this PT adequate? Which parts can be improved? | Comments  |
|-------------|--|--|---|
| LC0008      | No   | yes  | The vial with standard was not injected upon arrival but<br>was diluted and run September 21st (together with the<br>EURL PT corn sample) and then run again November<br>1st (together with there two PT samples).<br>The results are corrected for recovery, 92 % B1, 93 %<br>B2, 91 % G1, 90 % G2.  |
| LC0009      | No   | The purpose of the standard solution was not completely clear for LC-MS users.               |   |
| LC0010      | Centrifugation and filtration using GFA filter paper was<br>required following dilution of extract with PBS, otherwise<br>sample would not pass through IAC column | yes  | The unit for the standard is in ng/ml. The edit feature<br>for the unit tab for the standard injections was not<br>enabled hence unit could not be inserted   |
| LC0011      | No   | yes  |   |
| LC0012      | No   | yes  |   |
| LC0014      | No   | yes  |   |
| LC0015      |  | yes  |   |
| LC0016      | No   | I'm not sure I have correctly understood the procedure to follow for the standard injections |   |
| LC0017      | See above, do not know if there was any effect on results.<br>Applied to both sample A and B   | yes  | We have analysed defatted peanut powder in the past<br>without the problems which we encountered. was it<br>totally defatted?   |
| LC0018      | No   | yes  | The EURL-std solution was diluted 1/2 before injection.<br>(Calibration curve std max 50 ng/ml)   |
| LC0019      | No   | yes  |   |
| LC0021      | sample was defatted which we do not make with routine<br>samples   | parts for reporting Total aflatoxin amount   | Our recovery factor was 103 % obtained with ELISA kit<br>(Aflatoxin Veratox HS, LOT:227494) for sum of total<br>aflatoksin (B1+B2+G1+G2) with the determined cross -<br>reactivity of a test (B1- 100 %, B2-18 %, G1-13 % and<br>G2-1 %).Out results of recovery corrected results for<br>sum of aflatoxins in sample named<br>"A" =4,14 µg/kg<br>"B" =5,10 µg/kg<br>LOD=0,5 µg/kg<br>LOQ=1,0 µg/kg<br>Linearity area=1-8 µg/kg |
| LC0022      | No   | yes  | No  |
| LC0023      | No   | yes  |   |
| LC0024      | No   | yes  |   |
| LC0025      | No   | yes  |   |
| LC0026      | Problems when using KOBRA CELL   | yes  |   |
| LC0027      | No   | yes  |   |

| Lab<br>Code | Did you notice any unusual observations which, however,<br>did not seem to have any effect on the results?                           | Did you find the instructions distributed for this PT adequate? Which parts can be improved?  | Comments   |
|-------------|--|---|--|
| LC0028      | No   | yes   |  |
| LC0029      | No   | yes   |  |
| LC0030      | No   | yes   |  |
| LC0031      | Νο   | yes   | The stored standard deteriorated badly under<br>conditions set, we would not have accepted data<br>generated by it.<br>Lack of sample meant sample could not be repeated as<br>would be standard practice following an analysis run<br>where the kobra membrane has been depleted during<br>run. |
| LC0032      | No   |   |  |
| LC0033      | No   | yes   |  |
| LC0034      | Νο   | yes   | The analysis method is ELISA and the result is the sum of B1+B2+G1+G2 (not just AFLA   |
| LC0035      | No   | yes   |  |
| LC0036      | Probably due to the fact that the samples are defatted, we had to make some modifications in the slurry preparation.<br>Both samples | yes   |  |
| LC0037      | No   | yes   |  |
| LC0038      | No   | yes   | You may in your description of the use of RingDat<br>mention, that after finishing there will be a possibility<br>to print out the input data  |
| LC0039      | No   | yes   |  |
| LC0040      | No   | yes   |  |
| LC0041      | No   | yes   | The results are calculated from three replicates' results and corrected with recoveries.   |
| LC0042      | No   | yes   | For sample A: AfB2 result below LOQ (without Uc),<br>AfG1 and AfG2 both below LOD (without Uc)   |
| LC0043      | No   | yes   |  |
| LC0044      | No   | yes   |  |
| LC0045      | No   | Otherwise well instructed but some more advice with data<br>input would have been helpful (such as how many<br>decimals were needed, if use comma or column as<br>decimal separator etc.) |  |
| LC0046      | No   | yes   |  |

| Lab<br>Code | Did you notice any unusual observations which, however,<br>did not seem to have any effect on the results? | Did you find the instructions distributed for this PT<br>adequate? Which parts can be improved?   | Comments  |
|-------------|--|---|---|
| LC0047      | Νο   | yes   | Recoveries for aflatoxin B1 -92 %<br>aflatoxin B2 - 95.2 %<br>aflatoxin G1 -91.4 %<br>aflatoxin G2 -102.4 %   |
| LC0048      | No   | yes   |   |
| LC0049      |  |   | Recovery: Aflatoxin B1 : 53 %<br>Aflatoxin B2: 51 %<br>Aflatoxin G1: 31 %<br>Aflatoxin G2: 21 %   |
| LC0050      | No   | yes   |   |
| LC0051      | No   | yes   |   |
| LC0052      | No   | There is no column for the recoveries.  |   |
| LC0053      | Νο   | No particular suggestions   | With the samples lower than our reporting limit I<br>reported a measurement uncertainty of 0 which is of<br>course meaningless.<br>If samples are considered blank we do not report a<br>measurement uncertainty but the software refused to<br>accept this. Our general uncertainty is 40 %,<br>irrespective of the level. |
| LC0054      | No   | yes   | -   |
| LC0055      | No   | yes   | Downloading the software for the results attachment was a problem.  |
| LC0056      | No   | yes   |   |
| LC0057      | Νο   | Instructions regarding standard solutions were a bit<br>misleading.<br>Reported peak areas of aflatoxins in SET1-SET3 were<br>obtained in our laboratory using LC-MS (NOT using FLD). |   |

#### Annex 9. Recoveries

| Laboratory code | Recovery % AFLA B1 | Recovery % AFLA B2 | Recovery % AFLA G1 | Recovery % AFLA G2 | Recovery % total AFLA's | Technique |
|-----------------|--------------------|--------------------|--------------------|--------------------|-------------------------|-----------|
| LC0002          | 86                 | 86                 | 88                 | 73                 |                         | HPLC      |
| LC0003          | 98.5               | 95                 | 97.8               | 87                 |                         | HPLC      |
| LC0004          | 101                | 86                 | 92                 | 92                 |                         | LCMS      |
| LC0005          | 95                 | 94                 | 96                 | 94                 |                         | LCMS      |
| LC0006          | 97.4               | 97                 | 100.2 / 94.9       | 97.7               |                         | HPLC      |
| LC0007          | 100                | -                  | -                  |                    |                         | HPLC      |
| LC0008          | 92                 | 93                 | 91                 | 90                 |                         | HPLC      |
| LC0009          | 91                 | 93                 | 93                 | 95                 |                         | LCMS      |
| LC0010          | 77                 | 85                 | 84                 | 80                 |                         | HPLC      |
| LC0011          | 90                 | 93                 | 90.1               | 87.2               |                         | HPLC      |
| LC0012          | 88.1 / 89.4        | 88.7 / 92.7        | 97.1 / 95.1        | 77.2 / 67.3        |                         | LCMS      |
| LC0013          | no result          | no result          | no result          | no result          |                         | no result |
| LC0014          | 78.1               | 81                 | 77                 | 74                 |                         | HPLC      |
| LC0015          | 100                |                    | 80                 |                    |                         | LCMS      |
| LC0016          | 81.1               | 80.2               | 80.6               | 62.7               |                         | HPLC      |
| LC0017          | 65.7               | 77                 | 71.3               | 77.2               |                         | HPLC      |
| LC0018          | 88.4               | 87.4               | 84                 | 86.1               |                         | HPLC      |
| LC0019          | 82.3               | 82.8               | 84.9               | 67.2               |                         | HPLC      |
| LC0020          | no result          | no result          | no result          | no result          |                         | no result |
| LC0021          |                    |                    |                    |                    | 103                     | ELISA     |
| LC0022          | 92.3               | 92.8               | 87.2               | 78.4               |                         | HPLC      |
| LC0023          | 97                 | 97                 | 97                 | 97                 |                         | HPLC      |
| LC0024          | 65                 | 46                 | 57                 | 36                 |                         | HPLC      |
| LC0025          | 60                 | 65                 | 70                 | 75                 |                         | HPLC      |
| LC0026          | 95                 | 93                 | 87                 | 70                 |                         | HPLC      |
| LC0027          | 90                 | 93                 | 99                 | 76                 |                         | HPLC      |
| LC0028          | 84.9               | 84                 | 46.2               | 55.8               |                         | HPLC      |
| LC0029          | 84                 | 91                 | 84                 | 69                 |                         | HPLC      |

| Laboratory code | Recovery % AFLA B1 | Recovery % AFLA B2 | Recovery % AFLA G1 | Recovery % AFLA G2 | Recovery % total AFLA's | Technique |
|-----------------|--------------------|--------------------|--------------------|--------------------|-------------------------|-----------|
| LC0030          | 82                 | 87.7               | 87.2               | 89.6               |                         | HPLC      |
| LC0031          | 77.6               | 83.2               | 68.8               | 76.8               |                         | HPLC      |
| LC0032          | 87                 | 97                 | 85                 | 93                 |                         | HPLC      |
| LC0033          | 113                | 115                | 116                | 110                |                         | HPLC      |
| LC0034          |                    |                    |                    |                    | 112                     | ELISA     |
| LC0035          | 96                 | 103                | 101                | 100                |                         | HPLC      |
| LC0036          | 75.5               | 84.2               | 87.9               | 88                 |                         | HPLC      |
| LC0037          | 96                 | 99                 | 96                 | 114                |                         | LCMS      |
| LC0038          | 88                 | 88                 | 87                 | 92                 |                         | HPLC      |
| LC0039          | 83                 | 91                 | 90                 | 94                 |                         | HPLC      |
| LC0040          |                    |                    |                    |                    |                         | HPLC      |
| LC0041          | 87.5               | 85.2               | 106.3              | 106.4              |                         | HPLC      |
| LC0042          |                    |                    |                    |                    |                         | LCMS      |
| LC0043          | 90                 | 95                 | 90                 | 100                |                         | HPLC      |
| LC0044          | 88 / 77            | 81 / 87            | 85 / 88            | 78 / 73            |                         | HPLC      |
| LC0045          | 94                 | 96                 | 96                 | 96                 |                         | HPLC      |
| LC0046          | 86                 | 84                 | 83                 | 87                 |                         | HPLC      |
| LC0047          | 92                 | 95.2               | 91.4               | 102.4              |                         | HPLC      |
| LC0048          | 92                 | 97                 | 94                 | 96                 |                         | HPLC      |
| LC0049          | 53                 | 51                 | 31                 | 21                 |                         | LCMS      |
| LC0050          | 63.3               | 80.4               | 61.9               | 78.7               |                         | HPLC      |
| LC0051          | 78.1               | 85.5               | 75.1               | 81                 |                         | LCMS      |
| LC0052          | 92 / 99            | 95 / 98            | 87 / 96            | 94 / 80            |                         | HPLC      |
| LC0053          | 100                | 100                | 100                | 100                |                         | LCMS      |
| LC0054          | 91.5               | 91.5               | 91.5               | 91.5               |                         | HPLC      |
| LC0055          | 102.13             | 102.75             | 98                 | 90.55              |                         | HPLC      |
| LC0056          | 95                 | 90                 | 88                 | 90                 |                         | HPLC      |
| LC0057          | 91                 | 99                 | 94                 | 88                 |                         | LCMS      |

# *Europe Direct is a service to help you find answers to your questions about the European Union.*

Freephone number (\*):

## 00 800 6 7 8 9 10 11

(\*) The information given is free, as are most calls (though some operators, phone boxes or hotels may charge you).

More information on the European Union is available on the internet (<u>http://europa.eu</u>).

### HOW TO OBTAIN EU PUBLICATIONS

#### **Free publications:**

- one copy: via EU Bookshop (<u>http://bookshop.europa.eu</u>);
- more than one copy or posters/maps: from the European Union's representations (<u>http://ec.europa.eu/represent\_en.htm</u>); from the delegations in non-EU countries (<u>http://eeas.europa.eu/delegations/index\_en.htm</u>); by contacting the Europe Direct service (<u>http://europa.eu/europedirect/index\_en.htm</u>) or calling 00 800 6 7 8 9 10 11 (freephone number from anywhere in the EU) (\*).

(\*) The information given is free, as are most calls (though some operators, phone boxes or hotels may charge you).

#### **Priced publications:**

• via EU Bookshop (<u>http://bookshop.europa.eu</u>).

## **JRC Mission**

As the science and knowledge service of the European Commission, the Joint Research Centre's mission is to support EU policies with independent evidence throughout the whole policy cycle.



## EU Science Hub ec.europa.eu/jrc

- 9 @EU\_ScienceHub
- f EU Science Hub Joint Research Centre
- in Joint Research Centre
- EU Science Hub



doi:10.2760/56278 ISBN 978-92-79-72215-8