



JRC CRL - FCM Database / GC-MS

Sample Name: 2,4-Toluene diisocyanate dimmer
Solvent: Hexanel
Concentration: 100 µg/mL

Date: 21/09/2005
CAS: 026747-90-0
PM ref: 25270 [U.R.N. A129]



m/z	Abundance	Ion Intensity %	m/z	Abundance	Ion Intensity %
39.10	46288,0	12,50	118.10	60928,0	16,46
50.10	48944,0	13,22	119.10	41632,0	11,25
51.10	70728,0	19,11	132.10	73632,0	19,89
52.10	47384,0	12,80	145.10	273088,0	73,77
63.10	42896,0	11,59	146.10	131392,0	35,49
64.10	47280,0	12,77	173.00	98632,0	26,64
76.10	51616,0	13,94	174.10	370176,0	100,00
91.10	76104,0	20,56			

Spectrometer: HEWLETT PACKARD GC/MS 6890/5973
Inlet system: capillary GC/MS
Scan Range: 40-700 amu
Source temperature: 230 °C

Flow: 1.2 mL/min
Column: DB 17-HT (30 m x 0.25 mm x 0.15 µm)
Programme temperature: 40 °C (3 min); 20 °C/min
(350 °C); 350 °C (20 min)